

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070924\
 Data File : BF138490.D
 Acq On : 09 Jul 2024 17:48
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF070924

Quant Time: Jul 09 18:16:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 16:58:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	40.139	-0.3	96	0.00
3	Pyridine	40.000	40.835	-2.1	99	0.00
4	n-Nitrosodimethylamine	40.000	41.107	-2.8	98	0.00
5 S	2-Fluorophenol	80.000	79.645	0.4	97	0.00
6	Aniline	40.000	40.836	-2.1	99	0.00
7 S	Phenol-d6	80.000	79.529	0.6	98	0.00
8	2-Chlorophenol	40.000	40.380	-1.0	99	0.00
9	Benzaldehyde	40.000	40.711	-1.8	102	0.00
10 C	Phenol	40.000	40.602	-1.5	100	0.00
11	bis(2-Chloroethyl)ether	40.000	39.700	0.7	97	0.00
12	1,3-Dichlorobenzene	40.000	39.941	0.1	98	0.00
13 C	1,4-Dichlorobenzene	40.000	39.776	0.6	97	0.00
14	1,2-Dichlorobenzene	40.000	39.908	0.2	98	0.00
15	Benzyl Alcohol	40.000	40.532	-1.3	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.260	-0.6	98	0.00
17	2-Methylphenol	40.000	40.100	-0.3	98	0.00
18	Hexachloroethane	40.000	40.761	-1.9	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.990	2.5	97	0.00
20	3+4-Methylphenols	40.000	39.707	0.7	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	38.233	4.4	97	0.00
23 S	Nitrobenzene-d5	80.000	79.208	1.0	98	0.00
24	Nitrobenzene	40.000	39.835	0.4	98	0.00
25	Isophorone	40.000	39.434	1.4	99	0.00
26 C	2-Nitrophenol	40.000	41.592	-4.0	101	0.00
27	2,4-Dimethylphenol	40.000	39.758	0.6	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.619	1.0	99	0.00
29 C	2,4-Dichlorophenol	40.000	40.267	-0.7	99	0.00
30	1,2,4-Trichlorobenzene	40.000	39.393	1.5	98	0.00
31	Naphthalene	40.000	39.322	1.7	98	0.00
32	Benzoic acid	40.000	42.326	-5.8	99	0.00
33	4-Chloroaniline	40.000	39.293	1.8	99	0.00
34 C	Hexachlorobutadiene	40.000	39.466	1.3	99	0.00
35	Caprolactam	40.000	40.443	-1.1	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.506	1.2	99	0.00
37	2-Methylnaphthalene	40.000	38.822	2.9	99	0.00
38	1-Methylnaphthalene	40.000	39.133	2.2	98	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.498	1.3	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.662	-4.2	98	0.00
42 S	2,4,6-Tribromophenol	80.000	79.279	0.9	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.921	-2.3	101	0.00
44	2,4,5-Trichlorophenol	40.000	39.738	0.7	96	0.00
45 S	2-Fluorobiphenyl	80.000	77.047	3.7	99	0.00
46	1,1'-Biphenyl	40.000	39.057	2.4	98	0.00
47	2-Chloronaphthalene	40.000	39.424	1.4	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070924\
 Data File : BF138490.D
 Acq On : 09 Jul 2024 17:48
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF070924

Quant Time: Jul 09 18:16:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 16:58:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.939	-2.3	99	0.00
49	Acenaphthylene	40.000	39.575	1.1	99	0.00
50	Dimethylphthalate	40.000	39.937	0.2	99	0.00
51	2,6-Dinitrotoluene	40.000	40.225	-0.6	97	0.00
52 C	Acenaphthene	40.000	40.006	-0.0	101	0.00
53	3-Nitroaniline	40.000	40.446	-1.1	98	0.00
54 P	2,4-Dinitrophenol	40.000	41.193	-3.0	103	0.00
55	Dibenzofuran	40.000	39.287	1.8	99	0.00
56 P	4-Nitrophenol	40.000	41.532	-3.8	99	0.00
57	2,4-Dinitrotoluene	40.000	41.488	-3.7	101	0.00
58	Fluorene	40.000	39.187	2.0	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.976	-2.4	100	0.00
60	Diethylphthalate	40.000	40.141	-0.4	100	0.00
61	4-Chlorophenyl-phenylether	40.000	39.172	2.1	100	0.00
62	4-Nitroaniline	40.000	40.336	-0.8	99	0.00
63	Azobenzene	40.000	39.822	0.4	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.709	-11.8	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.035	2.4	98	0.00
67	4-Bromophenyl-phenylether	40.000	39.810	0.5	100	0.00
68	Hexachlorobenzene	40.000	39.477	1.3	100	0.00
69	Atrazine	40.000	39.053	2.4	98	0.00
70 C	Pentachlorophenol	40.000	41.807	-4.5	101	0.00
71	Phenanthrene	40.000	38.889	2.8	99	0.00
72	Anthracene	40.000	39.164	2.1	99	0.00
73	Carbazole	40.000	39.033	2.4	99	0.00
74	Di-n-butylphthalate	40.000	40.015	-0.0	101	0.00
75 C	Fluoranthene	40.000	38.797	3.0	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	46.352	-15.9	117	0.00
78	Pyrene	40.000	40.602	-1.5	100	0.00
79 S	Terphenyl-d14	80.000	81.756	-2.2	102	0.00
80	Butylbenzylphthalate	40.000	42.891	-7.2	105	0.00
81	Benzo(a)anthracene	40.000	38.676	3.3	99	0.00
82	3,3'-Dichlorobenzidine	40.000	40.598	-1.5	103	0.00
83	Chrysene	40.000	39.179	2.1	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.306	-5.8	102	0.00
85 c	Di-n-octyl phthalate	40.000	39.498	1.3	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.369	-8.4	101	0.00
88	Benzo(b)fluoranthene	40.000	40.193	-0.5	107	0.00
89	Benzo(k)fluoranthene	40.000	36.262	9.3	96	0.00
90 C	Benzo(a)pyrene	40.000	39.770	0.6	103	0.00
91	Dibenzo(a,h)anthracene	40.000	43.647	-9.1	101	0.00
92	Benzo(g,h,i)perylene	40.000	43.157	-7.9	101	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070924\
Data File : BF138490.D
Acq On : 09 Jul 2024 17:48
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF070924

Quant Time: Jul 09 18:16:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF070924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 16:58:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0