

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071015\
 Data File : BF080408.D
 Acq On : 11 Jul 2015 1:42
 Operator : TP/UM
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Quant Time: Jul 11 04:44:30 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 11 04:42:33 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.15	152	19747	20.00	ng	0.00
21) Naphthalene-d8	8.43	136	79499	20.00	ng	0.00
38) Acenaphthene-d10	10.19	164	43883	20.00	ng	0.00
63) Phenanthrene-d10	11.69	188	94172	20.00	ng	0.00
75) Chrysene-d12	14.48	240	97178	20.00	ng	-0.06
86) Perylene-d12	16.18	264	95326	20.00	ng	-0.10

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.81	112	4579	4.17	ng	0.02
7) Phenol-d6	6.81	99	6283	4.32	ng	0.01
23) Nitrobenzene-d5	7.71	82	5939	4.37	ng	0.00
41) 2,4,6-Tribromophenol	10.99	330	1613	3.52	ng	0.00
44) 2-Fluorobiphenyl	9.50	172	16914	5.21	ng	0.00
78) Terphenyl-d14	13.29	244	21230	5.05	ng	-0.03

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.01	88	1177	2.27	ng	# 34
6) Aniline	6.88	93	6436	3.27	ng	# 38
8) 2-Chlorophenol	6.96	128	2712	2.12	ng	94
9) Benzaldehyde	6.72	77	2101	2.61	ng	96
10) Phenol	6.82	94	3579	2.27	ng	97
11) bis(2-Chloroethyl)ether	6.88	93	6436	5.31	ng	97
12) 1,3-Dichlorobenzene	7.16	146	4011	2.61	ng	94
13) 1,4-Dichlorobenzene	7.16	146	4011	2.70	ng	92
14) 1,2-Dichlorobenzene	7.32	146	4248	2.87	ng	95
16) 2,2'-oxybis(1-Chloropropan	7.42	45	4751	2.48	ng	# 68
17) 2-Methylphenol	7.42	107	2261	2.22	ng	# 91
18) Hexachloroethane	7.66	117	1217	2.56	ng	89
22) Acetophenone	7.56	105	3784	2.04	ng	# 87
24) Nitrobenzene	7.73	77	2965	2.09	ng	# 89
25) Isophorone	7.98	82	5311	2.04	ng	# 88
28) bis(2-Chloroethoxy)methane	8.18	93	3745	2.31	ng	93
30) 1,2,4-Trichlorobenzene	8.37	180	3438	2.72	ng	99
31) Naphthalene	8.45	128	11098	2.71	ng	99
34) Hexachlorobutadiene	8.57	225	2027	2.67	ng	94
36) 4-Chloro-3-methylphenol	9.00	107	2265	2.08	ng	# 85
37) 2-Methylnaphthalene	9.15	142	6959	2.63	ng	98
39) 1,2,4,5-Tetrachlorobenzene	9.31	216	3517	2.50	ng	94
45) 1,1'-Biphenyl	9.61	154	9001	2.31	ng	98
46) 2-Chloronaphthalene	9.64	162	6578	2.23	ng	99
48) Acenaphthylene	10.05	152	10714	2.17	ng	99
49) Dimethylphthalate	9.90	163	7376	2.09	ng	# 94
51) Acenaphthene	10.22	154	7556	2.54	ng	97
54) Dibenzofuran	10.40	168	9712	2.32	ng	94
55) 4-Nitrophenol	10.40	139	4163	7.06	ng	# 1
57) Fluorene	10.74	166	8494	2.37	ng	# 94
59) Diethylphthalate	10.60	149	7540	2.18	ng	98
60) 4-Chlorophenyl-phenylether	10.73	204	4082	2.54	ng	97
62) Azobenzene	10.89	77	7005	2.08	ng	97
65) n-Nitrosodiphenylamine	10.85	169	6361	2.12	ng	97


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071015\
 Data File : BF080408.D
 Acq On : 11 Jul 2015 1:42
 Operator : TP/UM
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Quant Time: Jul 11 04:44:30 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 11 04:42:33 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
66) 4-Bromophenyl-phenylether	11.22	248	2374	2.22	ng	97
67) Hexachlorobenzene	11.30	284	2453	2.21	ng	96
70) Phenanthrene	11.71	178	14952	2.64	ng	100
71) Anthracene	11.77	178	12117	2.20	ng	97
72) Carbazole	11.93	167	11829	2.30	ng	97
74) Fluoranthene	12.92	202	13857	2.29	ng	99
77) Pyrene	13.16	202	14937	2.53	ng	98
79) Butylbenzylphthalate	13.80	149	5011	2.14	ng	90
80) Benzo(a)anthracene	14.47	228	15407	2.62	ng	99
82) Chrysene	14.50	228	14498	2.64	ng	98
83) Bis(2-ethylhexyl)phthalate	14.42	149	7744	2.17	ng	# 91
84) Di-n-octyl phthalate	15.15	149	14086	2.36	ng	# 68
85) Indeno(1,2,3-cd)pyrene	17.83	276	19908	3.08	ng	100
87) Benzo(b)fluoranthene	15.69	252	19336	3.21	ng	95
88) Benzo(k)fluoranthene	15.69	252	19336	3.41	ng	95
89) Benzo(a)pyrene	16.10	252	15619	2.85	ng	# 90
90) Dibenzo(a,h)anthracene	17.84	278	16436	3.04	ng	# 93
91) Benzo(g,h,i)perylene	18.33	276	15557	2.85	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 7.15 min Scan# 443

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion:152 Resp: 19747

Ion Ratio Lower Upper

152 100

150 156.9 122.3 183.5

115 53.2 41.0 61.4



#2

1,4-Dioxane

Concen: 2.27 ng

RT: 3.01 min Scan# 81

Delta R.T. 0.09 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42



Tgt Ion: 88 Resp: 1177


Ion Ratio Lower Upper

88 100

58 17.7 62.4 93.6#

43 0.0 24.4 36.6#





#5

2-Fluorophenol

Concen: 4.17 ng

RT: 5.81 min Scan# 326

Delta R.T. 0.02 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion: 112 Resp: 4579

Ion Ratio Lower Upper

112 100

64 55.9 54.3 81.5




63 31.6 26.7 40.1



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 3.27 ng

RT: 6.88 min Scan# 419

Delta R.T. 0.06 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion: 93 Resp: 6436

Ion Ratio Lower Upper

93 100

66 0.0 54.8 82.2#



65 25.3 31.4 47.2#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 4.32 ng

RT: 6.81 min Scan# 413

Delta R.T. 0.01 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion: 99 Resp: 6283

Ion Ratio Lower Upper

99 100

42 21.3 17.9 26.9

71 39.1 30.1 45.1



#8

2-Chlorophenol

Concen: 2.12 ng

RT: 6.96 min Scan# 426

Delta R.T. 0.01 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42




Tgt Ion: 128 Resp: 2712


Ion Ratio Lower Upper

128 100

130 37.0 11.7 51.7

64 46.0 28.0 68.0





#9

Benzaldehyde

Concen: 2.61 ng

RT: 6.72 min Scan# 405

Delta R.T. 0.01 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion: 77 Resp: 2101

Ion Ratio Lower Upper

77 100

105 84.5 69.6 109.6




106 81.6 63.6 103.6



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 2.27 ng

RT: 6.82 min Scan# 414

Delta R.T. 0.01 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion: 94 Resp: 3579

Ion Ratio Lower Upper

94 100

65 32.4 15.0 55.0



66 45.4 27.1 67.1




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0









#11
 bis(2-Chloroethyl)ether
 Concen: 5.31 ng
 RT: 6.88 min Scan# 419
 Delta R.T. 0.00 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5


Tgt Ion: 93 Resp: 6436
 Ion Ratio Lower Upper
 93 100
 63 79.4 62.9 102.9
 95 33.6 12.2 52.2



#12
 1,3-Dichlorobenzene
 Concen: 2.61 ng
 RT: 7.16 min Scan# 444
 Delta R.T. 0.08 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42

Tgt Ion: 146 Resp: 4011
 Ion Ratio Lower Upper
 146 100
 148 67.8 49.5 74.3
 75 39.0 33.4 50.2





#13

1,4-Dichlorobenzene

Concen: 2.70 ng

RT: 7.16 min Scan# 444

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion:146 Resp: 4011

Ion Ratio Lower Upper




146 100

148 67.8 50.3 75.5

111 47.7 33.1 49.7



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 2.87 ng

RT: 7.32 min Scan# 458

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion:146 Resp: 4248

Ion Ratio Lower Upper



146 100


148 60.7 52.2 78.4

111 42.7 32.2 48.2



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E









#16
2,2'-oxybis(1-chloropropane)
Concen: 2.48 ng
RT: 7.42 min Scan# 467
Delta R.T. 0.01 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 45 Resp: 4751
Ion Ratio Lower Upper
45 100
77 38.3 39.1 79.1#
79 30.5 37.2 77.2#



#17
2-Methylphenol
Concen: 2.22 ng
RT: 7.42 min Scan# 467
Delta R.T. 0.01 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion: 107 Resp: 2261
Ion Ratio Lower Upper
107 100
108 117.1 88.0 132.0
77 72.4 45.9 68.9#
79 57.7 44.5 66.7









#18
Hexachloroethane
Concen: 2.56 ng
RT: 7.66 min Scan# 488
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:117 Resp: 1217
Ion Ratio Lower Upper
117 100
119 108.7 78.1 117.1
201 119.6 86.5 129.7




Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E






#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.43 min Scan# 555
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion:136 Resp: 79499
Ion Ratio Lower Upper
136 100
137 11.3 9.0 13.4
54 10.8 9.1 13.7
68 7.2 5.9 8.9



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0





#22

Acetophenone

Concen: 2.04 ng

RT: 7.56 min Scan# 479

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5

Tgt Ion:105 Resp: 3784


Ion Ratio Lower Upper

105 100

71 16.3 5.2 7.8#

51 41.7 28.7 43.1

120 14.7 16.4 24.6#






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E



#23

Nitrobenzene-d5

Concen: 4.37 ng

RT: 7.71 min Scan# 492

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion: 82 Resp: 5939

Ion Ratio Lower Upper

82 100

128 29.6 26.9 40.3



54 58.7 47.7 71.5




Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF0









#24
 Nitrobenzene
 Concen: 2.09 ng
 RT: 7.73 min Scan# 494
 Delta R.T. 0.00 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tgt Ion: 77 Resp: 2965
 Ion Ratio Lower Upper
 77 100
 123 35.3 34.0 51.0
 65 10.0 11.2 16.8#




Abundance Ion 77.00 (76.70 to 77.70): BF0
 Ion 123.00 (122.70 to 123.70): BF0
 Ion 65.00 (64.70 to 65.70): BF0






#25
 Isophorone
 Concen: 2.04 ng
 RT: 7.98 min Scan# 516
 Delta R.T. 0.01 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42

Tgt Ion: 82 Resp: 5311
 Ion Ratio Lower Upper
 82 100
 95 7.7 5.4 8.2
 138 18.4 9.5 14.3#



Abundance Ion 82.00 (81.70 to 82.70): BF0
 Ion 95.00 (94.70 to 95.70): BF0
 Ion 138.00 (137.70 to 138.70): BF0









#28
bis(2-Chloroethoxy)methane
Concen: 2.31 ng
RT: 8.18 min Scan# 533
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 93 Resp: 3745
Ion Ratio Lower Upper
93 100
95 27.7 25.8 38.8
123 10.5 9.7 14.5



#30
1,2,4-Trichlorobenzene
Concen: 2.72 ng
RT: 8.37 min Scan# 550
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion: 180 Resp: 3438
Ion Ratio Lower Upper
180 100
182 94.9 75.6 113.4
145 28.1 23.3 34.9





#31

Naphthalene

Concen: 2.71 ng

RT: 8.45 min Scan# 557

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion:128 Resp: 11098

Ion Ratio Lower Upper

128 100

129 11.5 9.1 13.7




127 13.0 10.7 16.1



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#34

Hexachlorobutadiene

Concen: 2.67 ng

RT: 8.57 min Scan# 567

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion:225 Resp: 2027

Ion Ratio Lower Upper

225 100

223 64.5 49.5 74.3



227 67.0 48.0 72.0




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E









#36
 4-Chloro-3-methylphenol
 Concen: 2.08 ng
 RT: 9.00 min Scan# 605
 Delta R.T. 0.00 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5


Tgt Ion:107 Resp: 2265
 Ion Ratio Lower Upper
 107 100
 144 22.4 22.6 34.0#
 142 71.5 68.7 103.1



#37
 2-Methylnaphthalene
 Concen: 2.63 ng
 RT: 9.15 min Scan# 618
 Delta R.T. 0.00 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42

Tgt Ion:142 Resp: 6959
 Ion Ratio Lower Upper
 142 100
 141 92.0 73.0 109.4
 115 34.0 25.0 37.4





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 10.19 min Scan# 709

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion:164 Resp: 43883

Ion Ratio Lower Upper

164 100

162 104.8 82.7 124.1




160 45.0 35.5 53.3



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#39

1,2,4,5-Tetrachlorobenzene

Concen: 2.50 ng

RT: 9.31 min Scan# 632

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Tgt Ion:216 Resp: 3517


Ion Ratio Lower Upper

216 100

214 76.1 64.6 97.0

179 19.9 17.9 26.9

108 16.5 15.8 23.6






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E









#41
 2,4,6-Tribromophenol
 Concen: 3.52 ng
 RT: 10.99 min Scan# 779
 Delta R.T. 0.00 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tgt Ion: 330 Resp: 1613
 Ion Ratio Lower Upper
 330 100
 332 96.4 78.0 117.0
 141 36.5 28.6 43.0




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 5.21 ng
 RT: 9.50 min Scan# 649
 Delta R.T. 0.00 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42

Tgt Ion: 172 Resp: 16914
 Ion Ratio Lower Upper
 172 100
 171 34.7 28.4 42.6
 170 22.4 18.6 28.0



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 2.31 ng
 RT: 9.61 min Scan# 658
 Delta R.T. 0.00 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Tgt Ion:	154	Resp:	9001
Ion	Ratio	Lower	Upper
154	100		
153	41.8	20.6	60.6
76	13.2	0.0	32.4




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 2.23 ng
 RT: 9.64 min Scan# 661
 Delta R.T. 0.00 min
 Lab File: BF080408.D
 Acq: 11 Jul 2015 1:42

Tgt Ion:	162	Resp:	6578
Ion	Ratio	Lower	Upper
162	100		
127	35.4	28.6	42.8
164	35.1	27.2	40.8



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E





#48

Acenaphthylene

Concen: 2.17 ng

RT: 10.05 min Scan# 697

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion:152 Resp: 10714

Ion Ratio Lower Upper




152 100

151 20.3 15.8 23.8

153 13.1 10.2 15.4



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#49

Dimethylphthalate

Concen: 2.09 ng

RT: 9.90 min Scan# 684

Delta R.T. -0.01 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion:163 Resp: 7376

Ion Ratio Lower Upper



163 100


194 0.0 3.4 5.2#

164 10.0 7.4 11.0



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E





#51

Acenaphthene

Concen: 2.54 ng

RT: 10.22 min Scan# 712

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion:154 Resp: 7556

Ion Ratio Lower Upper

154 100

153 118.4 91.7 137.5




152 54.0 43.5 65.3



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#54

Dibenzofuran

Concen: 2.32 ng

RT: 10.40 min Scan# 727

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion:168 Resp: 9712

Ion Ratio Lower Upper

168 100

139 46.9 33.4 50.2



169 13.7 11.1 16.7




Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E





#55

4-Nitrophenol

Concen: 7.06 ng

RT: 10.40 min Scan# 727

Delta R.T. 0.07 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion:139 Resp: 4163

Ion Ratio Lower Upper

139 100

109 0.0 54.6 94.6#




65 0.0 97.8 137.8#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#57

Fluorene

Concen: 2.37 ng

RT: 10.74 min Scan# 757

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion:166 Resp: 8494

Ion Ratio Lower Upper

166 100

165 103.7 78.3 117.5



167 16.3 10.6 15.8#




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#59
Diethylphthalate
Concen: 2.18 ng
RT: 10.60 min Scan# 745
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:149 Resp: 7540
Ion Ratio Lower Upper
149 100
177 21.0 15.4 23.2
150 12.7 10.3 15.5




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 2.54 ng
RT: 10.73 min Scan# 756
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion:204 Resp: 4082
Ion Ratio Lower Upper
204 100
206 35.3 26.8 40.2
141 66.4 55.1 82.7



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E





#62

Azobenzene

Concen: 2.08 ng

RT: 10.89 min Scan# 770

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5

Tgt Ion: 77 Resp: 7005


Ion Ratio Lower Upper

77 100

182 19.8 0.0 39.4

105 22.9 0.6 40.6

51 32.5 11.5 51.5






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0



#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.69 min Scan# 840

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion: 188 Resp: 94172

Ion Ratio Lower Upper

188 100

94 8.1 6.7 10.1



80 8.5 7.5 11.3




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0









#65
n-Nitrosodiphenylamine
Concen: 2.12 ng
RT: 10.85 min Scan# 767
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:169 Resp: 6361
Ion Ratio Lower Upper
169 100
168 70.5 53.5 80.3
167 36.5 29.4 44.2




Abundance Ion 169.00 (168.70 to 169.70): E
Ion 168.00 (167.70 to 168.70): E
Ion 167.00 (166.70 to 167.70): E






#66
4-Bromophenyl-phenylether
Concen: 2.22 ng
RT: 11.22 min Scan# 799
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion:248 Resp: 2374
Ion Ratio Lower Upper
248 100
250 95.1 74.2 111.4
141 91.2 70.7 106.1



Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 2.21 ng
RT: 11.30 min Scan# 806
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 284 Resp: 2453
Ion Ratio Lower Upper
284 100
142 16.7 12.9 19.3
249 32.5 23.5 35.3




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#70
Phenanthrene
Concen: 2.64 ng
RT: 11.71 min Scan# 842
Delta R.T. 0.00 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion: 178 Resp: 14952
Ion Ratio Lower Upper
178 100
176 19.2 15.2 22.8
179 15.3 12.4 18.6



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 2.20 ng

RT: 11.77 min Scan# 847

Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :


BNA_F

ClientSampleId :

SSTDICC2.5

Tgt Ion:178 Resp: 12117




Ion	Ratio	Lower	Upper
178	100		
176	19.0	14.2	21.4
179	15.7	11.6	17.4



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 2.30 ng

RT: 11.93 min Scan# 861


Delta R.T. 0.00 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Tgt Ion:167 Resp: 11829



Ion	Ratio	Lower	Upper
167	100		
166	20.1	17.4	26.2
139	12.2	9.5	14.3




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#74

Fluoranthene

Concen: 2.29 ng

RT: 12.92 min Scan# 948

Delta R.T. -0.01 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion:202 Resp: 13857

Ion Ratio Lower Upper

202 100

101 8.6 0.0 28.8




203 18.0 0.0 37.3



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.48 min Scan# 1084

Delta R.T. -0.06 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion:240 Resp: 97178

Ion Ratio Lower Upper

240 100

120 7.5 7.4 11.2



236 25.4 19.7 29.5




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Pyrene

Concen: 2.53 ng

RT: 13.16 min Scan# 969

Delta R.T. -0.01 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion: 202 Resp: 14937

Ion Ratio Lower Upper

202 100

200 20.6 17.0 25.6




203 18.2 13.8 20.8



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 5.05 ng

RT: 13.29 min Scan# 980

Delta R.T. -0.03 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion: 244 Resp: 21230

Ion Ratio Lower Upper

244 100

212 7.8 4.9 7.3#



122 14.3 4.1 6.1#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 2.14 ng
RT: 13.80 min Scan# 1025
Delta R.T. -0.03 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:149 Resp: 5011
Ion Ratio Lower Upper
149 100
91 68.1 62.5 93.7
206 19.1 14.0 21.0




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 91.00 (90.70 to 91.70): BF0
Ion 206.00 (205.70 to 206.70): E






#80
Benzo(a)anthracene
Concen: 2.62 ng
RT: 14.47 min Scan# 1083
Delta R.T. -0.06 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion:228 Resp: 15407
Ion Ratio Lower Upper
228 100
226 27.2 21.7 32.5
229 20.3 15.5 23.3



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E





#82

Chrysene

Concen: 2.64 ng

RT: 14.50 min Scan# 1086

Delta R.T. -0.07 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5





Tgt Ion: 228 Resp: 14498

Ion Ratio Lower Upper

228 100

226 28.6 24.3 36.5

229 19.8 15.5 23.3



#83

Bis(2-ethylhexyl)phthalate

Concen: 2.17 ng

RT: 14.42 min Scan# 1079

Delta R.T. -0.07 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42




Tgt Ion: 149 Resp: 7744


Ion Ratio Lower Upper

149 100

167 23.7 22.5 33.7

279 0.0 3.5 5.3#









#84
Di-n-octyl phthalate
Concen: 2.36 ng
RT: 15.15 min Scan# 1143
Delta R.T. -0.09 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:149 Resp: 14086
Ion Ratio Lower Upper
149 100
167 0.0 1.2 1.8#
43 24.9 9.0 13.6#




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 43.00 (42.70 to 43.70): BF0






#85
Indeno(1,2,3-cd)pyrene
Concen: 3.08 ng
RT: 17.83 min Scan# 1377
Delta R.T. -0.13 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion:276 Resp: 19908
Ion Ratio Lower Upper
276 100
138 26.6 21.4 32.0
277 24.6 19.5 29.3



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#86
Perylene-d12
Concen: 20.00 ng
RT: 16.18 min Scan# 1233
Delta R.T. -0.10 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 264 Resp: 95326
Ion Ratio Lower Upper
264 100
260 23.6 19.4 29.2
265 20.9 17.8 26.8



#87
Benzo(b)fluoranthene
Concen: 3.21 ng
RT: 15.69 min Scan# 1190
Delta R.T. -0.10 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion: 252 Resp: 19336
Ion Ratio Lower Upper
252 100
253 25.4 17.8 26.8
125 10.1 8.6 12.8









#88
Benzo(k)fluoranthene
Concen: 3.41 ng
RT: 15.69 min Scan# 1190
Delta R.T. -0.14 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 252 Resp: 19336
Ion Ratio Lower Upper
252 100
253 25.4 17.7 26.5
125 10.1 8.4 12.6



#89
Benzo(a)pyrene
Concen: 2.85 ng
RT: 16.10 min Scan# 1226
Delta R.T. -0.11 min
Lab File: BF080408.D
Acq: 11 Jul 2015 1:42

Tgt Ion: 252 Resp: 15619
Ion Ratio Lower Upper
252 100
253 25.3 17.0 25.6
125 15.0 8.1 12.1#





#90

Dibenzo(a,h)anthracene

Concen: 3.04 ng

RT: 17.84 min Scan# 1378

Delta R.T. -0.13 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5


Tgt Ion: 278 Resp: 16436

Ion Ratio Lower Upper

278 100

139 19.9 11.8 17.8#




279 25.6 18.8 28.2



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 2.85 ng

RT: 18.33 min Scan# 1421

Delta R.T. -0.11 min

Lab File: BF080408.D

Acq: 11 Jul 2015 1:42


Tgt Ion: 276 Resp: 15557

Ion Ratio Lower Upper

276 100

277 25.0 18.2 27.4

138 23.1 20.3 30.5



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

