

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071015\
 Data File : BF080421.D
 Acq On : 11 Jul 2015 7:55
 Operator : TP/UM
 Sample : G2838-05DL 2X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEC-B7DL

Manual Integrations
 APPROVED

apatel
 7/13/2015 3:31:51 PM

Quant Time: Jul 13 01:23:41 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 11 05:35:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.14	152	20305	20.00	ng	-0.01
21) Naphthalene-d8	8.43	136	87610	20.00	ng	0.00
38) Acenaphthene-d10	10.19	164	43135	20.00	ng	0.00
63) Phenanthrene-d10	11.69	188	85089	20.00	ng	0.00
75) Chrysene-d12	14.58	240	97338	20.00	ng	0.05
86) Perylene-d12	16.36	264	95213	20.00	ng	0.08
System Monitoring Compounds						
5) 2-Fluorophenol	5.79	112	99569	84.03	ng	0.00
7) Phenol-d6	6.80	99	123216	78.05	ng	0.00
23) Nitrobenzene-d5	7.71	82	78868	51.34	ng	0.00
41) 2,4,6-Tribromophenol	10.98	330	28720	68.22	ng	-0.01
44) 2-Fluorobiphenyl	9.50	172	170615	53.27	ng	0.00
78) Terphenyl-d14	13.33	244	195958	43.72	ng	0.01
Target Compounds						Qvalue
48) Acenaphthylene	10.05	152	16173	3.41	ng	99
49) Dimethylphthalate	9.90	163	15065	4.62	ng	99
70) Phenanthrene	11.71	178	84580	16.55	ng	99
71) Anthracene	11.76	178	30608	6.47	ng	98
74) Fluoranthene	12.95	202	285872	53.92	ng	100
77) Pyrene	13.20	202	247677	39.72	ng	99
80) Benzo(a)anthracene	14.57	228	184414	30.06	ng	96
82) Chrysene	14.60	228	110881	19.63	ng	96
85) Indeno(1,2,3-cd)pyrene	18.02	276	54394	7.75	ng	96
87) Benzo(b)fluoranthene	15.87	252	170836m	25.83	ng	
88) Benzo(k)fluoranthene	15.89	252	34792m	6.43	ng	
89) Benzo(a)pyrene	16.29	252	95660	16.51	ng	97
90) Dibenzo(a,h)anthracene	18.03	278	15277	2.66	ng	# 90
91) Benzo(g,h,i)perylene	18.52	276	54720	9.60	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071015\
Data File : BF080421.D
Acq On : 11 Jul 2015 7:55
Operator : TP/UM
Sample : G2838-05DL 2X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEC-B7DL

Manual Integrations
APPROVED

apatel
7/13/2015 3:31:51 PM

Quant Time: Jul 13 01:23:41 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 11 05:35:30 2015
Response via : Initial Calibration

