

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071015\
 Data File : BF080422.D
 Acq On : 11 Jul 2015 8:24
 Operator : TP/UM
 Sample : G2925-06 2X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEC-E1

Manual Integrations
 APPROVED

apatel
 7/13/2015 3:31:55 PM

Quant Time: Jul 13 01:27:15 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 11 05:35:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.14	152	21321	20.00	ng	-0.01
21) Naphthalene-d8	8.43	136	87782	20.00	ng	0.00
38) Acenaphthene-d10	10.19	164	44054	20.00	ng	0.00
63) Phenanthrene-d10	11.69	188	77463	20.00	ng	0.00
75) Chrysene-d12	14.48	240	93715	20.00	ng	-0.06
86) Perylene-d12	16.18	264	92994	20.00	ng	-0.10
System Monitoring Compounds						
5) 2-Fluorophenol	5.79	112	53249	42.80	ng	0.00
7) Phenol-d6	6.80	99	73777	44.51	ng	0.00
23) Nitrobenzene-d5	7.71	82	45115	29.31	ng	0.00
41) 2,4,6-Tribromophenol	10.98	330	8770	20.40	ng	-0.01
44) 2-Fluorobiphenyl	9.50	172	97103	29.69	ng	0.00
78) Terphenyl-d14	13.29	244	109981	25.49	ng	-0.03
Target Compounds						Qvalue
49) Dimethylphthalate	9.90	163	34353	10.32	ng	99
70) Phenanthrene	11.71	178	67928	14.60	ng	99
71) Anthracene	11.76	178	19427	4.51	ng	98
74) Fluoranthene	12.91	202	154486	32.01	ng	95
77) Pyrene	13.15	202	149690	24.93	ng	99
80) Benzo(a)anthracene	14.47	228	94650	16.03	ng	99
82) Chrysene	14.50	228	62914	11.57	ng	97
85) Indeno(1,2,3-cd)pyrene	17.83	276	57088	8.45	ng	94
87) Benzo(b)fluoranthene	15.70	252	99934m	15.47	ng	
88) Benzo(k)fluoranthene	15.71	252	43409m	8.22	ng	
89) Benzo(a)pyrene	16.11	252	80466	14.22	ng	97
90) Dibenzo(a,h)anthracene	17.84	278	13615	2.43	ng	# 83
91) Benzo(g,h,i)perylene	18.33	276	60613	10.88	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071015\
Data File : BF080422.D
Acq On : 11 Jul 2015 8:24
Operator : TP/UM
Sample : G2925-06 2X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEC-E1

Manual Integrations
APPROVED

apatel
7/13/2015 3:31:55 PM

Quant Time: Jul 13 01:27:15 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 11 05:35:30 2015
Response via : Initial Calibration

