

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071017\
 Data File : BF096610.D
 Acq On : 10 Jul 2017 19:30
 Operator : SJ/JU
 Sample : I4101-03MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Jul 11 01:11:15 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 07 13:02:33 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.69	152	120500	20.00	ng	-0.01
21) Naphthalene-d8	7.97	136	517040	20.00	ng	-0.02
38) Acenaphthene-d10	9.72	164	228515	20.00	ng	-0.02
63) Phenanthrene-d10	11.20	188	380684	20.00	ng	-0.01
75) Chrysene-d12	13.83	240	230497	20.00	ng	-0.02
86) Perylene-d12	15.22	264	207968	20.00	ng	-0.03

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.30	112	992383	121.55	ng	0.00
7) Phenol-d6	6.33	99	1214772	121.03	ng	-0.01
23) Nitrobenzene-d5	7.25	82	772263	89.75	ng	-0.02
41) 2,4,6-Tribromophenol	10.51	330	249578	139.82	ng	-0.01
44) 2-Fluorobiphenyl	9.05	172	1265789	94.71	ng	-0.02
78) Terphenyl-d14	12.78	244	987533	91.55	ng	-0.02

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.39	88	131086	33.86	ng	# 77
3) Pyridine	3.08	79	362174	34.10	ng	# 65
4) n-Nitrosodimethylamine	3.02	42	219567	41.88	ng	# 85
6) Aniline	6.35	93	426407	32.87	ng	# 68
8) 2-Chlorophenol	6.47	128	363953	41.80	ng	# 97
9) Benzaldehyde	6.23	77	211475	33.67	ng	100
10) Phenol	6.35	94	465011	40.67	ng	81
11) bis(2-Chloroethyl)ether	6.43	93	355804	40.27	ng	95
12) 1,3-Dichlorobenzene	6.63	146	386876	42.38	ng	99
13) 1,4-Dichlorobenzene	6.70	146	395493	43.35	ng	97
14) 1,2-Dichlorobenzene	6.86	146	360340	43.01	ng	97
15) Benzyl Alcohol	6.83	79	305214	45.49	ng	98
16) 2,2'-oxybis(1-Chloropropan	6.97	45	764998	42.59	ng	92
17) 2-Methylphenol	6.95	107	311432	44.89	ng	99
18) Hexachloroethane	7.20	117	145889	44.10	ng	# 82
19) n-Nitroso-di-n-propylamine	7.11	70	249385	41.71	ng	# 84
20) 3+4-Methylphenols	7.10	107	361494	46.71	ng	# 86
22) Acetophenone	7.10	105	491912	43.54	ng	# 91
24) Nitrobenzene	7.27	77	396341	43.91	ng	93
25) Isophorone	7.51	82	689436	41.33	ng	96
26) 2-Nitrophenol	7.59	139	216945	45.41	ng	90
27) 2,4-Dimethylphenol	7.63	122	353472	43.45	ng	95
28) bis(2-Chloroethoxy)methane	7.72	93	494954	44.94	ng	100
29) 2,4-Dichlorophenol	7.83	162	327919	46.72	ng	96
30) 1,2,4-Trichlorobenzene	7.91	180	298631	45.10	ng	97
31) Naphthalene	7.99	128	1108909	45.43	ng	99
32) Benzoic acid	7.73	122	67773	9.92	ng	92
33) 4-Chloroaniline	8.03	127	324670	32.02	ng	98
34) Hexachlorobutadiene	8.12	225	154208	45.41	ng	98
35) Caprolactam	8.42	113	80933	33.44	ng	# 63
36) 4-Chloro-3-methylphenol	8.53	107	350570	45.14	ng	90
37) 2-Methylnaphthalene	8.68	142	729363	46.94	ng	98
39) 1,2,4,5-Tetrachlorobenzene	8.85	216	244061	41.13	ng	99
40) Hexachlorocyclopentadiene	8.84	237	247866	85.91	ng	99

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071017\
 Data File : BF096610.D
 Acq On : 10 Jul 2017 19:30
 Operator : SJ/JU
 Sample : I4101-03MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Jul 11 01:11:15 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 07 13:02:33 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	8.96	196	204753	43.62	nq	99
43) 2,4,5-Trichlorophenol	9.00	196	204966	44.16	nq	96
45) 1,1'-Biphenyl	9.15	154	818956	41.83	nq	97
46) 2-Chloronaphthalene	9.17	162	639951	43.86	nq	94
47) 2-Nitroaniline	9.26	65	252815	47.62	nq	93
48) Acenaphthylene	9.58	152	1044680	45.18	nq	99
49) Dimethylphthalate	9.45	163	913566	53.55	nq	100
50) 2,6-Dinitrotoluene	9.51	165	173345	45.95	nq	91
51) Acenaphthene	9.75	154	581678	40.81	nq	99
52) 3-Nitroaniline	9.67	138	165818	35.24	nq	91
53) 2,4-Dinitrophenol	9.78	184	133135	66.45	nq	# 74
54) Dibenzofuran	9.93	168	868965	46.18	nq	100
55) 4-Nitrophenol	9.85	139	311970	80.00	nq	# 65
56) 2,4-Dinitrotoluene	9.91	165	228665	47.87	nq	# 79
57) Fluorene	10.27	166	636432	47.89	nq	99
58) 2,3,4,6-Tetrachlorophenol	10.05	232	164984	51.38	nq	97
59) Diethylphthalate	10.15	149	735489	45.61	nq	99
60) 4-Chlorophenyl-phenylether	10.26	204	267469	46.26	nq	99
61) 4-Nitroaniline	10.29	138	215091	50.28	nq	90
62) Azobenzene	10.42	77	743331	47.68	nq	91
64) 4,6-Dinitro-2-methylphenol	10.32	198	119064	45.35	nq	81
65) n-Nitrosodiphenylamine	10.38	169	624726	46.77	nq	97
66) 4-Bromophenyl-phenylether	10.75	248	170048	45.89	nq	94
67) Hexachlorobenzene	10.82	284	181445	44.73	nq	# 89
68) Atrazine	10.91	200	169456	44.41	nq	94
69) Pentachlorophenol	11.01	266	191923	79.83	nq	96
70) Phenanthrene	11.22	178	947677	46.05	nq	99
71) Anthracene	11.27	178	974249	46.46	nq	99
72) Carbazole	11.43	167	943681	44.75	nq	99
73) Di-n-butylphthalate	11.77	149	1119782	43.84	nq	99
74) Fluoranthene	12.40	202	977231	48.43	nq	97
76) Benzidine	12.53	184	534653	48.34	nq	# 92
77) Pyrene	12.63	202	980190	52.98	nq	99
79) Butylbenzylphthalate	13.26	149	481198	51.38	nq	# 81
80) Benzo(a)anthracene	13.82	228	649840	48.69	nq	99
81) 3,3'-Dichlorobenzidine	13.78	252	176241	32.15	nq	# 96
82) Chrysene	13.85	228	636127	45.51	nq	100
83) Bis(2-ethylhexyl)phthalate	13.82	149	488502	46.73	nq	100
84) Di-n-octyl phthalate	14.43	149	877275	42.61	nq	# 94
85) Indeno(1,2,3-cd)pyrene	16.56	276	599716	54.35	nq	98
87) Benzo(b)fluoranthene	14.83	252	551636	42.33	nq	99
88) Benzo(k)fluoranthene	14.83	252	551636	47.32	nq	96
89) Benzo(a)pyrene	15.16	252	522419	44.90	nq	98
90) Dibenzo(a,h)anthracene	16.57	278	479015	52.34	nq	# 95
91) Benzo(g,h,i)perylene	16.96	276	505353	53.28	nq	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071017\
Data File : BF096610.D
Acq On : 10 Jul 2017 19:30
Operator : SJ/JU
Sample : I4101-03MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :

Quant Time: Jul 11 01:11:15 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF070117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 07 13:02:33 2017
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.69 min Scan# 782
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampleId :




Tot Ion:152 Resp: 120500
Ion Ratio Lower Upper
152 100
150 154.5 126.2 189.2
115 63.8 52.2 78.2




#2

1,4-Dioxane
Concen: 33.86 ng
RT: 2.39 min Scan# 52
Delta R.T. 0.03 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion: 88 Resp: 131086
Ion Ratio Lower Upper
88 100
58 98.8 61.4 92.0#
43 38.5 23.4 35.0#





#3

Pvridine

Concen: 34.10 ng

RT: 3.08 min Scan# 168

Delta R.T. 0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 79 Resp: 362174

Ion Ratio Lower Upper

79 100

52 101.4 54.8 82.2#

51 48.0 27.0 40.4#



#4

n-Nitrosodimethylamine

Concen: 41.88 ng

RT: 3.02 min Scan# 159

Delta R.T. 0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30




Tgt Ion: 42 Resp: 219567


Ion Ratio Lower Upper

42 100

74 111.4 103.9 155.9

44 7.6 5.7 8.5





#5

2-Fluorophenol

Concen: 121.55 ng

RT: 5.30 min Scan# 547

Delta R.T. -0.01 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampled :


Tot Ion:112 Resp: 992383

Ion Ratio Lower Upper

112 100

64 66.7 46.0 69.0




63 33.8 23.4 35.0



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 32.87 ng

RT: 6.35 min Scan# 725

Delta R.T. -0.01 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Tgt Ion: 93 Resp: 426407

Ion Ratio Lower Upper

93 100

66 60.5 32.9 49.3#



65 35.6 16.0 24.0#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 121.03 ng

RT: 6.33 min Scan# 722

Delta R.T. -0.01 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Instrument :

BNA_F




ClientSampled :

Tot Ion: 99 Resp: 1214772

Ion	Ratio	Lower	Upper
99	100		
42	22.5	13.5	20.3#
71	31.9	24.5	36.7



Abundance Ion 99.00 (98.70 to 99.70): BF0
Ion 42.00 (41.70 to 42.70): BF0
Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 41.80 ng

RT: 6.47 min Scan# 746


Delta R.T. -0.01 min

Lab File: BF096610.D



Acq: 10 Jul 2017 19:30


Tgt Ion: 128 Resp: 363953

Ion	Ratio	Lower	Upper
128	100		
130	33.0	12.9	52.9
64	52.1	28.4	68.4



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BF0









#9
Benzaldehyde
Concen: 33.67 ng
RT: 6.23 min Scan# 704
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampled :


Tot Ion: 77 Resp: 211475
Ion Ratio Lower Upper
77 100
105 103.9 83.8 123.8
106 99.2 79.1 119.1



#10
Phenol
Concen: 40.67 ng
RT: 6.35 min Scan# 724
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion: 94 Resp: 465011
Ion Ratio Lower Upper
94 100
65 39.3 9.9 49.9
66 56.2 23.1 63.1









#11
bis(2-Chloroethyl)ether
Concen: 40.27 ng
RT: 6.43 min Scan# 738
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampleId :


Tot Ion: 93 Resp: 355804
Ion Ratio Lower Upper
93 100
63 85.7 59.2 99.2
95 32.5 12.8 52.8



#12
1,3-Dichlorobenzene
Concen: 42.38 ng
RT: 6.63 min Scan# 772
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion: 146 Resp: 386876
Ion Ratio Lower Upper
146 100
148 65.2 51.8 77.6
75 28.3 23.1 34.7









#13
 1,4-Dichlorobenzene
 Concen: 43.35 ng
 RT: 6.70 min Scan# 785
 Delta R.T. -0.01 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:146 Resp: 395493
 Ion Ratio Lower Upper
 146 100
 148 64.5 52.2 78.2
 111 42.2 30.3 45.5



#14
 1,2-Dichlorobenzene
 Concen: 43.01 ng
 RT: 6.86 min Scan# 811
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion:146 Resp: 360340
 Ion Ratio Lower Upper
 146 100
 148 64.8 50.4 75.6
 111 45.7 38.2 57.4









#15
Benzyl Alcohol
Concen: 45.49 ng
RT: 6.83 min Scan# 807
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampled :


Tot Ion: 79 Resp: 305214
Ion Ratio Lower Upper
79 100
108 81.2 63.4 95.0
77 60.4 49.2 73.8



#16
2,2'-oxybis(1-chloropropane)
Concen: 42.59 ng
RT: 6.97 min Scan# 830
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion: 45 Resp: 764998
Ion Ratio Lower Upper
45 100
77 9.4 0.0 33.2
79 7.4 0.0 30.0









#17
2-Methylphenol
Concen: 44.89 ng
RT: 6.95 min Scan# 827
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampled :

Tot Ion:107 Resp: 311432
Ion Ratio Lower Upper
107 100
108 115.3 93.4 140.0
77 44.6 35.8 53.6
79 44.5 34.8 52.2




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF0
Ion 79.00 (78.70 to 79.70): BF0






#18
Hexachloroethane
Concen: 44.10 ng
RT: 7.20 min Scan# 869
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:117 Resp: 145889
Ion Ratio Lower Upper
117 100
119 99.0 77.4 116.0
201 84.7 94.6 141.8#



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 41.71 ng

RT: 7.11 min Scan# 854

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :

Tot Ion: 70 Resp: 249385


Ion Ratio Lower Upper

70 100

42 76.5 47.2 70.8#

101 10.0 7.2 10.8

130 20.5 15.9 23.9






Abundance Ion 70.00 (69.70 to 70.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#20

3+4-Methylphenols

Concen: 46.71 ng

RT: 7.10 min Scan# 853

Delta R.T. -0.01 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Tgt Ion: 107 Resp: 361494


Ion Ratio Lower Upper

107 100

108 90.7 71.0 111.0

77 80.2 90.5 130.5#

79 28.4 8.4 48.4






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 7.97 min Scan# 1000

Delta R.T. -0.02 min

Lab File: BF096610.D


Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136	Resp:	517040
Ion Ratio	Lower	Upper
136	100	
137	10.4	8.5 12.7
54	9.6	4.9 7.3#
68	4.8	4.1 6.1






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 43.54 ng


RT: 7.10 min Scan# 852

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Tgt Ion:105	Resp:	491912
Ion Ratio	Lower	Upper
105	100	
71	5.0	2.8 4.2#
51	46.9	30.7 46.1#
120	21.4	17.4 26.0






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E





#23

Nitrobenzene-d5

Concen: 89.75 ng

RT: 7.25 min Scan# 878

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampled :


Tot Ion: 82 Resp: 772263

Ion Ratio Lower Upper

82 100

128 47.7 34.0 51.0




54 66.1 42.2 63.2#



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): BF0

Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 43.91 ng

RT: 7.27 min Scan# 881

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Tgt Ion: 77 Resp: 396341

Ion Ratio Lower Upper

77 100

123 49.6 35.8 53.8



65 14.8 10.6 15.8




Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 123.00 (122.70 to 123.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 41.33 ng

RT: 7.51 min Scan# 922

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 82 Resp: 689436

Ion Ratio Lower Upper

82 100

95 6.7 5.3 7.9

138 18.8 13.1 19.7



#26

2-Nitrophenol

Concen: 45.41 ng

RT: 7.59 min Scan# 935

Delta R.T. -0.01 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30




Tgt Ion: 139 Resp: 216945


Ion Ratio Lower Upper

139 100

109 21.6 21.0 31.6

65 47.4 33.0 49.6









#27
 2,4-Dimethylphenol
 Concen: 43.45 ng
 RT: 7.63 min Scan# 943
 Delta R.T. -0.01 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:122 Resp: 353472
 Ion Ratio Lower Upper
 122 100
 107 104.6 89.2 133.8
 121 59.0 45.2 67.8




Abundance Ion 122.00 (121.70 to 122.70): E
 Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E






#28
 bis(2-Chloroethoxy)methane
 Concen: 44.94 ng
 RT: 7.72 min Scan# 958
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion: 93 Resp: 494954
 Ion Ratio Lower Upper
 93 100
 95 32.9 26.5 39.7
 123 11.0 9.0 13.4



Abundance Ion 93.00 (92.70 to 93.70): BF0
 Ion 95.00 (94.70 to 95.70): BF0
 Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 46.72 ng
 RT: 7.83 min Scan# 976
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:162 Resp: 327919
 Ion Ratio Lower Upper
 162 100
 164 63.6 44.6 84.6
 98 39.2 14.8 54.8




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): B






#30
 1,2,4-Trichlorobenzene
 Concen: 45.10 ng
 RT: 7.91 min Scan# 990
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion:180 Resp: 298631
 Ion Ratio Lower Upper
 180 100
 182 93.4 75.8 113.6
 145 36.0 25.3 37.9



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 45.43 ng

RT: 7.99 min Scan# 1004

Delta R.T. -0.01 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :


Tot Ion:128 Resp: 1108909

Ion Ratio Lower Upper

128 100

129 10.9 9.0 13.6




127 12.9 10.8 16.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 9.92 ng

RT: 7.73 min Scan# 959

Delta R.T. -0.06 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Tgt Ion:122 Resp: 67773

Ion Ratio Lower Upper

122 100

105 116.4 103.3 143.3



77 83.2 73.7 113.7




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 32.02 ng
RT: 8.03 min Scan# 1011
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampleId :


Tot Ion:	127	Resp:	324670
Ion	Ratio	Lower	Upper
127	100		
129	32.0	26.2	39.2
65	34.3	27.8	41.8
92	19.2	16.8	25.2



#34
Hexachlorobutadiene
Concen: 45.41 ng
RT: 8.12 min Scan# 1025
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:	225	Resp:	154208
Ion	Ratio	Lower	Upper
225	100		
223	64.1	48.8	73.2
227	63.3	51.1	76.7









#35
Caprolactam
Concen: 33.44 ng
RT: 8.42 min Scan# 1076
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampleId :


Tot Ion:113 Resp: 80933
Ion Ratio Lower Upper
113 100
55 231.7 150.2 190.2#
56 159.5 107.8 147.8#



#36
4-Chloro-3-methylphenol
Concen: 45.14 ng
RT: 8.53 min Scan# 1095
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:107 Resp: 350570
Ion Ratio Lower Upper
107 100
144 26.5 24.2 36.4
142 81.3 73.4 110.0









#37
2-Methylnaphthalene
Concen: 46.94 ng
RT: 8.68 min Scan# 1121
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampleId :

Tot Ion:142 Resp: 729363
Ion Ratio Lower Upper
142 100
141 89.5 71.3 106.9
115 31.1 27.1 40.7




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E






#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.72 min Scan# 1297
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:164 Resp: 228515
Ion Ratio Lower Upper
164 100
162 104.3 82.6 123.8
160 46.3 36.2 54.2



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E





#39

1,2,4,5-Tetrachlorobenzene

Concen: 41.13 ng

RT: 8.85 min Scan# 1150

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :

Tot Ion:216 Resp: 244061



Ion Ratio Lower Upper

216 100

214 78.7 63.4 95.2

179 22.5 17.9 26.9

108 22.6 16.5 24.7





Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 85.91 ng

RT: 8.84 min Scan# 1148

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30



Tgt Ion:237 Resp: 247866

Ion Ratio Lower Upper

237 100

235 63.3 42.6 82.6


272 12.6 0.0 31.9




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 139.82 ng
 RT: 10.51 min Scan# 1432
 Delta R.T. -0.01 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:330 Resp: 249578
 Ion Ratio Lower Upper
 330 100
 332 96.7 76.6 115.0
 141 38.8 31.4 47.2




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E





#42
 2,4,6-Trichlorophenol
 Concen: 43.62 ng
 RT: 8.96 min Scan# 1169
 Delta R.T. -0.01 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion:196 Resp: 204753
 Ion Ratio Lower Upper
 196 100
 198 93.9 74.2 111.4
 200 30.6 24.0 36.0



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#43
2,4,5-Trichlorophenol
Concen: 44.16 ng
RT: 9.00 min Scan# 1176
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
196	100		
198	96.2	75.7	113.5
97	54.5	47.7	71.5
132	33.3	28.0	42.0






Abundance

Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E


Ion 97.00 (96.70 to 97.70): BF0

Ion 132.00 (131.70 to 132.70): E



#44
2-Fluorobiphenyl
Concen: 94.71 ng
RT: 9.05 min Scan# 1183
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.4	29.6	44.4
170	24.3	19.8	29.8






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 41.83 ng

RT: 9.15 min Scan# 1200

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :


Tot Ion:154 Resp: 818956

Ion Ratio Lower Upper

154 100

153 40.3 21.2 61.2




76 12.9 0.0 29.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 43.86 ng

RT: 9.17 min Scan# 1204

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Tgt Ion:162 Resp: 639951

Ion Ratio Lower Upper

162 100

127 42.3 28.3 42.5



164 33.5 27.0 40.4




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 47.62 ng
RT: 9.26 min Scan# 1220
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampleId :

Tot Ion: 65 Resp: 252815
Ion Ratio Lower Upper
65 100
92 64.8 53.9 80.9
138 108.1 78.8 118.2




Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): E






#48
Acenaphthylene
Concen: 45.18 ng
RT: 9.58 min Scan# 1274
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion: 152 Resp: 1044680
Ion Ratio Lower Upper
152 100
151 20.6 16.8 25.2
153 13.3 10.8 16.2



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 53.55 ng
RT: 9.45 min Scan# 1252
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
163	100		
194	3.1	2.6	4.0
164	10.3	8.4	12.6




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 45.95 ng
RT: 9.51 min Scan# 1262
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion	Ratio	Lower	Upper
165	100		
63	51.5	34.3	51.5
89	43.0	37.1	55.7



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 40.81 ng

RT: 9.75 min Scan# 1303

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :


Tot Ion:154 Resp: 581678

Ion Ratio Lower Upper

154 100

153 114.5 90.5 135.7




152 54.4 43.6 65.4



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 35.24 ng

RT: 9.67 min Scan# 1289

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Tgt Ion:138 Resp: 165818

Ion Ratio Lower Upper

138 100

108 10.0 9.1 13.7



92 106.9 93.8 140.6




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 66.45 ng
RT: 9.78 min Scan# 1308
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampleId :


Tot Ion:184 Resp: 133135
Ion Ratio Lower Upper
184 100
63 81.3 62.7 94.1
154 58.3 81.1 121.7#



#54
Dibenzofuran
Concen: 46.18 ng
RT: 9.93 min Scan# 1333
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:168 Resp: 868965
Ion Ratio Lower Upper
168 100
139 38.0 30.6 45.8
169 13.2 10.8 16.2





#55

4-Nitrophenol

Concen: 80.00 ng

RT: 9.85 min Scan# 1319


Delta R.T. -0.01 min

Lab File: BF096610.D




Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampleId :

Tot Ion:139 Resp: 311970
Ion Ratio Lower Upper
139 100
109 54.8 34.1 74.1
65 100.9 31.4 71.4#



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0



#56

2,4-Dinitrotoluene

Concen: 47.87 ng

RT: 9.91 min Scan# 1330




Delta R.T. -0.02 min


Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Tgt Ion:165 Resp: 228665
Ion Ratio Lower Upper
165 100
63 52.6 25.4 38.0#
89 67.5 46.4 69.6
182 2.1 1.8 2.6

Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0
Ion 182.00 (181.70 to 182.70): E









#57
 Fluorene
 Concen: 47.89 ng
 RT: 10.27 min Scan# 1391
 Delta R.T. -0.01 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:166 Resp: 636432
 Ion Ratio Lower Upper
 166 100
 165 99.1 78.8 118.2
 167 13.6 10.6 16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 51.38 ng
 RT: 10.05 min Scan# 1353
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion:232 Resp: 164984
 Ion Ratio Lower Upper
 232 100
 131 54.0 44.8 67.2
 130 2.5 2.3 3.5
 166 34.3 29.0 43.4









#59
Diethylphthalate
Concen: 45.61 ng
RT: 10.15 min Scan# 1371
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampled :

Tot Ion:149 Resp: 735489
Ion Ratio Lower Upper
149 100
177 21.3 16.7 25.1
150 12.4 10.2 15.2




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 46.26 ng
RT: 10.26 min Scan# 1389
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:204 Resp: 267469
Ion Ratio Lower Upper
204 100
206 32.0 26.2 39.2
141 75.0 60.8 91.2



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 50.28 ng
RT: 10.29 min Scan# 1394
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampled :


Tot Ion:138 Resp: 215091
Ion Ratio Lower Upper
138 100
92 47.9 21.8 61.8
108 54.7 42.3 82.3



#62
Azobenzene
Concen: 47.68 ng
RT: 10.42 min Scan# 1417
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion: 77 Resp: 743331
Ion Ratio Lower Upper
77 100
182 24.3 0.1 40.1
105 25.5 3.6 43.6
51 36.3 9.4 49.4









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.20 min Scan# 1549
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampleId :

Tot Ion:188 Resp: 380684
Ion Ratio Lower Upper
188 100
94 10.9 9.4 14.2
80 10.8 10.2 15.2




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0






#64
4,6-Dinitro-2-methylphenol
Concen: 45.35 ng
RT: 10.32 min Scan# 1400
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:198 Resp: 119064
Ion Ratio Lower Upper
198 100
51 60.8 53.2 93.2
105 47.7 45.8 85.8



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF0
Ion 105.00 (104.70 to 105.70): E









#65
 n-Nitrosodiphenylamine
 Concen: 46.77 ng
 RT: 10.38 min Scan# 1410
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:169 Resp: 624726
 Ion Ratio Lower Upper
 169 100
 168 68.7 53.2 79.8
 167 38.2 29.5 44.3



#66
 4-Bromophenyl-phenylether
 Concen: 45.89 ng
 RT: 10.75 min Scan# 1472
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion:248 Resp: 170048
 Ion Ratio Lower Upper
 248 100
 250 98.7 76.8 115.2
 141 93.9 68.6 103.0









#67
Hexachlorobenzene
Concen: 44.73 ng
RT: 10.82 min Scan# 1484
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Instrument :
BNA_F
ClientSampled :

Tot Ion:284 Resp: 181445
Ion Ratio Lower Upper
284 100
142 39.0 22.8 34.2#
249 28.8 24.0 36.0




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 44.41 ng
RT: 10.91 min Scan# 1500
Delta R.T. -0.01 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:200 Resp: 169456
Ion Ratio Lower Upper
200 100
173 31.1 6.3 46.3
215 44.4 27.2 67.2



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 79.83 ng
 RT: 11.01 min Scan# 1517
 Delta R.T. -0.01 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:266 Resp: 191923
 Ion Ratio Lower Upper
 266 100
 268 61.0 51.1 76.7
 264 61.2 51.7 77.5




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 46.05 ng
 RT: 11.22 min Scan# 1553
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion:178 Resp: 947677
 Ion Ratio Lower Upper
 178 100
 176 19.9 16.2 24.4
 179 15.4 12.6 19.0



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 46.46 ng

RT: 11.27 min Scan# 1562

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Instrument :

BNA_F

ClientSampleId :

Tot Ion:178 Resp: 974249



Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.8	23.8
179	15.2	12.7	19.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 44.75 ng

RT: 11.43 min Scan# 1588



Delta R.T. -0.02 min


Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Tgt Ion:167 Resp: 943681

Ion	Ratio	Lower	Upper
167	100		
166	22.5	18.1	27.1
139	14.0	11.7	17.5





#73

Di-n-butylphthalate

Concen: 43.84 ng

RT: 11.77 min Scan# 1646

Delta R.T. -0.01 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 1119782

Ion Ratio Lower Upper

149 100

150 9.4 7.7 11.5




104 4.2 4.0 6.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 48.43 ng

RT: 12.40 min Scan# 1754

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Tgt Ion:202 Resp: 977231

Ion Ratio Lower Upper

202 100

101 15.0 0.0 33.2



203 17.3 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.83 min Scan# 1996

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :


Tot Ion:240 Resp: 230497

Ion Ratio Lower Upper

240 100

120 14.0 5.8 8.8#




236 25.7 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 48.34 ng

RT: 12.53 min Scan# 1775

Delta R.T. -0.01 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Tgt Ion:184 Resp: 534653

Ion Ratio Lower Upper

184 100

185 13.9 15.5 23.3#



183 12.0 9.8 14.6




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
Pvrene
Concen: 52.98 ng
RT: 12.63 min Scan# 1793
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampleId :


Tot Ion:202 Resp: 980190
Ion Ratio Lower Upper
202 100
200 21.2 17.6 26.4
203 17.7 14.4 21.6



#78
Terphenyl-d14
Concen: 91.55 ng
RT: 12.78 min Scan# 1818
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:244 Resp: 987533
Ion Ratio Lower Upper
244 100
212 7.1 6.0 9.0
122 15.2 10.3 15.5









#79
Butylbenzylphthalate
Concen: 51.38 ng
RT: 13.26 min Scan# 1899
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
149	100		
91	72.9	45.0	67.4#
206	17.1	17.0	25.6



#80
Benzo(a)anthracene
Concen: 48.69 ng
RT: 13.82 min Scan# 1994
Delta R.T. -0.02 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.6	33.8
229	20.2	16.3	24.5









#81
 3,3'-Dichlorobenzidine
 Concen: 32.15 ng
 RT: 13.78 min Scan# 1988
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:252 Resp: 176241
 Ion Ratio Lower Upper
 252 100
 254 64.1 51.5 77.3
 126 12.8 15.8 23.8#




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#82
 Chrysene
 Concen: 45.51 ng
 RT: 13.85 min Scan# 2000
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion:228 Resp: 636127
 Ion Ratio Lower Upper
 228 100
 226 31.4 24.9 37.3
 229 19.8 15.8 23.6



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 46.73 ng
 RT: 13.82 min Scan# 1995
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:149 Resp: 488502
 Ion Ratio Lower Upper
 149 100
 167 26.1 21.0 31.4
 279 2.7 1.9 2.9



#84
 Di-n-octyl phthalate
 Concen: 42.61 ng
 RT: 14.43 min Scan# 2098
 Delta R.T. -0.02 min
 Lab File: BF096610.D
 Acq: 10 Jul 2017 19:30

Tgt Ion:149 Resp: 877275
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.2 1.8
 43 13.0 8.5 12.7#









#85
Indeno(1,2,3-cd)pyrene
Concen: 54.35 ng
RT: 16.56 min Scan# 2460
Delta R.T. -0.05 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
276	100		
138	36.9	27.8	41.6
277	24.9	20.2	30.4



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.22 min Scan# 2232
Delta R.T. -0.03 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.5	29.3
265	22.2	17.2	25.8





#87

Benzo(b)fluoranthene

Concen: 42.33 ng

RT: 14.83 min Scan# 2166

Delta R.T. -0.02 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :


Tot Ion:252 Resp: 551636

Ion Ratio Lower Upper

252 100

253 22.4 18.1 27.1




125 13.6 9.8 14.8



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 47.32 ng

RT: 14.83 min Scan# 2166

Delta R.T. -0.05 min

Lab File: BF096610.D

Acq: 10 Jul 2017 19:30


Tgt Ion:252 Resp: 551636

Ion Ratio Lower Upper

252 100

253 22.4 18.4 27.6



125 13.6 13.1 19.7




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(a)pyrene
Concen: 44.90 ng
RT: 15.16 min Scan# 2223
Delta R.T. -0.03 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30




Instrument :
BNA_F
ClientSampleId :


Tot Ion:252 Resp: 522419
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 15.2 11.0 16.4



#90
Dibenzo(a,h)anthracene
Concen: 52.34 ng
RT: 16.57 min Scan# 2463
Delta R.T. -0.04 min
Lab File: BF096610.D
Acq: 10 Jul 2017 19:30

Tgt Ion:278 Resp: 479015
Ion Ratio Lower Upper
278 100
139 25.0 16.6 24.8#
279 23.4 19.3 28.9





#91

Benzo(a,h,i)perylene

Concen: 53.28 ng

RT: 16.96 min Scan# 2529

Delta R.T. -0.04 min


Lab File: BF096610.D

Acq: 10 Jul 2017 19:30

Instrument :

BNA_F

ClientSampleId :




Tot Ion: 276 Resp: 505353

Ion Ratio Lower Upper

276 100

277 23.0 18.6 28.0

138 29.7 25.4 38.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

