

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071019\
 Data File : BF115451.D
 Acq On : 10 Jul 2019 13:05
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 10 16:51:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 10 16:47:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.635	0.593	6.6	105	0.00
3	Pyridine	1.756	1.624	7.5	105	0.00
4	n-Nitrosodimethylamine	0.712	0.598	16.0	93	0.00
5 S	2-Fluorophenol	1.294	1.196	7.6	102	0.00
6	Aniline	2.323	2.163	6.9	101	0.00
7 S	Phenol-d6	1.646	1.533	6.9	102	0.00
8	2-Chlorophenol	1.462	1.402	4.1	104	0.00
9	Benzaldehyde	1.049	0.978	6.8	112	0.00
10 C	Phenol	1.970	1.821	7.6	101	0.00
11	bis(2-Chloroethyl)ether	1.461	1.349	7.7	101	0.00
12	1,3-Dichlorobenzene	1.601	1.513	5.5	103	0.00
13 C	1,4-Dichlorobenzene	1.604	1.531	4.6	105	0.00
14	1,2-Dichlorobenzene	1.485	1.414	4.8	104	0.00
15	Benzyl Alcohol	1.321	1.226	7.2	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.043	1.814	11.2	98	0.00
17	2-Methylphenol	1.252	1.165	6.9	102	0.00
18	Hexachloroethane	0.598	0.568	5.0	105	0.00
19 P	n-Nitroso-di-n-propylamine	1.114	0.978	12.2	97	0.00
20	3+4-Methylphenols	1.479	1.381	6.6	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
22	Acetophenone	0.479	0.442	7.7	98	0.00
23 S	Nitrobenzene-d5	0.339	0.335	1.2	105	0.00
24	Nitrobenzene	0.382	0.373	2.4	103	0.00
25	Isophorone	0.734	0.668	9.0	99	0.00
26 C	2-Nitrophenol	0.157	0.183	-16.6	122	0.00
27	2,4-Dimethylphenol	0.300	0.267	11.0	96	0.00
28	bis(2-Chloroethoxy)methane	0.452	0.417	7.7	100	0.00
29 C	2,4-Dichlorophenol	0.298	0.287	3.7	103	0.00
30	1,2,4-Trichlorobenzene	0.332	0.325	2.1	105	0.00
31	Naphthalene	0.993	0.950	4.3	102	0.00
32	Benzoic acid	0.205	0.186	9.3	92	0.00
33	4-Chloroaniline	0.443	0.427	3.6	103	0.00
34 C	Hexachlorobutadiene	0.188	0.188	0.0	106	0.00
35	Caprolactam	0.097	0.092	5.2	99	0.00
36 C	4-Chloro-3-methylphenol	0.316	0.302	4.4	101	0.00
37	2-Methylnaphthalene	0.664	0.630	5.1	100	0.00
38	1-Methylnaphthalene	0.634	0.602	5.0	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	0.575	0.583	-1.4	105	0.00
41 P	Hexachlorocyclopentadiene	0.333	0.325	2.4	102	0.00
42 S	2,4,6-Tribromophenol	0.221	0.226	-2.3	103	0.00
43 C	2,4,6-Trichlorophenol	0.419	0.424	-1.2	104	0.00
44	2,4,5-Trichlorophenol	0.434	0.438	-0.9	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071019\
 Data File : BF115451.D
 Acq On : 10 Jul 2019 13:05
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 10 16:51:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 10 16:47:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.140	1.157	-1.5	100	0.00
46	1,1'-Biphenyl	1.572	1.531	2.6	101	0.00
47	2-Chloronaphthalene	1.299	1.277	1.7	102	0.00
48	2-Nitroaniline	0.382	0.395	-3.4	103	0.00
49	Acenaphthylene	1.967	1.905	3.2	100	0.00
50	Dimethylphthalate	1.501	1.470	2.1	101	0.00
51	2,6-Dinitrotoluene	0.326	0.335	-2.8	104	0.00
52 C	Acenaphthene	1.238	1.205	2.7	100	0.00
53	3-Nitroaniline	0.378	0.389	-2.9	103	0.00
54 P	2,4-Dinitrophenol	0.114	0.143	-25.4#	130	0.00
55	Dibenzofuran	1.744	1.684	3.4	100	0.00
56 P	4-Nitrophenol	0.293	0.308	-5.1	104	0.00
57	2,4-Dinitrotoluene	0.414	0.446	-7.7	107	0.00
58	Fluorene	1.375	1.241	9.7	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.375	0.376	-0.3	101	0.00
60	Diethylphthalate	1.478	1.402	5.1	96	0.00
61	4-Chlorophenyl-phenylether	0.671	0.639	4.8	101	0.00
62	4-Nitroaniline	0.371	0.386	-4.0	102	0.00
63	Azobenzene	1.523	1.390	8.7	93	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	0.087	0.109	-25.3#	123	0.00
66 c	n-Nitrosodiphenylamine	0.630	0.612	2.9	98	0.00
67	4-Bromophenyl-phenylether	0.224	0.219	2.2	100	0.00
68	Hexachlorobenzene	0.252	0.253	-0.4	102	0.00
69	Atrazine	0.195	0.193	1.0	96	0.00
70 C	Pentachlorophenol	0.153	0.154	-0.7	98	0.00
71	Phenanthrene	1.052	0.999	5.0	96	0.00
72	Anthracene	1.055	1.023	3.0	98	0.00
73	Carbazole	0.965	0.924	4.2	96	0.00
74	Di-n-butylphthalate	1.166	1.094	6.2	94	0.00
75 C	Fluoranthene	1.108	1.058	4.5	95	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
77	Benzidine	0.780	0.678	13.1	82	0.00
78	Pyrene	1.555	1.558	-0.2	96	0.00
79 S	Terphenyl-d14	0.977	0.966	1.1	96	0.00
80	Butylbenzylphthalate	0.680	0.674	0.9	93	0.00
81	Benzo(a)anthracene	1.281	1.269	0.9	94	0.00
82	3,3'-Dichlorobenzidine	0.465	0.477	-2.6	91	0.00
83	Chrysene	1.317	1.291	2.0	92	0.00
84	Bis(2-ethylhexyl)phthalate	0.842	0.822	2.4	94	0.00
85 c	Di-n-octyl phthalate	1.498	1.407	6.1	87	0.00
86	Indeno(1,2,3-cd)pyrene	1.140	1.117	2.0	94	0.00
87 I	Perylene-d12	1.000	1.000	0.0	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071019\
Data File : BF115451.D
Acq On : 10 Jul 2019 13:05
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 10 16:51:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 10 16:47:05 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.302	1.226	5.8	83	0.00
89	Benzo(k)fluoranthene	1.158	1.147	0.9	93	0.00
90 C	Benzo(a)pyrene	1.123	1.090	2.9	88	0.00
91	Dibenzo(a,h)anthracene	0.960	0.929	3.2	93	0.00
92	Benzo(g,h,i)perylene	0.908	0.903	0.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0