


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071019\
 Data File : BF115450.D
 Acq On : 10 Jul 2019 12:05
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070519.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 10 16:47:05 2019



AutoFind: Scans 1630, 1631, 1632; Background Corrected with Scan 1625

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.1	33929	PASS
68	69	0.00	2	1.7	629	PASS
69	198	0.00	100	32.7	36766	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.4	50007	PASS
197	198	0.00	2	0.3	348	PASS
198	198	100	100	100.0	112560	PASS
199	198	5	9	6.8	7677	PASS
275	198	10	60	28.4	32003	PASS
365	198	1	100	3.5	3954	PASS
441	198	0.01	100	15.3	17263	PASS
442	198	50	100	97.5	109765	PASS
443	442	15	24	19.0	20840	PASS