

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071020\  
 Data File : BF120837.D  
 Acq On : 10 Jul 2020 18:07  
 Operator : JU/CG  
 Sample : L3255-12  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 CHRT-31

Quant Time: Jul 10 18:58:37 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF070220.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jul 02 18:51:52 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.85  | 152  | 192189   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.13  | 136  | 679231   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.90  | 164  | 170137   | 20.00  | ng    | 0.01     |
| 64) Phenanthrene-d10        | 11.42 | 188  | 125530   | 20.00  | ng    | 0.05     |
| 76) Chrysene-d12            | 14.04 | 240  | 273064   | 20.00  | ng    | 0.03     |
| 86) Perylene-d12            | 15.47 | 264  | 388794   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.45  | 112  | 984460   | 79.07  | ng    | 0.00     |
| 7) Phenol-d6                | 6.47  | 99   | 1302909  | 81.95  | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.41  | 82   | 903327   | 77.05  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.70 | 330  | 124503   | 72.84  | ng    | 0.02     |
| 45) 2-Fluorobiphenyl        | 9.22  | 172  | 1218122  | 108.32 | ng    | 0.00     |
| 79) Terphenyl-d14           | 13.00 | 244  | 565164   | 40.52  | ng    | 0.04     |
| Target Compounds            |       |      |          |        |       |          |
| 46) 1,1'-Biphenyl           | 9.31  | 154  | 38642    | 2.816  | ng    | 94       |
| 50) Dimethylphthalate       | 9.60  | 163  | 117058   | 9.432  | ng    | # 62     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071020\  
Data File : BF120837.D  
Acq On : 10 Jul 2020 18:07  
Operator : JU/CG  
Sample : L3255-12  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
CHRT-31

Quant Time: Jul 10 18:58:37 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF070220.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Jul 02 18:51:52 2020  
Response via : Initial Calibration

