

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071117\
 Data File : BF096667.D
 Acq On : 12 Jul 2017 2:50
 Operator : SJ/JU
 Sample : I4130-03
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-P6-WSW

Manual Integrations
 APPROVED

mohammad
 7/12/2017 4:57:46 PM

Quant Time: Jul 12 04:44:31 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 07 13:02:33 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	68756	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	149623	20.00	ng	0.00
38) Acenaphthene-d10	9.76	164	50243	20.00	ng	0.03
63) Phenanthrene-d10	11.23	188	112260	20.00	ng	0.02
75) Chrysene-d12	13.84	240	139208	20.00	ng	0.00
86) Perylene-d12	15.24	264	177724	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.29	112	622282	133.58	ng	-0.02
7) Phenol-d6	6.33	99	700853	122.38	ng	-0.02
23) Nitrobenzene-d5	7.26	82	333122	133.79	ng	-0.01
41) 2,4,6-Tribromophenol	10.54	330	68674	174.98	ng	0.02
44) 2-Fluorobiphenyl	9.08	172	289273	98.44	ng	0.02
78) Terphenyl-d14	12.80	244	316961	48.65	ng	0.00
Target Compounds						
37) 2-Methylnaphthalene	8.73	142	199898	44.458	ng	# 84
70) Phenanthrene	11.25	178	55466	9.140	ng	# 67
74) Fluoranthene	12.44	202	32533	5.468	ng	# 80
77) Pyrene	12.66	202	35171	3.148	ng	# 84
80) Benzo(a)anthracene	13.84	228	16842m	2.089	ng	
82) Chrysene	13.87	228	20955	2.482	ng	# 59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071117\
Data File : BF096667.D
Acq On : 12 Jul 2017 2:50
Operator : SJ/JU
Sample : I4130-03
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-P6-WSW

Manual Integrations
APPROVED

mohammad
7/12/2017 4:57:46 PM

Quant Time: Jul 12 04:44:31 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 07 13:02:33 2017
Response via : Initial Calibration

