

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071118\
 Data File : BF107296.D
 Acq On : 11 Jul 2018 13:32
 Operator : JU/SJ
 Sample : J3933-08
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EP-8-S1

Manual Integrations
 APPROVED

Sohil
 7/12/2018 2:19:09 PM

Quant Time: Jul 11 15:33:06 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 29 14:38:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.93	152	73483	20.00	ng	-0.03
21) Naphthalene-d8	8.21	136	284490	20.00	ng	-0.04
38) Acenaphthene-d10	9.98	164	148550	20.00	ng	-0.03
63) Phenanthrene-d10	11.48	188	296535	20.00	ng	-0.02
75) Chrysene-d12	14.14	240	221020	20.00	ng	-0.03
86) Perylene-d12	15.68	264	237176	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	406080	93.58	ng	-0.03
7) Phenol-d6	6.58	99	518939	95.76	ng	-0.04
23) Nitrobenzene-d5	7.50	82	338980	66.22	ng	-0.04
41) 2,4,6-Tribromophenol	10.78	330	183439	106.54	ng	-0.02
44) 2-Fluorobiphenyl	9.29	172	567799	60.43	ng	-0.03
78) Terphenyl-d14	13.06	244	629549	69.00	ng	-0.02
Target Compounds						
10) Phenol	6.59	94	24806	4.011	ng	91
49) Dimethylphthalate	9.68	163	156717	14.066	ng	98
70) Phenanthrene	11.50	178	133371	9.325	ng	97
71) Anthracene	11.55	178	32587	2.232	ng	99
74) Fluoranthene	12.70	202	265803	16.861	ng	96
77) Pyrene	12.93	202	242009	16.605	ng	99
80) Benzo(a)anthracene	14.12	228	124084	9.211	ng	98
82) Chrysene	14.16	228	106883	8.625	ng	98
83) Bis(2-ethylhexyl)phthalate	14.08	149	150069	19.588	ng	96
85) Indeno(1,2,3-cd)pyrene	17.27	276	77932	7.410	ng	# 88
87) Benzo(b)fluoranthene	15.22	252	149931m	10.176	ng	
88) Benzo(k)fluoranthene	15.24	252	49753m	3.546	ng	
89) Benzo(a)pyrene	15.61	252	115071m	8.786	ng	
91) Benzo(a,h,i)perylene	17.75	276	80278	7.399	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071118\
Data File : BF107296.D
Acq On : 11 Jul 2018 13:32
Operator : JU/SJ
Sample : J3933-08
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EP-8-S1

Manual Integrations
APPROVED

Sohil
7/12/2018 2:19:09 PM

Quant Time: Jul 11 15:33:06 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 29 14:38:29 2018
Response via : Initial Calibration

