


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF071118\
 Data File : BF107294.D
 Acq On : 11 Jul 2018 11:42
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF061918.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 11 16:39:37 2018



AutoFind: Scans 1592, 1593, 1594; Background Corrected with Scan 1587

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.2	15759	PASS
68	69	0.00	2	0.7	105	PASS
69	198	0.00	100	26.5	14779	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	35.4	19759	PASS
197	198	0.00	2	0.5	306	PASS
198	198	100	100	100.0	55816	PASS
199	198	5	9	7.1	3955	PASS
275	198	10	60	27.2	15173	PASS
365	198	1	100	3.3	1861	PASS
441	198	0.01	100	15.0	8382	PASS
442	198	50	100	96.0	53597	PASS
443	442	15	24	19.2	10267	PASS