

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071216\
 Data File : BF088951.D
 Acq On : 13 Jul 2016 8:17
 Operator : UM/SJ
 Sample : H3936-09
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-6(11-12)

Quant Time: Jul 13 08:40:02 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF063016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 11 18:13:32 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	150257	20.00	ng	0.15
21) Naphthalene-d8	0.00	136	0	0.00	ng	-8.00
38) Acenaphthene-d10	0.00	164	0	0.00	ng	-9.75
63) Phenanthrene-d10	11.27	188	667	20.00	ng	0.05
75) Chrysene-d12	0.00	240	0	0.00	ng	-13.85
86) Perylene-d12	0.00	264	0	0.00	ng	-15.21
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	378657	37.77	ng	0.01
7) Phenol-d6	6.35	99	488741	37.73	ng	0.00
23) Nitrobenzene-d5	7.27	82	266358	0.00	ng	-0.01
41) 2,4,6-Tribromophenol	10.52	330	73690	0.00	ng	-0.01
44) 2-Fluorobiphenyl	9.07	172	419340	0.00	ng	0.00
78) Terphenyl-d14	12.80	244	297177	0.00	ng	0.00
Target Compounds						
19) n-Nitroso-di-n-propylamine	7.27	70	36226	4.16	ng	# 74
65) n-Nitrosodiphenylamine	10.52	169	2602	115.26	ng	# 43
70) Phenanthrene	11.23	178	717	19.66	ng	# 59
71) Anthracene	11.23	178	717	19.78	ng	# 59
73) Di-n-butylphthalate	11.78	149	915	23.05	ng	# 77
74) Fluoranthene	12.51	202	232	6.28	ng	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071216\
Data File : BF088951.D
Acq On : 13 Jul 2016 8:17
Operator : UM/SJ
Sample : H3936-09
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-6(11-12)

Quant Time: Jul 13 08:40:02 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF063016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 11 18:13:32 2016
Response via : Initial Calibration

