



Data Path : Z:\HPCHEM1\BNA F\DATA\BF071217\
Data File : BF096713.D
Acq On : 13 Jul 2017 2:09
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 13 06:12:42 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF070117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 07 13:02:33 2017
Response via : Initial Calibration



TIC: BF096713.D

(76) Benzidine

12.504min (-0.035) 0.00ng

response 294053

Ion	Exp%	Act%
184.00	100	100
185.00	19.40	14.68#
183.00	12.20	11.95
0.00	0.00	0.00