

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF071217\  
 Data File : BF096695.D  
 Acq On : 12 Jul 2017 16:51  
 Operator : SJ/JU  
 Sample : I4126-03  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PP-WW-1

Quant Time: Jul 13 07:26:35 2017  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF070117.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jul 07 13:02:33 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.67  | 152  | 125556   | 20.00 | ng    | -0.03     |
| 21) Naphthalene-d8          | 7.95  | 136  | 491205   | 20.00 | ng    | -0.04     |
| 38) Acenaphthene-d10        | 9.70  | 164  | 203165   | 20.00 | ng    | -0.04     |
| 63) Phenanthrene-d10        | 11.17 | 188  | 300341   | 20.00 | ng    | -0.04     |
| 75) Chrysene-d12            | 13.80 | 240  | 200982   | 20.00 | ng    | -0.04     |
| 86) Perylene-d12            | 15.20 | 264  | 175841   | 20.00 | ng    | -0.05     |
| System Monitoring Compounds |       |      |          |       |       |           |
| 5) 2-Fluorophenol           | 5.28  | 112  | 690874   | 81.22 | ng    | -0.03     |
| 7) Phenol-d6                | 6.31  | 99   | 879389   | 84.09 | ng    | -0.04     |
| 23) Nitrobenzene-d5         | 7.23  | 82   | 505517   | 61.84 | ng    | -0.04     |
| 41) 2,4,6-Tribromophenol    | 10.49 | 330  | 145008   | 91.37 | ng    | -0.04     |
| 44) 2-Fluorobiphenyl        | 9.03  | 172  | 885819   | 74.55 | ng    | -0.04     |
| 78) Terphenyl-d14           | 12.76 | 244  | 460379   | 48.95 | ng    | -0.04     |
| Target Compounds            |       |      |          |       |       |           |
| 49) Dimethylphthalate       | 9.42  | 163  | 85724    | 5.652 | ng    | Qvalue 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF071217\  
Data File : BF096695.D  
Acq On : 12 Jul 2017 16:51  
Operator : SJ/JU  
Sample : I4126-03  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PP-WW-1

Quant Time: Jul 13 07:26:35 2017  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF070117.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Jul 07 13:02:33 2017  
Response via : Initial Calibration

