

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071217\
 Data File : BF096687.D
 Acq On : 12 Jul 2017 13:02
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Jul 12 15:53:02 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 07 13:02:33 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.67	152	141099	20.00	ng	-0.03
21) Naphthalene-d8	7.95	136	509663	20.00	ng	-0.04
38) Acenaphthene-d10	9.70	164	222768	20.00	ng	-0.03
63) Phenanthrene-d10	11.18	188	430479	20.00	ng	-0.03
75) Chrysene-d12	13.82	240	283650	20.00	ng	-0.03
86) Perylene-d12	15.20	264	261493	20.00	ng	-0.04

System Monitoring Compounds

5) 2-Fluorophenol	5.27	112	594590	62.20	ng	-0.04
7) Phenol-d6	6.32	99	876457	74.58	ng	-0.03
23) Nitrobenzene-d5	7.24	82	771232	90.93	ng	-0.03
41) 2,4,6-Tribromophenol	10.49	330	153058	87.96	ng	-0.03
44) 2-Fluorobiphenyl	9.03	172	1186320	91.05	ng	-0.03
78) Terphenyl-d14	12.77	244	1045513	78.76	ng	-0.03

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.32	88	164307	36.247	ng	# 75
3) Pyridine	3.00	79	431531	34.696	ng	# 61
4) n-Nitrosodimethylamine	2.95	42	217745	35.469	ng	85
6) Aniline	6.33	93	547053	36.015	ng	# 59
8) 2-Chlorophenol	6.46	128	392432	38.495	ng	94
9) Benzaldehyde	6.22	77	252696	34.356	ng	100
10) Phenol	6.33	94	482722	36.054	ng	# 70
11) bis(2-Chloroethyl)ether	6.41	93	390854	37.775	ng	91
12) 1,3-Dichlorobenzene	6.61	146	410469	38.401	ng	98
13) 1,4-Dichlorobenzene	6.69	146	418402	39.162	ng	94
14) 1,2-Dichlorobenzene	6.85	146	380921	38.831	ng	96
15) Benzyl Alcohol	6.82	79	298671	38.018	ng	99
16) 2,2'-oxybis(1-Chloropropan	6.95	45	828344	39.380	ng	91
17) 2-Methylphenol	6.93	107	310247	38.187	ng	99
18) Hexachloroethane	7.18	117	151396	39.087	ng	# 77
19) n-Nitroso-di-n-propylamine	7.10	70	263452	37.628	ng	# 85
20) 3+4-Methylphenols	7.09	107	358816	39.597	ng	# 78
22) Acetophenone	7.09	105	510956	45.882	ng	# 94
24) Nitrobenzene	7.26	77	405951	45.624	ng	91
25) Isophorone	7.50	82	714480	43.447	ng	98
26) 2-Nitrophenol	7.57	139	218099	46.311	ng	# 85
27) 2,4-Dimethylphenol	7.62	122	345066	43.031	ng	97
28) bis(2-Chloroethoxy)methane	7.71	93	396605	36.535	ng	100
29) 2,4-Dichlorophenol	7.82	162	274328	39.651	ng	96
30) 1,2,4-Trichlorobenzene	7.90	180	262902	40.283	ng	98
31) Naphthalene	7.97	128	942974	39.191	ng	100
32) Benzoic acid	7.76	122	232303	34.493	ng	88
33) 4-Chloroaniline	8.03	127	387834	38.806	ng	97
34) Hexachlorobutadiene	8.10	225	138425	41.353	ng	98
35) Caprolactam	8.41	113	90053	37.746	ng	# 55
36) 4-Chloro-3-methylphenol	8.52	107	334977	43.756	ng	86
37) 2-Methylnaphthalene	8.67	142	681019	44.465	ng	98
39) 1,2,4,5-Tetrachlorobenzene	8.83	216	264120	45.662	ng	99
40) Hexachlorocyclopentadiene	8.83	237	98540	35.037	ng	98


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071217\
 Data File : BF096687.D
 Acq On : 12 Jul 2017 13:02
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Jul 12 15:53:02 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 07 13:02:33 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	8.95	196	201351	44.004	ng	97
43) 2,4,5-Trichlorophenol	8.99	196	203637	45.010	ng	93
45) 1,1'-Biphenyl	9.13	154	773109	40.504	ng	98
46) 2-Chloronaphthalene	9.16	162	572083	40.218	ng	95
47) 2-Nitroaniline	9.25	65	201703	38.970	ng	91
48) Acenaphthylene	9.57	152	882782	39.166	ng	99
49) Dimethylphthalate	9.44	163	704396	42.357	ng	99
50) 2,6-Dinitrotoluene	9.50	165	167060	45.428	ng #	91
51) Acenaphthene	9.74	154	500341	36.013	ng	98
52) 3-Nitroaniline	9.66	138	174432	38.030	ng	86
53) 2,4-Dinitrophenol	9.77	184	75743	38.778	ng #	75
54) Dibenzofuran	9.91	168	711232	38.774	ng	99
55) 4-Nitrophenol	9.83	139	133743	35.180	ng #	66
56) 2,4-Dinitrotoluene	9.90	165	187683	40.301	ng #	83
57) Fluorene	10.25	166	546897	42.215	ng	100
58) 2,3,4,6-Tetrachlorophenol	10.03	232	127854	40.846	ng	96
59) Diethylphthalate	10.13	149	622698	39.615	ng	99
60) 4-Chlorophenyl-phenylether	10.24	204	228696	40.574	ng	97
61) 4-Nitroaniline	10.27	138	169145	40.556	ng	90
62) Azobenzene	10.40	77	561026	36.912	ng	92
64) 4,6-Dinitro-2-methylphenol	10.31	198	105235	35.446	ng	85
65) n-Nitrosodiphenylamine	10.37	169	511446	33.861	ng	97
66) 4-Bromophenyl-phenylether	10.73	248	151646	36.188	ng	97
67) Hexachlorobenzene	10.80	284	169276	36.901	ng #	91
68) Atrazine	10.90	200	151087	35.013	ng	96
69) Pentachlorophenol	10.99	266	103670	38.131	ng	95
70) Phenanthrene	11.21	178	881047	37.861	ng	98
71) Anthracene	11.26	178	901377	38.011	ng	100
72) Carbazole	11.42	167	912289	38.254	ng	98
73) Di-n-butylphthalate	11.76	149	1117861	38.700	ng	98
74) Fluoranthene	12.39	202	882147	38.664	ng	98
76) Benzidine	12.51	184	466895	34.304	ng #	93
77) Pyrene	12.62	202	894710	39.297	ng	99
79) Butylbenzylphthalate	13.24	149	418216	36.286	ng #	84
80) Benzo(a)anthracene	13.80	228	646960	39.387	ng	99
81) 3,3'-Dichlorobenzidine	13.77	252	274850	40.740	ng	98
82) Chrysene	13.84	228	612005	35.579	ng	99
83) Bis(2-ethylhexyl)phthalate	13.81	149	527697	41.018	ng	97
84) Di-n-octyl phthalate	14.42	149	945466	37.319	ng #	95
85) Indeno(1,2,3-cd)pyrene	16.54	276	520067	38.303	ng	96
87) Benzo(b)fluoranthene	14.82	252	595337	36.334	ng	99
88) Benzo(k)fluoranthene	14.84	252	519860	35.469	ng	97
89) Benzo(a)pyrene	15.15	252	565795	38.677	ng	97
90) Dibenzo(a,h)anthracene	16.56	278	418352	36.352	ng #	94
91) Benzo(g,h,i)perylene	16.94	276	419135	35.147	ng	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.67 min Scan# 779
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:152 Resp: 141099
 Ion Ratio Lower Upper
 152 100
 150 160.4 126.2 189.2
 115 64.9 52.2 78.2



#2
 1,4-Dioxane
 Concen: 36.247 ng
 RT: 2.32 min Scan# 39
 Delta R.T. -0.05 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion: 88 Resp: 164307
 Ion Ratio Lower Upper
 88 100
 58 101.0 61.4 92.0#
 43 38.9 23.4 35.0#









#3
 Pyridine
 Concen: 34.696 ng
 RT: 3.00 min Scan# 155
 Delta R.T. -0.05 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 79 Resp: 431531
 Ion Ratio Lower Upper
 79 100
 52 104.5 54.8 82.2#
 51 49.3 27.0 40.4#



#4
 n-Nitrosodimethylamine
 Concen: 35.469 ng
 RT: 2.95 min Scan# 146
 Delta R.T. -0.06 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion: 42 Resp: 217745
 Ion Ratio Lower Upper
 42 100
 74 111.4 103.9 155.9
 44 8.0 5.7 8.5





#5

2-Fluorophenol

Concen: 62.197 ng

RT: 5.27 min Scan# 542

Delta R.T. -0.04 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 112 Resp: 594590

Ion Ratio Lower Upper

112 100

64 66.3 46.0 69.0




63 33.6 23.4 35.0



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 36.015 ng

RT: 6.33 min Scan# 722

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion: 93 Resp: 547053

Ion Ratio Lower Upper

93 100

66 65.3 32.9 49.3#



65 40.7 16.0 24.0#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 74.577 ng

RT: 6.32 min Scan# 719

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 876457

Ion Ratio Lower Upper

99 100

42 23.5 13.5 20.3#

71 31.4 24.5 36.7



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0

800000

600000

400000

200000



0

6.32

6.30

6.40

Time-->



#8

2-Chlorophenol

Concen: 38.495 ng

RT: 6.46 min Scan# 743

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion: 128 Resp: 392432

Ion Ratio Lower Upper

128 100

130 32.0 12.9 52.9

64 54.8 28.4 68.4



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0

600000

500000

400000

300000

200000

100000

0

6.46


6.40


6.45

6.50

6.55

Time-->





#9

Benzaldehyde

Concen: 34.356 ng

RT: 6.22 min Scan# 702

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 77 Resp: 252696

Ion Ratio Lower Upper

77 100

105 104.6 83.8 123.8




106 99.1 79.1 119.1



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 36.054 ng

RT: 6.33 min Scan# 721

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion: 94 Resp: 482722

Ion Ratio Lower Upper

94 100

65 43.5 9.9 49.9



66 64.1 23.1 63.1#




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0





#11

bis(2-Chloroethyl)ether

Concen: 37.775 ng

RT: 6.41 min Scan# 735

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 93 Resp: 390854

Ion Ratio Lower Upper

93 100

63 89.9 59.2 99.2

95 32.8 12.8 52.8



#12

1,3-Dichlorobenzene

Concen: 38.401 ng

RT: 6.61 min Scan# 769

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02




Tgt Ion: 146 Resp: 410469


Ion Ratio Lower Upper

146 100

148 63.5 51.8 77.6

75 29.8 23.1 34.7





#13

1,4-Dichlorobenzene

Concen: 39.162 ng

RT: 6.69 min Scan# 782

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:146 Resp: 418402

Ion Ratio Lower Upper

146 100

148 62.7 52.2 78.2




111 43.8 30.3 45.5



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 38.831 ng

RT: 6.85 min Scan# 809

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:146 Resp: 380921

Ion Ratio Lower Upper

146 100

148 64.7 50.4 75.6



111 43.0 38.2 57.4




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 38.018 ng

RT: 6.82 min Scan# 804

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 298671

Ion Ratio Lower Upper

79 100

108 80.4 63.4 95.0

77 60.4 49.2 73.8



#16

2,2'-oxybis(1-chloropropane)

Concen: 39.380 ng

RT: 6.95 min Scan# 827

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02




Tgt Ion: 45 Resp: 828344


Ion Ratio Lower Upper

45 100

77 9.2 0.0 33.2

79 7.4 0.0 30.0





#17

2-Methylphenol

Concen: 38.187 ng

RT: 6.93 min Scan# 824

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:107 Resp: 310247


Ion Ratio Lower Upper

107 100

108 115.5 93.4 140.0

77 43.5 35.8 53.6

79 43.9 34.8 52.2






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0



#18

Hexachloroethane

Concen: 39.087 ng

RT: 7.18 min Scan# 866

Delta R.T. -0.04 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:117 Resp: 151396

Ion Ratio Lower Upper

117 100

119 93.2 77.4 116.0



201 76.3 94.6 141.8#




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 37.628 ng

RT: 7.10 min Scan# 852

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 70 Resp: 263452


Ion Ratio Lower Upper

70 100

42 74.9 47.2 70.8#

101 9.5 7.2 10.8

130 19.7 15.9 23.9






Abundance Ion 70.00 (69.70 to 70.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#20

3+4-Methylphenols

Concen: 39.597 ng

RT: 7.09 min Scan# 851

Delta R.T. -0.02 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Tgt Ion: 107 Resp: 358816


Ion Ratio Lower Upper

107 100

108 94.4 71.0 111.0

77 64.8 90.5 130.5#

79 29.0 8.4 48.4






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 7.95 min Scan# 997
Delta R.T. -0.04 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:	136	Resp:	509663
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.5	12.7
54	10.8	4.9	7.3#
68	5.5	4.1	6.1




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0






#22
Acetophenone
Concen: 45.882 ng
RT: 7.09 min Scan# 850
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion:	105	Resp:	510956
Ion	Ratio	Lower	Upper
105	100		
71	4.5	2.8	4.2#
51	44.4	30.7	46.1
120	22.0	17.4	26.0



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BF0
Ion 51.00 (50.70 to 51.70): BF0
Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 90.931 ng
 RT: 7.24 min Scan# 876
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 82 Resp: 771232
 Ion Ratio Lower Upper
 82 100
 128 48.9 34.0 51.0
 54 66.1 42.2 63.2#



#24
 Nitrobenzene
 Concen: 45.624 ng
 RT: 7.26 min Scan# 879
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion: 77 Resp: 405951
 Ion Ratio Lower Upper
 77 100
 123 51.5 35.8 53.8
 65 15.2 10.6 15.8





#25

Isophorone

Concen: 43.447 ng

RT: 7.50 min Scan# 920

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 714480

Ion Ratio Lower Upper

82 100

95 6.2 5.3 7.9

138 17.7 13.1 19.7



#26

2-Nitrophenol

Concen: 46.311 ng

RT: 7.57 min Scan# 932

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02




Tgt Ion: 139 Resp: 218099


Ion Ratio Lower Upper

139 100

109 23.3 21.0 31.6

65 54.7 33.0 49.6#





#27

2,4-Dimethylphenol

Concen: 43.031 ng

RT: 7.62 min Scan# 940

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:122 Resp: 345066

Ion Ratio Lower Upper

122 100

107 107.1 89.2 133.8




121 57.5 45.2 67.8



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 36.535 ng

RT: 7.71 min Scan# 956

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion: 93 Resp: 396605

Ion Ratio Lower Upper

93 100

95 32.8 26.5 39.7



123 11.2 9.0 13.4




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 39.651 ng
 RT: 7.82 min Scan# 974
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
162	100		
164	63.4	44.6	84.6
98	39.5	14.8	54.8




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 40.283 ng
 RT: 7.90 min Scan# 988
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion	Ratio	Lower	Upper
180	100		
182	92.7	75.8	113.6
145	33.6	25.3	37.9



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 39.191 ng

RT: 7.97 min Scan# 1001

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:128 Resp: 942974

Ion Ratio Lower Upper

128 100

129 11.3 9.0 13.6




127 13.5 10.8 16.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 34.493 ng

RT: 7.76 min Scan# 965

Delta R.T. -0.02 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:122 Resp: 232303

Ion Ratio Lower Upper

122 100

105 111.6 103.3 143.3



77 79.4 73.7 113.7




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 38.806 ng
RT: 8.03 min Scan# 1010
Delta R.T. -0.02 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	387834
Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.2	39.2
65	31.5	27.8	41.8
92	19.3	16.8	25.2



#34
Hexachlorobutadiene
Concen: 41.353 ng
RT: 8.10 min Scan# 1022
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion:	225	Resp:	138425
Ion	Ratio	Lower	Upper
225	100		
223	62.6	48.8	73.2
227	65.3	51.1	76.7









#35
 Caprolactam
 Concen: 37.746 ng
 RT: 8.41 min Scan# 1075
 Delta R.T. -0.02 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:113 Resp: 90053
 Ion Ratio Lower Upper
 113 100
 55 237.6 150.2 190.2#
 56 174.6 107.8 147.8#



#36
 4-Chloro-3-methylphenol
 Concen: 43.756 ng
 RT: 8.52 min Scan# 1093
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:107 Resp: 334977
 Ion Ratio Lower Upper
 107 100
 144 25.7 24.2 36.4
 142 76.9 73.4 110.0









#37
2-Methylnaphthalene
Concen: 44.465 ng
RT: 8.67 min Scan# 1119
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 681019
Ion Ratio Lower Upper
142 100
141 90.4 71.3 106.9
115 32.0 27.1 40.7



#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.70 min Scan# 1295
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion:164 Resp: 222768
Ion Ratio Lower Upper
164 100
162 102.8 82.6 123.8
160 43.3 36.2 54.2





#39

1,2,4,5-Tetrachlorobenzene

Concen: 45.662 ng

RT: 8.83 min Scan# 1147

Delta R.T. -0.04 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:216 Resp: 264120



Ion Ratio Lower Upper

216 100

214 79.9 63.4 95.2

179 22.7 17.9 26.9

108 21.1 16.5 24.7





Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E

Time-->



#40

Hexachlorocyclopentadiene

Concen: 35.037 ng

RT: 8.83 min Scan# 1146

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02



Tgt Ion:237 Resp: 98540

Ion Ratio Lower Upper

237 100

235 64.0 42.6 82.6

272 12.5 0.0 31.9





Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E

Time-->









#41
 2,4,6-Tribromophenol
 Concen: 87.960 ng
 RT: 10.49 min Scan# 1429
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:330 Resp: 153058
 Ion Ratio Lower Upper
 330 100
 332 97.4 76.6 115.0
 141 37.6 31.4 47.2



#42
 2,4,6-Trichlorophenol
 Concen: 44.004 ng
 RT: 8.95 min Scan# 1167
 Delta R.T. -0.02 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:196 Resp: 201351
 Ion Ratio Lower Upper
 196 100
 198 96.3 74.2 111.4
 200 31.3 24.0 36.0









#43
 2,4,5-Trichlorophenol
 Concen: 45.010 ng
 RT: 8.99 min Scan# 1174
 Delta R.T. -0.02 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	196	Resp:	203637
Ion	Ratio	Lower	Upper
196	100		
198	98.8	75.7	113.5
97	50.9	47.7	71.5
132	32.5	28.0	42.0




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 91.052 ng
 RT: 9.03 min Scan# 1181
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:	172	Resp:	1186320
Ion	Ratio	Lower	Upper
172	100		
171	37.0	29.6	44.4
170	25.0	19.8	29.8



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 40.504 ng
 RT: 9.13 min Scan# 1198
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:154 Resp: 773109
 Ion Ratio Lower Upper
 154 100
 153 42.1 21.2 61.2
 76 12.3 0.0 29.9




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 40.218 ng
 RT: 9.16 min Scan# 1202
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:162 Resp: 572083
 Ion Ratio Lower Upper
 162 100
 127 40.2 28.3 42.5
 164 33.1 27.0 40.4



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 38.970 ng
RT: 9.25 min Scan# 1218
Delta R.T. -0.02 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 65 Resp: 201703
Ion Ratio Lower Upper
65 100
92 65.1 53.9 80.9
138 111.4 78.8 118.2




Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): BF0






#48
Acenaphthylene
Concen: 39.166 ng
RT: 9.57 min Scan# 1272
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion: 152 Resp: 882782
Ion Ratio Lower Upper
152 100
151 20.6 16.8 25.2
153 13.1 10.8 16.2



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 42.357 ng
RT: 9.44 min Scan# 1250
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.4	2.6	4.0
164	9.9	8.4	12.6




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 45.428 ng
RT: 9.50 min Scan# 1260
Delta R.T. -0.02 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion	Ratio	Lower	Upper
165	100		
63	53.1	34.3	51.5#
89	44.4	37.1	55.7



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 36.013 ng

RT: 9.74 min Scan# 1301

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 500341

Ion Ratio Lower Upper

154 100

153 116.1 90.5 135.7




152 55.5 43.6 65.4



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 38.030 ng

RT: 9.66 min Scan# 1288

Delta R.T. -0.02 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:138 Resp: 174432

Ion Ratio Lower Upper

138 100

108 9.4 9.1 13.7



92 100.9 93.8 140.6




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
 2,4-Dinitrophenol
 Concen: 38.778 ng
 RT: 9.77 min Scan# 1306
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:184 Resp: 75743
 Ion Ratio Lower Upper
 184 100
 63 76.8 62.7 94.1
 154 57.6 81.1 121.7#



#54
 Dibenzofuran
 Concen: 38.774 ng
 RT: 9.91 min Scan# 1330
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:168 Resp: 711232
 Ion Ratio Lower Upper
 168 100
 139 39.0 30.6 45.8
 169 13.8 10.8 16.2





#55

4-Nitrophenol

Concen: 35.180 ng

RT: 9.83 min Scan# 1316

Delta R.T. -0.02 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:139 Resp: 133743

Ion Ratio Lower Upper

139 100

109 54.4 34.1 74.1




65 99.4 31.4 71.4#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#56

2,4-Dinitrotoluene

Concen: 40.301 ng

RT: 9.90 min Scan# 1328

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Tgt Ion:165 Resp: 187683


Ion Ratio Lower Upper

165 100

63 50.0 25.4 38.0#

89 64.5 46.4 69.6

182 2.3 1.8 2.6






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E





#57

Fluorene

Concen: 42.215 ng

RT: 10.25 min Scan# 1388

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:166 Resp: 546897

Ion Ratio Lower Upper

166 100

165 98.9 78.8 118.2




167 13.6 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 40.846 ng

RT: 10.03 min Scan# 1351

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Tgt Ion:232 Resp: 127854


Ion Ratio Lower Upper

232 100

131 53.5 44.8 67.2

130 2.4 2.3 3.5

166 33.9 29.0 43.4






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#59
Diethylphthalate
Concen: 39.615 ng
RT: 10.13 min Scan# 1368
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 622698
Ion Ratio Lower Upper
149 100
177 20.5 16.7 25.1
150 12.4 10.2 15.2



#60
4-Chlorophenyl-phenylether
Concen: 40.574 ng
RT: 10.24 min Scan# 1387
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion:204 Resp: 228696
Ion Ratio Lower Upper
204 100
206 32.9 26.2 39.2
141 72.2 60.8 91.2





#61

4-Nitroaniline

Concen: 40.556 ng

RT: 10.27 min Scan# 1392

Delta R.T. -0.02 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:138 Resp: 169145

Ion Ratio Lower Upper

138 100

92 44.4 21.8 61.8




108 51.0 42.3 82.3



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 36.912 ng

RT: 10.40 min Scan# 1414

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Tgt Ion: 77 Resp: 561026


Ion Ratio Lower Upper

77 100

182 22.4 0.1 40.1

105 24.8 3.6 43.6

51 37.1 9.4 49.4






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.18 min Scan# 1546

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:188 Resp: 430479

Ion Ratio Lower Upper

188 100

94 13.2 9.4 14.2




80 13.5 10.2 15.2



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0



#64

4,6-Dinitro-2-methylphenol

Concen: 35.446 ng

RT: 10.31 min Scan# 1398

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:198 Resp: 105235

Ion Ratio Lower Upper

198 100

51 67.1 53.2 93.2



105 47.2 45.8 85.8




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF0

Ion 105.00 (104.70 to 105.70): E









#65
 n-Nitrosodiphenylamine
 Concen: 33.861 ng
 RT: 10.37 min Scan# 1408
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:169 Resp: 511446
 Ion Ratio Lower Upper
 169 100
 168 69.3 53.2 79.8
 167 38.2 29.5 44.3



#66
 4-Bromophenyl-phenylether
 Concen: 36.188 ng
 RT: 10.73 min Scan# 1470
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:248 Resp: 151646
 Ion Ratio Lower Upper
 248 100
 250 97.6 76.8 115.2
 141 89.8 68.6 103.0





#67

Hexachlorobenzene

Concen: 36.901 ng

RT: 10.80 min Scan# 1482

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:284 Resp: 169276

Ion Ratio Lower Upper

284 100

142 38.4 22.8 34.2#




249 30.1 24.0 36.0



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 35.013 ng

RT: 10.90 min Scan# 1498

Delta R.T. -0.02 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:200 Resp: 151087

Ion Ratio Lower Upper

200 100

173 30.3 6.3 46.3



215 45.7 27.2 67.2




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 38.131 ng
 RT: 10.99 min Scan# 1514
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:266 Resp: 103670
 Ion Ratio Lower Upper
 266 100
 268 59.6 51.1 76.7
 264 61.8 51.7 77.5




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 37.861 ng
 RT: 11.21 min Scan# 1551
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:178 Resp: 881047
 Ion Ratio Lower Upper
 178 100
 176 19.6 16.2 24.4
 179 15.1 12.6 19.0



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 38.011 ng
 RT: 11.26 min Scan# 1560
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:178 Resp: 901377
 Ion Ratio Lower Upper
 178 100
 176 20.0 15.8 23.8
 179 15.8 12.7 19.1



#72
 Carbazole
 Concen: 38.254 ng
 RT: 11.42 min Scan# 1586
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:167 Resp: 912289
 Ion Ratio Lower Upper
 167 100
 166 21.7 18.1 27.1
 139 13.6 11.7 17.5





#73

Di-n-butylphthalate

Concen: 38.700 ng

RT: 11.76 min Scan# 1644

Delta R.T. -0.02 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 1117861

Ion Ratio Lower Upper




149 100

150 9.0 7.7 11.5

104 4.3 4.0 6.0



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 38.664 ng

RT: 12.39 min Scan# 1752

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:202 Resp: 882147

Ion Ratio Lower Upper



202 100


101 14.6 0.0 33.2

203 17.4 0.0 38.1



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 13.82 min Scan# 1994

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:240 Resp: 283650

Ion Ratio Lower Upper

240 100

120 16.5 5.8 8.8#




236 25.2 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 34.304 ng

RT: 12.51 min Scan# 1772

Delta R.T. -0.03 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:184 Resp: 466895

Ion Ratio Lower Upper

184 100

185 14.9 15.5 23.3#



183 11.7 9.8 14.6




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
 Pyrene
 Concen: 39.297 ng
 RT: 12.62 min Scan# 1791
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion: 202 Resp: 894710
 Ion Ratio Lower Upper
 202 100
 200 22.4 17.6 26.4
 203 17.4 14.4 21.6




Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 200.00 (199.70 to 200.70): E
 Ion 203.00 (202.70 to 203.70): E






#78
 Terphenyl-d14
 Concen: 78.762 ng
 RT: 12.77 min Scan# 1816
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion: 244 Resp: 1045513
 Ion Ratio Lower Upper
 244 100
 212 6.8 6.0 9.0
 122 14.8 10.3 15.5



Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 36.286 ng
RT: 13.24 min Scan# 1897
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 418216
Ion Ratio Lower Upper
149 100
91 69.6 45.0 67.4#
206 17.5 17.0 25.6



#80
Benzo(a)anthracene
Concen: 39.387 ng
RT: 13.80 min Scan# 1992
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion:228 Resp: 646960
Ion Ratio Lower Upper
228 100
226 27.7 22.6 33.8
229 19.6 16.3 24.5









#81
 3,3'-Dichlorobenzidine
 Concen: 40.740 ng
 RT: 13.77 min Scan# 1986
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
252	100		
254	64.8	51.5	77.3
126	15.8	15.8	23.8




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#82
 Chrysene
 Concen: 35.579 ng
 RT: 13.84 min Scan# 1999
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.8	24.9	37.3
229	18.8	15.8	23.6



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 41.018 ng
 RT: 13.81 min Scan# 1993
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	149	Resp:	527697
Ion	Ratio	Lower	Upper
149	100		
167	24.8	21.0	31.4
279	2.6	1.9	2.9



#84
 Di-n-octyl phthalate
 Concen: 37.319 ng
 RT: 14.42 min Scan# 2096
 Delta R.T. -0.03 min
 Lab File: BF096687.D
 Acq: 12 Jul 2017 13:02

Tgt Ion:	149	Resp:	945466
Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	12.8	8.5	12.7#









#85
Indeno(1,2,3-cd)pyrene
Concen: 38.303 ng
RT: 16.54 min Scan# 2458
Delta R.T. -0.06 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
276	100		
138	38.1	27.8	41.6
277	24.8	20.2	30.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.20 min Scan# 2230
Delta R.T. -0.04 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.6	19.5	29.3
265	20.6	17.2	25.8









#87
Benzo(b)fluoranthene
Concen: 36.334 ng
RT: 14.82 min Scan# 2165
Delta R.T. -0.03 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 595337
Ion Ratio Lower Upper
252 100
253 22.0 18.1 27.1
125 12.9 9.8 14.8



#88
Benzo(k)fluoranthene
Concen: 35.469 ng
RT: 14.84 min Scan# 2169
Delta R.T. -0.04 min
Lab File: BF096687.D
Acq: 12 Jul 2017 13:02

Tgt Ion:252 Resp: 519860
Ion Ratio Lower Upper
252 100
253 22.2 18.4 27.6
125 14.2 13.1 19.7





#89

Benzo(a)pyrene

Concen: 38.677 ng

RT: 15.15 min Scan# 2221

Delta R.T. -0.04 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 565795

Ion Ratio Lower Upper

252 100

253 21.0 17.5 26.3




125 15.6 11.0 16.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 36.352 ng

RT: 16.56 min Scan# 2460

Delta R.T. -0.06 min

Lab File: BF096687.D

Acq: 12 Jul 2017 13:02


Tgt Ion:278 Resp: 418352

Ion Ratio Lower Upper

278 100

139 26.2 16.6 24.8#



279 24.4 19.3 28.9




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 35.147 ng

RT: 16.94 min Scan# 2526

Delta R.T. -0.06 min

Lab File: BF096687.D


Acq: 12 Jul 2017 13:02

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040




Tgt Ion: 276 Resp: 419135

Ion Ratio Lower Upper

276 100

277 23.7 18.6 28.0

138 34.9 25.4 38.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

