


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $2.17 \times 10^{-1} \times \text{Amt} - 5.40 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071317.M
Calibration Table Last Updated: Thu Jul 13 14:49:34 2017

2,4-Dinitrophenol

Response Ratio




Resp Ratio = $1.76 \times 10^{-1} \times \text{Amt} - 5.94 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071317.M
Calibration Table Last Updated: Thu Jul 13 14:49:34 2017


4-Chlorophenyl-phenylether

Response Ratio



$R = -5.71e-002 A^2 + 6.28e-001 A + 2.24e-002$
Coef of Det (r^2) = 0.995 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071317.M
Calibration Table Last Updated: Thu Jul 13 14:49:34 2017



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071317.M
Calibration Table Last Updated: Thu Jul 13 14:49:34 2017