

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071318\
 Data File : BF107377.D
 Acq On : 13 Jul 2018 19:42
 Operator : JU/SJ
 Sample : J3944-01 5X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FOUNDRY-WEST

Quant Time: Jul 16 11:28:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 13 11:54:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.93	152	44999	20.00	ng	0.00
21) Naphthalene-d8	8.21	136	162056	20.00	ng	0.00
38) Acenaphthene-d10	9.98	164	81909	20.00	ng	0.00
63) Phenanthrene-d10	11.48	188	150862	20.00	ng	0.00
75) Chrysene-d12	14.19	240	37120	20.00	ng	0.06
86) Perylene-d12	15.75	264	84307	20.00	ng	0.08
System Monitoring Compounds						
5) 2-Fluorophenol	5.58	112	34826	13.11	ng	0.01
7) Phenol-d6	6.58	99	38504	11.60	ng	0.00
23) Nitrobenzene-d5	7.50	82	22608	7.75	ng	0.00
41) 2,4,6-Tribromophenol	10.78	330	12363	13.02	ng	0.00
44) 2-Fluorobiphenyl	0.00	172	0d	0.00	ng	
78) Terphenyl-d14	13.10	244	20026	13.07	ng	0.04
Target Compounds						
51) Acenaphthene	10.01	154	10327	2.022	ng	96
54) Dibenzofuran	10.18	168	38198	5.461	ng	94
57) Fluorene	10.53	166	20063	3.614	ng	# 96
70) Phenanthrene	11.51	178	324765	44.633	ng	100
74) Fluoranthene	12.73	202	51067	6.367	ng	89
77) Pyrene	12.97	202	24060	9.829	ng	# 73
83) Bis(2-ethylhexyl)phthalate	14.12	149	26904	20.910	ng	# 73
85) Indeno(1,2,3-cd)pyrene	17.32	276	4776	2.704	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071318\
Data File : BF107377.D
Acq On : 13 Jul 2018 19:42
Operator : JU/SJ
Sample : J3944-01 5X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FOUNDRY-WEST

Quant Time: Jul 16 11:28:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 13 11:54:53 2018
Response via : Initial Calibration

