

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071318\
 Data File : BF107396.D
 Acq On : 14 Jul 2018 5:21
 Operator : JU/SJ
 Sample : J3970-10
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 071118-TIPC-F-U-S

Quant Time: Jul 14 06:37:53 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 13 11:54:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.92	152	38883	20.00	ng	0.00
21) Naphthalene-d8	8.21	136	135881	20.00	ng	0.00
38) Acenaphthene-d10	9.97	164	65025	20.00	ng	0.00
63) Phenanthrene-d10	11.47	188	143467	20.00	ng	0.00
75) Chrysene-d12	14.13	240	133129	20.00	ng	0.00
86) Perylene-d12	15.67	264	89907	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	95316	41.51	ng	0.00
7) Phenol-d6	6.58	99	69856	24.36	ng	0.00
23) Nitrobenzene-d5	7.50	82	218943	89.54	ng	0.00
41) 2,4,6-Tribromophenol	10.78	330	71547	94.93	ng	0.00
44) 2-Fluorobiphenyl	9.28	172	398152	107.55	ng	0.00
78) Terphenyl-d14	13.06	244	604668	110.02	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.68	163	45834	9.398	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071318\
Data File : BF107396.D
Acq On : 14 Jul 2018 5:21
Operator : JU/SJ
Sample : J3970-10
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
071118-TIPC-F-U-S

Quant Time: Jul 14 06:37:53 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 13 11:54:53 2018
Response via : Initial Calibration

