

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071321\
 Data File : BF124544.D
 Acq On : 13 Jul 2021 12:51
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 13 14:56:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	139	0.00
2	1,4-Dioxane	0.378	0.399	-5.6	139	0.04
3	Pyridine	0.968	1.048	-8.3	147	0.04
4	n-Nitrosodimethylamine	0.787	0.835	-6.1	140	0.04
5 S	2-Fluorophenol	1.154	1.162	-0.7	139	0.00
6	Aniline	1.480	1.495	-1.0	141	0.00
7 S	Phenol-d6	1.404	1.470	-4.7	142	0.00
8	2-Chlorophenol	1.294	1.301	-0.5	135	0.00
9	Benzaldehyde	0.824	0.745	9.6	137	0.00
10 C	Phenol	1.371	1.402	-2.3	140	0.00
11	bis(2-Chloroethyl)ether	1.063	1.132	-6.5	147	0.00
12	1,3-Dichlorobenzene	1.416	1.497	-5.7	147	0.00
13 C	1,4-Dichlorobenzene	1.447	1.481	-2.3	139	0.00
14	1,2-Dichlorobenzene	1.375	1.411	-2.6	142	0.00
15	Benzyl Alcohol	0.972	1.008	-3.7	138	0.00
16	2,2'-oxybis(1-Chloropropane	1.832	1.925	-5.1	143	0.00
17	2-Methylphenol	0.947	0.995	-5.1	143	0.00
18	Hexachloroethane	0.545	0.550	-0.9	133	0.00
19 P	n-Nitroso-di-n-propylamine	0.790	0.811	-2.7	139	0.00
20	3+4-Methylphenols	1.247	1.301	-4.3	148	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	141	0.00
22	Acetophenone	0.454	0.458	-0.9	141	0.00
23 S	Nitrobenzene-d5	0.317	0.326	-2.8	141	0.00
24	Nitrobenzene	0.314	0.321	-2.2	138	0.00
25	Isophorone	0.582	0.592	-1.7	142	0.00
26 C	2-Nitrophenol	0.169	0.181	-7.1	145	0.00
27	2,4-Dimethylphenol	0.281	0.282	-0.4	139	0.00
28	bis(2-Chloroethoxy)methane	0.339	0.339	0.0	142	0.00
29 C	2,4-Dichlorophenol	0.290	0.302	-4.1	141	0.00
30	1,2,4-Trichlorobenzene	0.321	0.328	-2.2	144	0.00
31	Naphthalene	1.034	1.057	-2.2	144	0.00
32	Benzoic acid	0.196	0.219	-11.7	150#	0.02
33	4-Chloroaniline	0.391	0.397	-1.5	138	0.00
34 C	Hexachlorobutadiene	0.181	0.186	-2.8	141	0.00
35	Caprolactam	0.073	0.082	-12.3	174#	0.00
36 C	4-Chloro-3-methylphenol	0.286	0.293	-2.4	138	0.00
37	2-Methylnaphthalene	0.698	0.720	-3.2	144	0.00
38	1-Methylnaphthalene	0.656	0.666	-1.5	144	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	139	0.00
40	1,2,4,5-Tetrachlorobenzene	0.533	0.568	-6.6	145	0.00
41 P	Hexachlorocyclopentadiene	0.333	0.361	-8.4	146	0.00
42 S	2,4,6-Tribromophenol	0.156	0.162	-3.8	142	0.00
43 C	2,4,6-Trichlorophenol	0.385	0.408	-6.0	142	0.00
44	2,4,5-Trichlorophenol	0.397	0.413	-4.0	147	0.00
45 S	2-Fluorobiphenyl	1.372	1.393	-1.5	140	0.00
46	1,1'-Biphenyl	1.519	1.569	-3.3	143	0.00
47	2-Chloronaphthalene	1.191	1.230	-3.3	142	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071321\
 Data File : BF124544.D
 Acq On : 13 Jul 2021 12:51
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 13 14:56:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.286	0.310	-8.4	137	0.00
49	Acenaphthylene	1.858	1.876	-1.0	140	0.00
50	Dimethylphthalate	1.391	1.439	-3.5	143	0.00
51	2,6-Dinitrotoluene	0.274	0.316	-15.3	141	0.00
52 C	Acenaphthene	1.188	1.235	-4.0	144	0.00
53	3-Nitroaniline	0.306	0.343	-12.1	151#	0.00
54 P	2,4-Dinitrophenol	0.143	0.156	-9.1	151#	0.00
55	Dibenzofuran	1.747	1.804	-3.3	143	0.00
56 P	4-Nitrophenol	0.255	0.272	-6.7	139	0.00
57	2,4-Dinitrotoluene	0.376	0.417	-10.9	146	0.00
58	Fluorene	1.351	1.383	-2.4	145	0.00
59	2,3,4,6-Tetrachlorophenol	0.307	0.323	-5.2	144	0.00
60	Diethylphthalate	1.455	1.479	-1.6	141	0.00
61	4-Chlorophenyl-phenylether	0.613	0.624	-1.8	139	0.00
62	4-Nitroaniline	0.309	0.328	-6.1	146	0.00
63	Azobenzene	1.272	1.301	-2.3	144	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.00
65	4,6-Dinitro-2-methylphenol	0.114	0.125	-9.6	146	0.00
66 c	n-Nitrosodiphenylamine	0.651	0.677	-4.0	141	0.00
67	4-Bromophenyl-phenylether	0.207	0.221	-6.8	146	0.00
68	Hexachlorobenzene	0.212	0.216	-1.9	135	0.00
69	Atrazine	0.199	0.204	-2.5	139	0.00
70 C	Pentachlorophenol	0.134	0.141	-5.2	144	0.00
71	Phenanthrene	1.091	1.116	-2.3	139	0.00
72	Anthracene	1.095	1.105	-0.9	138	0.00
73	Carbazole	1.011	1.034	-2.3	140	0.00
74	Di-n-butylphthalate	1.363	1.408	-3.3	140	0.00
75 C	Fluoranthene	1.101	1.112	-1.0	135	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	130	0.00
77	Benzydine	0.748	0.706	5.6	123	0.00
78	Pyrene	1.662	1.713	-3.1	135	0.00
79 S	Terphenyl-d14	1.191	1.246	-4.6	140	0.00
80	Butylbenzylphthalate	0.803	0.812	-1.1	130	0.00
81	Benzo(a)anthracene	1.326	1.369	-3.2	135	0.00
82	3,3'-Dichlorobenzidine	0.347	0.360	-3.7	130	0.00
83	Chrysene	1.271	1.340	-5.4	136	0.00
84	Bis(2-ethylhexyl)phthalate	1.079	1.128	-4.5	135	0.00
85 c	Di-n-octyl phthalate	1.561	1.632	-4.5	135	0.00
86 I	Perylene-d12	1.000	1.000	0.0	129	0.00
87	Indeno(1,2,3-cd)pyrene	1.160	1.313	-13.2	138	0.00
88	Benzo(b)fluoranthene	1.436	1.590	-10.7	139	0.00
89	Benzo(k)fluoranthene	1.317	1.450	-10.1	129	0.00
90 C	Benzo(a)pyrene	1.185	1.297	-9.5	138	0.00
91	Dibenzo(a,h)anthracene	0.987	1.100	-11.4	138	0.00
92	Benzo(g,h,i)perylene	0.964	1.071	-11.1	136	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071321\
Data File : BF124544.D
Acq On : 13 Jul 2021 12:51
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 13 14:56:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 08 07:49:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0