



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071321\
 Data File : BF124543.D
 Acq On : 13 Jul 2021 11:44
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 08 07:49:42 2021






AutoFind: Scans 1634, 1635, 1636; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.3	6097	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	36.2	4225	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.9	6052	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	11664	PASS
199	198	5	9	6.7	783	PASS
275	198	10	60	23.6	2754	PASS
365	198	1	100	1.8	215	PASS
441	198	0.01	100	10.1	1179	PASS
442	442	50	100	100.0	8129	PASS
443	442	15	24	19.4	1578	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071321\
 Data File : BF124543.D
 Acq On : 13 Jul 2021 11:44
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Jul 13 12:50:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration



TIC: BF124543.D\data.ms

(77) Benzidine

12.792min (+ 0.006) 0.00 ng




response 53320

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	13.80
183.00	10.10	11.43
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071321\
Data File : BF124543.D
Acq On : 13 Jul 2021 11:44
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jul 13 12:50:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 08 07:49:42 2021
Response via : Initial Calibration



TIC: BF124543.D\data.ms

(70) Pentachlorophenol (C)

11.251min (+ 0.000) 0.00 ng

response 9783

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	59.80
264.00	69.90	58.71
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/13/2021	BNA_F	<u>BF124543.D</u>
Compound Name	Response	Retention Time
DDT	24579	13.61
DDD	0	13.05
DDE	0	12.71
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
0	24579	0.00

Instrument :
BNA_F
ClientSampleId :
DFTPP