





Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071415.M
Calibration Table Last Updated: Wed Jul 15 03:58:42 2015


n-Nitrosodimethylamine

Response Ratio



$R = 9.57e-002 A^2 + 4.51e-001 A + 3.83e-002$
Coef of Det (r^2) = 0.991 Curve Fit: Quadratic


Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071415.M
Calibration Table Last Updated: Wed Jul 15 03:58:42 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071415.M
Calibration Table Last Updated: Wed Jul 15 03:58:42 2015

Hexachlorocyclopentadiene


Response Ratio



$R = 6.16e-002 A^2 + 1.20e-001 A + 6.45e-002$
Coef of Det (r^2) = 0.997 Curve Fit: Quadratic


Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071415.M
Calibration Table Last Updated: Wed Jul 15 03:58:42 2015

2,4-Dinitrophenol




$R = 1.89e-002 A^2 + 7.85e-002 A - 3.99e-003$
Coef of Det (r^2) = 0.995 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071415.M
Calibration Table Last Updated: Wed Jul 15 03:58:42 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071415.M
Calibration Table Last Updated: Wed Jul 15 03:58:42 2015

4,6-Dinitro-2-methylphenol




$R = 1.00e-002 A^2 + 8.08e-002 A - 1.63e-003$
Coef of Det (r^2) = 0.997 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071415.M
Calibration Table Last Updated: Wed Jul 15 03:58:42 2015

Indeno(1,2,3-cd)pyrene

Response Ratio



$R = 3.33e-001 A^2 + 1.08e+000 A + 5.32e-002$
Coef of Det (r^2) = 0.999 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071415.M
Calibration Table Last Updated: Wed Jul 15 03:58:42 2015