



Data Path : Z:\HPCHEM1\BNA F\DATA\BF071417\
Data File : BF096778.D
Acq On : 14 Jul 2017 18:16
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 15 06:03:06 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF071317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 13 14:49:34 2017
Response via : Initial Calibration



TIC: BF096778.D

(69) Pentachlorophenol (C)

10.945min (-0.012) 103212.72ng

response 79377

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	61.51
264.00	64.60	61.49
0.00	0.00	0.00