

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
 Data File : BF124563.D
 Acq On : 14 Jul 2021 14:19
 Operator : JU/CG
 Sample : PB137690BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB137690BL

Quant Time: Jul 14 14:51:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.922	152	8139	20.000	ng	0.00
21) Naphthalene-d8	8.204	136	33123	20.000	ng	# 0.00
39) Acenaphthene-d10	9.963	164	18313	20.000	ng	0.00
64) Phenanthrene-d10	11.451	188	33267	20.000	ng	0.00
76) Chrysene-d12	14.092	240	22203	20.000	ng	0.00
86) Perylene-d12	15.586	264	16303	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.563	112	63568	135.319	ng	0.02
7) Phenol-d6	6.545	99	73429	128.492	ng	0.00
23) Nitrobenzene-d5	7.487	82	45516	86.746	ng	0.00
42) 2,4,6-Tribromophenol	10.757	330	17522	122.954	ng	0.00
45) 2-Fluorobiphenyl	9.286	172	106354	84.688	ng	0.00
79) Terphenyl-d14	13.045	244	116950	88.449	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
Data File : BF124563.D
Acq On : 14 Jul 2021 14:19
Operator : JU/CG
Sample : PB137690BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB137690BL

Quant Time: Jul 14 14:51:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 08 07:49:42 2021
Response via : Initial Calibration

