




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
Data File : BF124559.D
Acq On : 14 Jul 2021 11:50
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 14 12:33:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 08 07:49:42 2021
Response via : Initial Calibration



TIC: BF124559.D\data.ms

(70) Pentachlorophenol (C)




11.251min (-0.000) 0.00 ng

response 11143

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	61.70
264.00	69.90	60.07
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
 Data File : BF124559.D
 Acq On : 14 Jul 2021 11:50
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 14 12:33:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration



TIC: BF124559.D\data.ms

(77) Benzidine

12.798min (+ 0.012) 0.00 ng

response 63568

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	13.12
183.00	10.10	10.87
0.00	0.00	0.00