

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
 Data File : BF124560.D
 Acq On : 14 Jul 2021 12:22
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 13:02:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	125	0.00
2	1,4-Dioxane	40.000	44.756	-11.9	135	0.02
3	Pyridine	40.000	43.385	-8.5	133	0.02
4	n-Nitrosodimethylamine	40.000	44.646	-11.6	132	0.02
5 S	2-Fluorophenol	80.000	83.194	-4.0	130	0.00
6	Aniline	40.000	40.899	-2.2	128	0.00
7 S	Phenol-d6	80.000	81.958	-2.4	125	0.00
8	2-Chlorophenol	40.000	41.836	-4.6	127	0.00
9	Benzaldehyde	40.000	37.016	7.5	126	0.00
10 C	Phenol	40.000	40.106	-0.3	123	0.00
11	bis(2-Chloroethyl)ether	40.000	42.313	-5.8	131	0.00
12	1,3-Dichlorobenzene	40.000	42.570	-6.4	133	0.00
13 C	1,4-Dichlorobenzene	40.000	42.149	-5.4	129	0.00
14	1,2-Dichlorobenzene	40.000	42.591	-6.5	133	0.00
15	Benzyl Alcohol	40.000	41.449	-3.6	124	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.118	-2.8	126	0.00
17	2-Methylphenol	40.000	41.193	-3.0	126	0.00
18	Hexachloroethane	40.000	42.495	-6.2	126	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.377	-3.4	126	0.00
20	3+4-Methylphenols	40.000	41.162	-2.9	131	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	126	0.00
22	Acetophenone	40.000	40.549	-1.4	126	0.00
23 S	Nitrobenzene-d5	80.000	83.089	-3.9	128	0.00
24	Nitrobenzene	40.000	41.710	-4.3	125	0.00
25	Isophorone	40.000	39.560	1.1	123	0.00
26 C	2-Nitrophenol	40.000	41.894	-4.7	127	0.00
27	2,4-Dimethylphenol	40.000	42.254	-5.6	131	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.602	1.0	125	0.00
29 C	2,4-Dichlorophenol	40.000	42.644	-6.6	128	0.00
30	1,2,4-Trichlorobenzene	40.000	39.912	0.2	125	0.00
31	Naphthalene	40.000	41.164	-2.9	129	0.00
32	Benzoic acid	40.000	42.165	-5.4	135	0.02
33	4-Chloroaniline	40.000	41.698	-4.2	127	0.00
34 C	Hexachlorobutadiene	40.000	42.030	-5.1	129	0.00
35	Caprolactam	40.000	43.939	-9.8	152	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.992	0.0	121	0.00
37	2-Methylnaphthalene	40.000	41.463	-3.7	129	0.00
38	1-Methylnaphthalene	40.000	42.128	-5.3	133	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	124	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.204	-3.0	125	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.405	-6.0	128	0.00
42 S	2,4,6-Tribromophenol	80.000	90.207	-12.8	138	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.827	-7.1	128	0.00
44	2,4,5-Trichlorophenol	40.000	43.240	-8.1	136	0.00
45 S	2-Fluorobiphenyl	80.000	83.211	-4.0	128	0.00
46	1,1'-Biphenyl	40.000	41.400	-3.5	128	0.00
47	2-Chloronaphthalene	40.000	41.592	-4.0	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
 Data File : BF124560.D
 Acq On : 14 Jul 2021 12:22
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 13:02:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.586	-11.5	126	0.00
49	Acenaphthylene	40.000	41.611	-4.0	128	0.00
50	Dimethylphthalate	40.000	42.135	-5.3	130	0.00
51	2,6-Dinitrotoluene	40.000	46.505	-16.3	127	0.00
52 C	Acenaphthene	40.000	42.959	-7.4	133	0.00
53	3-Nitroaniline	40.000	44.849	-12.1	135	0.00
54 P	2,4-Dinitrophenol	40.000	47.999	-20.0	149	0.00
55	Dibenzofuran	40.000	42.534	-6.3	131	0.00
56 P	4-Nitrophenol	40.000	45.785	-14.5	133	0.00
57	2,4-Dinitrotoluene	40.000	46.527	-16.3	137	0.00
58	Fluorene	40.000	40.852	-2.1	129	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.777	-9.4	133	0.00
60	Diethylphthalate	40.000	41.987	-5.0	130	0.00
61	4-Chlorophenyl-phenylether	40.000	41.834	-4.6	128	0.00
62	4-Nitroaniline	40.000	46.103	-15.3	141	0.00
63	Azobenzene	40.000	41.344	-3.4	130	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	130	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.229	-8.1	134	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.662	-1.7	128	0.00
67	4-Bromophenyl-phenylether	40.000	39.563	1.1	126	0.00
68	Hexachlorobenzene	40.000	39.615	1.0	122	0.00
69	Atrazine	40.000	40.830	-2.1	129	0.00
70 C	Pentachlorophenol	40.000	43.669	-9.2	140	0.00
71	Phenanthrene	40.000	40.457	-1.1	128	0.00
72	Anthracene	40.000	41.039	-2.6	131	0.00
73	Carbazole	40.000	41.305	-3.3	132	0.00
74	Di-n-butylphthalate	40.000	40.354	-0.9	127	0.00
75 C	Fluoranthene	40.000	41.394	-3.5	129	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	121	0.00
77	Benzydine	40.000	41.604	-4.0	126	0.00
78	Pyrene	40.000	41.723	-4.3	127	0.00
79 S	Terphenyl-d14	80.000	84.290	-5.4	131	0.00
80	Butylbenzylphthalate	40.000	41.957	-4.9	125	0.00
81	Benzo(a)anthracene	40.000	41.664	-4.2	126	0.00
82	3,3'-Dichlorobenzidine	40.000	41.091	-2.7	119	0.00
83	Chrysene	40.000	42.483	-6.2	127	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.575	-1.4	122	0.00
85 c	Di-n-octyl phthalate	40.000	41.240	-3.1	123	0.00
86 I	Perylene-d12	20.000	20.000	0.0	130	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.435	-13.6	140	0.00
88	Benzo(b)fluoranthene	40.000	41.724	-4.3	133	0.00
89	Benzo(k)fluoranthene	40.000	40.740	-1.9	121	0.00
90 C	Benzo(a)pyrene	40.000	41.659	-4.1	133	0.00
91	Dibenzo(a,h)anthracene	40.000	44.546	-11.4	140	0.00
92	Benzo(g,h,i)perylene	40.000	44.043	-10.1	136	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
Data File : BF124560.D
Acq On : 14 Jul 2021 12:22
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 14 13:02:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 08 07:49:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0