

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
 Data File : BF124567.D
 Acq On : 14 Jul 2021 16:59
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 17:25:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 14 13:25:43 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	123	0.00
2	1,4-Dioxane	40.000	44.911	-12.3	131	0.02
3	Pyridine	40.000	43.277	-8.2	130	0.02
4	n-Nitrosodimethylamine	40.000	44.737	-11.8	131	0.02
5 S	2-Fluorophenol	80.000	83.031	-3.8	128	0.00
6	Aniline	40.000	41.099	-2.7	127	0.00
7 S	Phenol-d6	80.000	84.874	-6.1	128	0.00
8	2-Chlorophenol	40.000	42.215	-5.5	126	0.00
9	Benzaldehyde	40.000	35.979	10.1	121	0.00
10 C	Phenol	40.000	41.057	-2.6	125	0.00
11	bis(2-Chloroethyl)ether	40.000	40.571	-1.4	124	0.00
12	1,3-Dichlorobenzene	40.000	43.397	-8.5	134	0.00
13 C	1,4-Dichlorobenzene	40.000	44.029	-10.1	133	0.00
14	1,2-Dichlorobenzene	40.000	42.126	-5.3	129	0.00
15	Benzyl Alcohol	40.000	40.800	-2.0	121	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.994	-7.5	130	0.00
17	2-Methylphenol	40.000	41.392	-3.5	125	0.00
18	Hexachloroethane	40.000	41.569	-3.9	121	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.862	-4.7	126	0.00
20	3+4-Methylphenols	40.000	41.665	-4.2	131	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	127	0.00
22	Acetophenone	40.000	40.847	-2.1	129	0.00
23 S	Nitrobenzene-d5	80.000	82.029	-2.5	127	0.00
24	Nitrobenzene	40.000	41.223	-3.1	125	0.00
25	Isophorone	40.000	39.779	0.6	125	0.00
26 C	2-Nitrophenol	40.000	43.403	-8.5	133	0.00
27	2,4-Dimethylphenol	40.000	41.656	-4.1	131	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.565	1.1	127	0.00
29 C	2,4-Dichlorophenol	40.000	41.501	-3.8	126	0.00
30	1,2,4-Trichlorobenzene	40.000	40.905	-2.3	130	0.00
31	Naphthalene	40.000	41.239	-3.1	131	0.00
32	Benzoic acid	40.000	45.973	-14.9	140	0.02
33	4-Chloroaniline	40.000	40.780	-2.0	125	0.00
34 C	Hexachlorobutadiene	40.000	42.311	-5.8	131	0.00
35	Caprolactam	40.000	37.719	5.7	132	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.660	-1.6	124	0.00
37	2-Methylnaphthalene	40.000	40.671	-1.7	128	0.00
38	1-Methylnaphthalene	40.000	40.366	-0.9	129	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	129	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.137	-2.8	130	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.192	-3.0	129	0.00
42 S	2,4,6-Tribromophenol	80.000	83.813	-4.8	133	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.236	-5.6	132	0.00
44	2,4,5-Trichlorophenol	40.000	40.776	-1.9	134	0.00
45 S	2-Fluorobiphenyl	80.000	80.472	-0.6	129	0.00
46	1,1'-Biphenyl	40.000	40.809	-2.0	132	0.00
47	2-Chloronaphthalene	40.000	39.747	0.6	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
 Data File : BF124567.D
 Acq On : 14 Jul 2021 16:59
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 17:25:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 14 13:25:43 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.459	-11.1	131	0.00
49	Acenaphthylene	40.000	40.490	-1.2	130	0.00
50	Dimethylphthalate	40.000	40.724	-1.8	131	0.00
51	2,6-Dinitrotoluene	40.000	45.084	-12.7	128	0.00
52 C	Acenaphthene	40.000	41.456	-3.6	133	0.00
53	3-Nitroaniline	40.000	43.285	-8.2	136	0.00
54 P	2,4-Dinitrophenol	40.000	45.378	-13.4	147	0.00
55	Dibenzofuran	40.000	40.555	-1.4	130	0.00
56 P	4-Nitrophenol	40.000	44.191	-10.5	134	0.00
57	2,4-Dinitrotoluene	40.000	46.238	-15.6	142	0.00
58	Fluorene	40.000	39.682	0.8	131	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.827	-4.6	133	0.00
60	Diethylphthalate	40.000	39.959	0.1	128	0.00
61	4-Chlorophenyl-phenylether	40.000	41.086	-2.7	131	0.00
62	4-Nitroaniline	40.000	43.894	-9.7	140	0.00
63	Azobenzene	40.000	39.751	0.6	130	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	134	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.903	-4.8	134	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.305	-0.8	131	0.00
67	4-Bromophenyl-phenylether	40.000	39.849	0.4	131	0.00
68	Hexachlorobenzene	40.000	40.498	-1.2	128	0.00
69	Atrazine	40.000	40.968	-2.4	134	0.00
70 C	Pentachlorophenol	40.000	42.544	-6.4	140	0.00
71	Phenanthrene	40.000	40.062	-0.2	130	0.00
72	Anthracene	40.000	39.538	1.2	130	0.00
73	Carbazole	40.000	40.181	-0.5	133	0.00
74	Di-n-butylphthalate	40.000	40.369	-0.9	131	0.00
75 C	Fluoranthene	40.000	40.726	-1.8	131	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	126	0.00
77	Benzidine	40.000	39.953	0.1	126	0.00
78	Pyrene	40.000	41.534	-3.8	131	0.00
79 S	Terphenyl-d14	80.000	83.696	-4.6	135	0.00
80	Butylbenzylphthalate	40.000	40.964	-2.4	127	0.00
81	Benzo(a)anthracene	40.000	41.363	-3.4	130	0.00
82	3,3'-Dichlorobenzidine	40.000	42.514	-6.3	128	0.00
83	Chrysene	40.000	41.734	-4.3	130	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.780	-4.5	130	0.00
85 c	Di-n-octyl phthalate	40.000	43.007	-7.5	134	0.00
86 I	Perylene-d12	20.000	20.000	0.0	140	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.836	-7.1	143	0.00
88	Benzo(b)fluoranthene	40.000	40.490	-1.2	139	0.00
89	Benzo(k)fluoranthene	40.000	39.765	0.6	127	0.00
90 C	Benzo(a)pyrene	40.000	40.901	-2.3	141	0.00
91	Dibenzo(a,h)anthracene	40.000	43.862	-9.7	149	0.00
92	Benzo(g,h,i)perylene	40.000	42.795	-7.0	143	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071421\
Data File : BF124567.D
Acq On : 14 Jul 2021 16:59
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 14 17:25:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 14 13:25:43 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0