

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071516\
 Data File : BF089043.D
 Acq On : 16 Jul 2016 6:03
 Operator : UM/SJ
 Sample : H3972-14DL 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LSS-SS-SS04(0-2in)DL

Quant Time: Jul 18 01:01:18 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF063016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 11 18:13:32 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.67	152	68131	20.00	ng	-0.03
21) Naphthalene-d8	7.96	136	262357	20.00	ng	-0.03
38) Acenaphthene-d10	9.71	164	101796	20.00	ng	-0.03
63) Phenanthrene-d10	11.19	188	157310	20.00	ng	-0.03
75) Chrysene-d12	13.81	240	121745	20.00	ng	-0.05
86) Perylene-d12	15.17	264	83611	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	38757	8.53	ng	-0.02
7) Phenol-d6	6.32	99	54240	9.23	ng	-0.03
23) Nitrobenzene-d5	7.23	82	31004	6.19	ng	-0.05
41) 2,4,6-Tribromophenol	10.49	330	6210	6.57	ng	-0.05
44) 2-Fluorobiphenyl	9.04	172	56204	8.72	ng	-0.03
78) Terphenyl-d14	12.77	244	29295	5.36	ng	-0.03
Target Compounds						Qvalue
74) Fluoranthene	12.39	202	37425	4.29	ng	93
77) Pyrene	12.61	202	31708	3.07	ng	97
79) Butylbenzylphthalate	13.25	149	105816	22.98	ng	93
80) Benzo(a)anthracene	13.79	228	17502m	2.23	ng	
82) Chrysene	13.83	228	16487m	2.13	ng	
83) Bis(2-ethylhexyl)phthalate	13.81	149	58938	11.21	ng	98
87) Benzo(b)fluoranthene	14.80	252	19366m	3.69	ng	
89) Benzo(a)pyrene	15.11	252	11372	2.56	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071516\
Data File : BF089043.D
Acq On : 16 Jul 2016 6:03
Operator : UM/SJ
Sample : H3972-14DL 10X
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
LSS-SS-SS04(0-2in)DL

Quant Time: Jul 18 01:01:18 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF063016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 11 18:13:32 2016
Response via : Initial Calibration

