

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071521\
 Data File : BF124582.D
 Acq On : 15 Jul 2021 20:51
 Operator : JU/CG
 Sample : M3026-11RE
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ABN-JG-4RE

Quant Time: Jul 16 02:52:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.922	152	5483	20.000	ng	0.00
21) Naphthalene-d8	8.204	136	21454	20.000	ng	0.00
39) Acenaphthene-d10	9.963	164	11489	20.000	ng	0.00
64) Phenanthrene-d10	11.451	188	19909	20.000	ng	# 0.00
76) Chrysene-d12	14.092	240	10116	20.000	ng	# 0.00
86) Perylene-d12	15.586	264	10499	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	2266	7.160	ng	0.00
7) Phenol-d6	6.534	99	11142	28.942	ng	-0.01
23) Nitrobenzene-d5	7.481	82	22242	65.446	ng	-0.01
42) 2,4,6-Tribromophenol	10.745	330	283	3.165	ng	-0.01
45) 2-Fluorobiphenyl	9.280	172	46607	59.156	ng	0.00
79) Terphenyl-d14	13.039	244	37091	61.569	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071521\
Data File : BF124582.D
Acq On : 15 Jul 2021 20:51
Operator : JU/CG
Sample : M3026-11RE
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ABN-JG-4RE

Quant Time: Jul 16 02:52:00 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 08 07:49:42 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.922 min Scan# 822
Delta R.T. -0.006 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51




Instrument :
BNA_F
ClientSampleId :
ABN-JG-4RE


Tgt Ion:152 Resp: 5483
Ion Ratio Lower Upper
152 100
150 152.8 117.6 176.4
115 62.6 47.1 70.7



#5
2-Fluorophenol
Concen: 7.160 ng
RT: 5.534 min Scan# 586
Delta R.T. -0.005 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51

Tgt Ion:112 Resp: 2266
Ion Ratio Lower Upper
112 100
64 52.8 44.2 66.2
63 27.3 26.2 39.2









#7
Phenol-d6
Concen: 28.942 ng
RT: 6.534 min Scan# 756
Delta R.T. -0.011 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51

Instrument :
BNA_F
ClientSampleId :
ABN-JG-4RE

Tgt Ion: 99 Resp: 11142




Ion	Ratio	Lower	Upper
99	100		
42	29.2	23.5	35.3
71	34.0	25.6	38.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.204 min Scan# 1040
Delta R.T. -0.006 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51

Tgt Ion: 136 Resp: 21454

Ion	Ratio	Lower	Upper
136	100		
137	10.3	9.8	14.6
54	8.6	6.4	9.6
68	3.9	3.3	4.9





#23

Nitrobenzene-d5

Concen: 65.446 ng

RT: 7.481 min Scan# 917

Delta R.T. -0.011 min

Lab File: BF124582.D

Acq: 15 Jul 2021 20:51





Tgt Ion: 82 Resp: 22242

Ion Ratio Lower Upper

82 100

128 53.0 40.6 60.8

54 66.6 49.6 74.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.963 min Scan# 1339

Delta R.T. -0.006 min

Lab File: BF124582.D

Acq: 15 Jul 2021 20:51




Tgt Ion:164 Resp: 11489

Ion Ratio Lower Upper

164 100

162 106.6 81.3 121.9

160 43.9 38.0 57.0




Instrument :

BNA_F


ClientSampleId :

ABN-JG-4RE






#42
2,4,6-Tribromophenol
Concen: 3.165 ng
RT: 10.745 min Scan# 1472
Delta R.T. -0.012 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51

Instrument :
BNA_F
ClientSampleId :
ABN-JG-4RE







Tgt Ion: 330 Resp: 283
Ion Ratio Lower Upper
330 100
332 56.9 76.2 114.4#
141 0.0 44.6 66.8#



#45
2-Fluorobiphenyl
Concen: 59.156 ng
RT: 9.280 min Scan# 1223
Delta R.T. -0.006 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51

Tgt Ion: 172 Resp: 46607
Ion Ratio Lower Upper
172 100
171 34.9 26.5 39.7
170 24.5 18.8 28.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.451 min Scan# 1592
Delta R.T. -0.006 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51




Instrument :
BNA_F
ClientSampleId :
ABN-JG-4RE


Tgt Ion:188 Resp: 19909
Ion Ratio Lower Upper
188 100
94 6.6 6.7 10.1#
80 8.9 7.5 11.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.092 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51

Tgt Ion:240 Resp: 10116
Ion Ratio Lower Upper
240 100
120 9.1 10.8 16.2#
236 27.4 21.6 32.4









#79
Terphenyl-d14
Concen: 61.569 ng
RT: 13.039 min Scan# 1862
Delta R.T. -0.006 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51

Instrument :
BNA_F
ClientSampleId :
ABN-JG-4RE

Tgt Ion:244 Resp: 37091
Ion Ratio Lower Upper
244 100
212 6.7 5.9 8.9
122 11.6 9.0 13.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.586 min Scan# 2295
Delta R.T. -0.006 min
Lab File: BF124582.D
Acq: 15 Jul 2021 20:51

Tgt Ion:264 Resp: 10499
Ion Ratio Lower Upper
264 100
260 24.4 19.6 29.4
265 23.6 17.2 25.8

