

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071522\
 Data File : BF129257.D
 Acq On : 16 Jul 2022 01:14
 Operator : CG\JU
 Sample : N3716-01 5X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-1

Quant Time: Jul 16 01:44:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 12:55:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	208636	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	678381	20.000	ng	# 0.00
39) Acenaphthene-d10	9.957	164	318164	20.000	ng	0.02
64) Phenanthrene-d10	11.457	188	439453	20.000	ng	# 0.03
76) Chrysene-d12	14.086	240	493491	20.000	ng	0.01
86) Perylene-d12	15.574	264	363716	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	215686	16.510	ng	0.00
7) Phenol-d6	6.522	99	259353	15.350	ng	0.00
23) Nitrobenzene-d5	7.463	82	205771	16.214	ng	0.00
42) 2,4,6-Tribromophenol	10.751	330	40895	13.848	ng	0.02
45) 2-Fluorobiphenyl	9.263	172	207188	11.033	ng	0.00
79) Terphenyl-d14	13.027	244	218646	8.288	ng	0.01
Target Compounds						Qvalue
9) Benzaldehyde	6.457	77	45383	4.377	ng	# 1
18) Hexachloroethane	7.434	117	243693	44.080	ng	# 1
19) n-Nitroso-di-n-propyla...	7.334	70	45823	4.611	ng	# 69
22) Acetophenone	7.292	105	583125	37.248	ng	# 55
24) Nitrobenzene	7.469	77	37259	3.086	ng	# 17
25) Isophorone	7.716	82	211496	9.916	ng	# 1
26) 2-Nitrophenol	7.804	139	12365	2.039	ng	# 1
28) bis(2-Chloroethoxy)met...	7.951	93	32153	2.263	ng	# 1
31) Naphthalene	8.210	128	164552	5.120	ng	# 52
35) Caprolactam	8.645	113	21715	6.973	ng	# 1
37) 2-Methylnaphthalene	8.910	142	348809	16.651	ng	99
38) 1-Methylnaphthalene	9.004	142	266512	13.155	ng	99
46) 1,1'-Biphenyl	9.369	154	81700	3.631	ng	96
51) 2,6-Dinitrotoluene	9.757	165	14649	3.279	ng	# 63
52) Acenaphthene	9.992	154	43410	2.489	ng	# 55
53) 3-Nitroaniline	9.910	138	17563	3.292	ng	# 51
56) 4-Nitrophenol	10.063	139	44162	10.458	ng	# 1
57) 2,4-Dinitrotoluene	10.163	165	24476	4.178	ng	# 63
58) Fluorene	10.510	166	125139	6.646	ng	# 69
60) Diethylphthalate	10.369	149	52095	2.609	ng	# 36
66) n-Nitrosodiphenylamine	10.616	169	254256	18.134	ng	# 58
69) Atrazine	11.145	200	27200	6.998	ng	# 50
71) Phenanthrene	11.480	178	149730	6.711	ng	# 75
72) Anthracene	11.569	178	70438	3.178	ng	# 1
75) Fluoranthene	12.669	202	57365	2.582	ng	70
78) Pyrene	12.898	202	124369	3.112	ng	# 93
84) Bis(2-ethylhexyl)phtha...	14.057	149	403448	17.031	ng	# 62
85) Di-n-octyl phthalate	14.663	149	586357	17.257	ng	# 96
88) Benzo(b)fluoranthene	15.133	252	45664	2.044	ng	# 1
89) Benzo(k)fluoranthene	15.133	252	45664	2.092	ng	# 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071522\
Data File : BF129257.D
Acq On : 16 Jul 2022 01:14
Operator : CG\JU
Sample : N3716-01 5X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP-1

Quant Time: Jul 16 01:44:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 15 12:55:49 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.904 min Scan# 71
 Delta R.T. 0.006 min
 Lab File: BF129257.D
 Acq: 16 Jul 2022 01:14




Instrument :
 BNA_F
 ClientSampleId :
 COMP-1

Tgt Ion:152 Resp: 208636
 Ion Ratio Lower Upper
 152 100
 150 155.1 123.7 185.5
 115 63.7 49.9 74.9



#5
 2-Fluorophenol
 Concen: 16.510 ng
 RT: 5.516 min Scan# 549
 Delta R.T. 0.000 min
 Lab File: BF129257.D
 Acq: 16 Jul 2022 01:14

Tgt Ion:112 Resp: 215686
 Ion Ratio Lower Upper
 112 100
 64 57.6 53.3 79.9
 63 30.9 28.0 42.0









#7
Phenol-d6
Concen: 15.350 ng
RT: 6.522 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion: 99 Resp: 259353
Ion Ratio Lower Upper
99 100
42 21.0 17.3 25.9
71 35.6 28.9 43.3



#9
Benzaldehyde
Concen: 4.377 ng
RT: 6.457 min Scan# 709
Delta R.T. 0.006 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion: 77 Resp: 45383
Ion Ratio Lower Upper
77 100
105 836.7 81.6 121.6#
106 72.1 76.1 116.1#









#18
Hexachloroethane
Concen: 44.080 ng
RT: 7.434 min Scan# 81
Delta R.T. 0.024 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion:117 Resp: 243693
Ion Ratio Lower Upper
117 100
119 389.9 77.4 116.2#
201 0.0 84.6 127.0#



#19
n-Nitroso-di-n-propylamine
Concen: 4.611 ng
RT: 7.334 min Scan# 858
Delta R.T. 0.018 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion: 70 Resp: 45823
Ion Ratio Lower Upper
70 100
42 43.6 52.3 78.5#
101 0.0 7.5 11.3#
130 0.0 16.1 24.1#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 10
Delta R.T. 0.006 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion:136 Resp: 678381
Ion Ratio Lower Upper
136 100
137 12.5 8.8 13.2
54 12.4 7.5 11.3#
68 12.1 4.7 7.1#



#22
Acetophenone
Concen: 37.248 ng
RT: 7.292 min Scan# 851
Delta R.T. -0.018 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion:105 Resp: 583125
Ion Ratio Lower Upper
105 100
71 16.8 5.0 7.4#
51 5.0 25.0 37.4#
120 0.2 17.8 26.8#





#23

Nitrobenzene-d5

Concen: 16.214 ng

RT: 7.463 min Scan# 81

Delta R.T. -0.006 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Instrument :

BNA_F

ClientSampleId :

COMP-1





Tgt Ion: 82 Resp: 205771

Ion Ratio Lower Upper

82 100

128 39.5 34.0 51.0

54 44.7 46.9 70.3#



#24

Nitrobenzene

Concen: 3.086 ng

RT: 7.469 min Scan# 881

Delta R.T. -0.018 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14




Tgt Ion: 77 Resp: 37259


Ion Ratio Lower Upper

77 100

123 20.0 35.4 53.2#

65 105.9 11.8 17.8#









#25
Isophorone
Concen: 9.916 ng
RT: 7.716 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion: 82 Resp: 211496
Ion Ratio Lower Upper
82 100
95 169.3 5.7 8.5#
138 8.7 14.2 21.2#



#26
2-Nitrophenol
Concen: 2.039 ng
RT: 7.804 min Scan# 938
Delta R.T. 0.006 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion:139 Resp: 12365
Ion Ratio Lower Upper
139 100
109 227.2 25.0 37.4#
65 229.1 45.1 67.7#









#28
bis(2-Chloroethoxy)methane
Concen: 2.263 ng
RT: 7.951 min Scan# 90
Delta R.T. 0.024 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion: 93 Resp: 32153
Ion Ratio Lower Upper
93 100
95 302.2 26.8 40.2#
123 98.5 9.2 13.8#



#31
Naphthalene
Concen: 5.120 ng
RT: 8.210 min Scan# 1007
Delta R.T. 0.006 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion: 128 Resp: 164552
Ion Ratio Lower Upper
128 100
129 38.9 9.2 13.8#
127 24.1 10.5 15.7#









#35
 Caprolactam
 Concen: 6.973 ng
 RT: 8.645 min Scan# 1080
 Delta R.T. 0.006 min
 Lab File: BF129257.D
 Acq: 16 Jul 2022 01:14




Instrument :
 BNA_F
 ClientSampleId :
 COMP-1


Tgt Ion:113 Resp: 21715
 Ion Ratio Lower Upper
 113 100
 55 1283.0 188.5 228.5#
 56 523.6 130.6 170.6#



#37
 2-Methylnaphthalene
 Concen: 16.651 ng
 RT: 8.910 min Scan# 1126
 Delta R.T. 0.012 min
 Lab File: BF129257.D
 Acq: 16 Jul 2022 01:14

Tgt Ion:142 Resp: 348809
 Ion Ratio Lower Upper
 142 100
 141 89.4 71.0 106.4





#38

1-Methylnaphthalene

Concen: 13.155 ng

RT: 9.004 min Scan# 1142

Delta R.T. 0.006 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Instrument :

BNA_F

ClientSampleId :





COMP-1

Tgt Ion:142 Resp: 266512

Ion Ratio Lower Upper

142 100

141 93.4 74.2 111.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.957 min Scan# 1304

Delta R.T. 0.018 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14




Tgt Ion:164 Resp: 318164


Ion Ratio Lower Upper

164 100

162 97.0 79.5 119.3

160 46.7 34.2 51.4





#42

2,4,6-Tribromophenol

Concen: 13.848 ng

RT: 10.751 min Scan# 1439

Delta R.T. 0.018 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Instrument :

BNA_F

ClientSampleId :

COMP-1





Tgt Ion:330 Resp: 40895

Ion Ratio Lower Upper

330 100

332 103.2 76.8 115.2

141 24.5 28.0 42.0#



#45

2-Fluorobiphenyl

Concen: 11.033 ng

RT: 9.263 min Scan# 1186

Delta R.T. 0.000 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14




Tgt Ion:172 Resp: 207188


Ion Ratio Lower Upper

172 100

171 34.3 28.8 43.2

170 23.3 19.8 29.8









#46
1,1'-Biphenyl
Concen: 3.631 ng
RT: 9.369 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion:154 Resp: 81700
Ion Ratio Lower Upper
154 100
153 37.6 20.6 60.6
76 13.6 0.0 33.3



#51
2,6-Dinitrotoluene
Concen: 3.279 ng
RT: 9.757 min Scan# 1270
Delta R.T. 0.029 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion:165 Resp: 14649
Ion Ratio Lower Upper
165 100
63 34.0 44.9 67.3#
89 23.3 43.6 65.4#









#52
Acenaphthene
Concen: 2.489 ng
RT: 9.992 min Scan# 11
Delta R.T. 0.018 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion:154 Resp: 43410
Ion Ratio Lower Upper
154 100
153 141.7 89.6 134.4#
152 110.8 42.4 63.6#



#53
3-Nitroaniline
Concen: 3.292 ng
RT: 9.910 min Scan# 1296
Delta R.T. 0.012 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion:138 Resp: 17563
Ion Ratio Lower Upper
138 100
108 93.5 10.9 16.3#
92 86.3 98.2 147.2#









#56
4-Nitrophenol
Concen: 10.458 ng
RT: 10.063 min Scan# 1322
Delta R.T. 0.012 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion:139 Resp: 44162
Ion Ratio Lower Upper
139 100
109 283.1 50.7 90.7#
65 88.8 74.5 114.5



#57
2,4-Dinitrotoluene
Concen: 4.178 ng
RT: 10.163 min Scan# 1339
Delta R.T. 0.035 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion:165 Resp: 24476
Ion Ratio Lower Upper
165 100
63 56.2 39.8 59.6
89 30.0 60.2 90.2#
182 30.3 2.2 3.2#





#58

Fluorene

Concen: 6.646 ng

RT: 10.510 min Scan# 11

Delta R.T. 0.018 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Instrument :

BNA_F

ClientSampleId :

COMP-1





Tgt Ion:166 Resp: 125139

Ion Ratio Lower Upper

166 100

165 124.7 78.1 117.1#

167 34.7 10.5 15.7#



#60

Diethylphthalate

Concen: 2.609 ng

RT: 10.369 min Scan# 1374

Delta R.T. 0.006 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14




Tgt Ion:149 Resp: 52095


Ion Ratio Lower Upper

149 100

177 50.4 17.3 25.9#

150 40.5 10.2 15.2#









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.457 min Scan# 11
Delta R.T. 0.029 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion:188 Resp: 439453
Ion Ratio Lower Upper
188 100
94 16.5 7.0 10.6#
80 17.2 7.6 11.4#



#66
n-Nitrosodiphenylamine
Concen: 18.134 ng
RT: 10.616 min Scan# 1416
Delta R.T. 0.012 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion:169 Resp: 254256
Ion Ratio Lower Upper
169 100
168 30.8 50.1 75.1#
167 9.5 27.9 41.9#





#69

Atrazine

Concen: 6.998 ng

RT: 11.145 min Scan# 1

Delta R.T. 0.018 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Instrument :

BNA_F

ClientSampleId :

COMP-1





Tgt Ion:200 Resp: 27200

Ion Ratio Lower Upper

200 100

173 86.9 6.3 46.3#

215 57.1 28.3 68.3



#71

Phenanthrene

Concen: 6.711 ng

RT: 11.480 min Scan# 1563

Delta R.T. 0.024 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14




Tgt Ion:178 Resp: 149730


Ion Ratio Lower Upper

178 100

176 22.2 16.0 24.0

179 36.9 12.6 19.0#





#72

Anthracene

Concen: 3.178 ng

RT: 11.569 min Scan# 1

Delta R.T. 0.059 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Instrument :

BNA_F

ClientSampleId :

COMP-1





Tgt Ion:178 Resp: 70438

Ion Ratio Lower Upper

178 100

176 28.0 15.4 23.2#

179 195.1 12.6 19.0#



#75

Fluoranthene

Concen: 2.582 ng

RT: 12.669 min Scan# 1765

Delta R.T. 0.024 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14




Tgt Ion:202 Resp: 57365


Ion Ratio Lower Upper

202 100

101 16.8 0.0 31.4

203 35.4 0.0 38.0





#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.086 min Scan# 20

Delta R.T. 0.012 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Instrument :

BNA_F

ClientSampleId :

COMP-1





Tgt Ion:240 Resp: 493491

Ion Ratio Lower Upper

240 100

120 12.5 8.9 13.3

236 25.5 20.8 31.2



#78

Pyrene

Concen: 3.112 ng

RT: 12.898 min Scan# 1804

Delta R.T. 0.024 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14




Tgt Ion:202 Resp: 124369


Ion Ratio Lower Upper

202 100

200 21.3 17.7 26.5

203 24.2 14.6 21.8#









#79
 Terphenyl-d14
 Concen: 8.288 ng
 RT: 13.027 min Scan# 1
 Delta R.T. 0.012 min
 Lab File: BF129257.D
 Acq: 16 Jul 2022 01:14




Instrument :
 BNA_F
 ClientSampleId :
 COMP-1


Tgt Ion:244 Resp: 218646
 Ion Ratio Lower Upper
 244 100
 212 8.2 6.1 9.1
 122 18.5 8.8 13.2#



#84
 Bis(2-ethylhexyl)phthalate
 Concen: 17.031 ng
 RT: 14.057 min Scan# 2001
 Delta R.T. 0.018 min
 Lab File: BF129257.D
 Acq: 16 Jul 2022 01:14

Tgt Ion:149 Resp: 403448
 Ion Ratio Lower Upper
 149 100
 167 4.6 21.0 31.4#
 279 0.8 3.1 4.7#









#85
Di-n-octyl phthalate
Concen: 17.257 ng
RT: 14.663 min Scan# 2104
Delta R.T. 0.006 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14




Instrument :
BNA_F
ClientSampleId :
COMP-1


Tgt Ion:149 Resp: 586357
Ion Ratio Lower Upper
149 100
167 2.0 1.3 1.9#
43 9.9 9.2 13.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.574 min Scan# 2259
Delta R.T. 0.012 min
Lab File: BF129257.D
Acq: 16 Jul 2022 01:14

Tgt Ion:264 Resp: 363716
Ion Ratio Lower Upper
264 100
260 24.0 18.4 27.6
265 22.2 17.2 25.8





#88

Benzo(b)fluoranthene

Concen: 2.044 ng

RT: 15.133 min Scan# 2184

Delta R.T. 0.012 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Instrument :

BNA_F

ClientSampleId :

COMP-1





Tgt Ion:252 Resp: 45664

Ion Ratio Lower Upper

252 100

253 32.2 17.7 26.5#

125 113.7 8.0 12.0#



#89

Benzo(k)fluoranthene

Concen: 2.092 ng

RT: 15.133 min Scan# 2184

Delta R.T. -0.024 min

Lab File: BF129257.D

Acq: 16 Jul 2022 01:14

Tgt Ion:252 Resp: 45664

Ion Ratio Lower Upper

252 100

253 32.2 17.8 26.6#

125 113.7 8.3 12.5#

