


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071615\
 Data File : BF080503.D
 Acq On : 16 Jul 2015 10:41
 Operator : TP/UM
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071615.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 16 17:11:01 2015



AutoFind: Scans 853, 854, 855; Background Corrected with Scan 850

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.0	4775	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	37.1	4220	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.5	5289	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	11372	PASS
199	198	5	9	6.3	721	PASS
275	198	10	60	26.0	2956	PASS
365	198	1	100	2.3	262	PASS
441	442	0.01	24	18.0	1159	PASS
442	198	50	100	56.6	6441	PASS
443	442	15	24	20.7	1336	PASS