


Benzaldehyde

Response Ratio




$R = -7.91e-002 A^2 + 8.77e-001 A + 8.93e-002$
Coef of Det (r^2) = 0.992 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071620.M
Calibration Table Last Updated: Thu Jul 16 14:20:32 2020

2,4-Dinitrophenol


Response Ratio



Resp Ratio = $1.95 \times 10^{-1} \times \text{Amt} - 6.67 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071620.M
Calibration Table Last Updated: Thu Jul 16 14:20:32 2020

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.26 \times 10^{-1} \times \text{Amt} - 2.42 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071620.M
Calibration Table Last Updated: Thu Jul 16 14:20:32 2020