


Pyridine

Response Ratio




$R = 2.76e-001 A^2 + 2.47e-001 A + 3.60e-001$
Coef of Det (r^2) = 0.993 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015


n-Nitroso-di-n-propylamine

Response Ratio




$R = 1.55e-001 A^2 + 4.77e-001 A + 8.00e-002$
Coef of Det (r^2) = 0.998 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015




Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015

2-Nitroaniline

Response Ratio




$R = 4.85e-002 A^2 + 1.92e-001 A + 1.18e-002$
Coef of Det (r^2) = 0.996 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015

2,6-Dinitrotoluene

Response Ratio




$R = 5.05e-002 A^2 + 1.41e-001 A + 2.52e-002$
Coef of Det (r^2) = 0.995 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015


3-Nitroaniline

Response Ratio



$R = 4.11e-002 A^2 + 2.10e-001 A + 1.15e-002$
Coef of Det (r^2) = 0.998 Curve Fit: Quadratic


Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015

4-Nitrophenol

Response Ratio




Resp Ratio = $2.75 \times 10^{-1} \times \text{Amt} - 8.97 \times 10^{-2}$
Coef of Det (r^2) = 0.993 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015

2,4-Dinitrotoluene

Response Ratio




Resp Ratio = $4.27 \times 10^{-1} \times \text{Amt} - 6.28 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015

2,3,4,6-Tetrachlorophenol

Response Ratio




$R = 3.72e-002 A^2 + 1.72e-001 A + 2.99e-002$
Coef of Det (r^2) = 0.996 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015

4-Nitroaniline


Response Ratio



$R = 2.87e-002 A^2 + 2.63e-001 A - 1.04e-002$
Coef of Det (r^2) = 0.997 Curve Fit: Quadratic


Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015

4,6-Dinitro-2-methylphenol




$R = 2.78e-002 A^2 + 2.28e-002 A + 1.93e-002$
Coef of Det (r^2) = 0.996 Curve Fit: Quadratic


Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015




Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015


Bis(2-ethylhexyl)phthalate

Response Ratio



$R = 1.55e-001 A^2 + 2.72e-001 A + 5.77e-002$
Coef of Det (r^2) = 0.993 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
Calibration Table Last Updated: Sat Jul 18 00:40:00 2015