


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF071715\
 Data File : BF080534.D
 Acq On : 17 Jul 2015 15:23
 Operator : TP/UM
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071715.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jul 20 13:29:49 2015



AutoFind: Scans 853, 854, 855; Background Corrected with Scan 851

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	62.9	4259	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	49.5	3349	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	61.9	4187	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	6766	PASS
199	198	5	9	7.4	501	PASS
275	198	10	60	28.6	1938	PASS
365	198	1	100	2.4	164	PASS
441	442	0.01	24	16.3	958	PASS
442	198	50	100	87.1	5893	PASS
443	442	15	24	19.3	1136	PASS