

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071718\
 Data File : BF107451.D
 Acq On : 17 Jul 2018 16:04
 Operator : JU/SJ
 Sample : J3832-03 2X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TR-04-071618

Manual Integrations
 APPROVED

Sohil
 7/18/2018 10:44:06 AM

Quant Time: Jul 17 18:59:22 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 17 11:29:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.98	152	138716	20.00	ng	0.00
21) Naphthalene-d8	8.27	136	566787	20.00	ng	0.00
38) Acenaphthene-d10	10.03	164	282438	20.00	ng	0.00
63) Phenanthrene-d10	11.52	188	524980	20.00	ng	0.00
75) Chrysene-d12	14.18	240	413274	20.00	ng	0.00
86) Perylene-d12	15.72	264	290659	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	324062	35.48	ng	0.00
7) Phenol-d6	6.61	99	420291	35.59	ng	0.00
23) Nitrobenzene-d5	7.54	82	282974	26.62	ng	-0.01
41) 2,4,6-Tribromophenol	10.82	330	88281	35.57	ng	0.00
44) 2-Fluorobiphenyl	9.34	172	496822	21.19	ng	0.00
78) Terphenyl-d14	13.11	244	450426	27.08	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.74	163	110553	4.926	ng	Qvalue # 96
70) Phenanthrene	11.55	178	93244	3.564	ng	97
74) Fluoranthene	12.74	202	132774	4.465	ng	95
77) Pyrene	12.98	202	152019	5.409	ng	98
80) Benzo(a)anthracene	14.17	228	77102m	3.049	ng	
82) Chrysene	14.20	228	80660	3.339	ng	97
87) Benzo(b)fluoranthene	15.25	252	76451m	4.494	ng	
89) Benzo(a)pyrene	15.65	252	46556	2.981	ng	# 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071718\
Data File : BF107451.D
Acq On : 17 Jul 2018 16:04
Operator : JU/SJ
Sample : J3832-03 2X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TR-04-071618

Manual Integrations
APPROVED

Sohil
7/18/2018 10:44:06 AM

Quant Time: Jul 17 18:59:22 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 17 11:29:08 2018
Response via : Initial Calibration

