

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071720\
 Data File : BF120914.D
 Acq On : 17 Jul 2020 10:51
 Operator : JU/CG
 Sample : PB130249TB
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB130249TB

Quant Time: Jul 17 11:20:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 16 14:20:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	173101	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	656738	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	333957	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	627099	20.00	ng	0.00
76) Chrysene-d12	13.97	240	503593	20.00	ng	0.00
86) Perylene-d12	15.42	264	463318	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	1129017	106.92	ng	0.01
7) Phenol-d6	6.45	99	1330244	97.53	ng	0.00
23) Nitrobenzene-d5	7.37	82	839413	73.96	ng	0.00
42) 2,4,6-Tribromophenol	10.64	330	376143	102.90	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	1531990	68.49	ng	0.00
79) Terphenyl-d14	12.92	244	1507695	65.08	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071720\
Data File : BF120914.D
Acq On : 17 Jul 2020 10:51
Operator : JU/CG
Sample : PB130249TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB130249TB

Quant Time: Jul 17 11:20:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 16 14:20:32 2020
Response via : Initial Calibration

