

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071917\
 Data File : BF096888.D
 Acq On : 19 Jul 2017 23:44
 Operator : SJ/JU
 Sample : I4258-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OK-02-071817-A

Manual Integrations
 APPROVED

Sohil
 7/20/2017 3:05:13 PM

Quant Time: Jul 20 16:30:04 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF071317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 18 13:33:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.59	152	117156	20.00	ng	-0.01
21) Naphthalene-d8	7.87	136	394077	20.00	ng	-0.02
38) Acenaphthene-d10	9.62	164	176096	20.00	ng	-0.02
63) Phenanthrene-d10	11.09	188	260429	20.00	ng	-0.01
75) Chrysene-d12	13.72	240	210679	20.00	ng	-0.01
86) Perylene-d12	15.10	264	124353	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	726094	93.19	ng	0.00
7) Phenol-d6	6.24	99	881453	94.27	ng	-0.01
23) Nitrobenzene-d5	7.15	82	494287	77.69	ng	-0.02
41) 2,4,6-Tribromophenol	10.40	330	136696	90.94	ng	-0.02
44) 2-Fluorobiphenyl	8.95	172	747251	67.04	ng	-0.01
78) Terphenyl-d14	12.68	244	557827	45.92	ng	-0.01
Target Compounds						
10) Phenol	6.25	94	39668	3.753	ng	Qvalue 100
49) Dimethylphthalate	9.35	163	132882	10.022	ng	97
74) Fluoranthene	12.30	202	33486	2.471	ng	97
83) Bis(2-ethylhexyl)phthalate	13.73	149	42115	3.807	ng	# 96
85) Indeno(1,2,3-cd)pyrene	16.38	276	6189	3.826	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF071917\
Data File : BF096888.D
Acq On : 19 Jul 2017 23:44
Operator : SJ/JU
Sample : I4258-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OK-02-071817-A

Manual Integrations
APPROVED

Sohil
7/20/2017 3:05:13 PM

Quant Time: Jul 20 16:30:04 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF071317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 18 13:33:22 2017
Response via : Initial Calibration

