


Data Path : Z:\HPCHEM1\BNA F\Data\BF071917\
 Data File : BF096867.D
 Acq On : 19 Jul 2017 13:52
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 18 13:33:22 2017



AutoFind: Scans 1595, 1596, 1597; Background Corrected with Scan 1590

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.8	27786	PASS
68	69	0.00	2	0.9	184	PASS
69	198	0.00	100	37.8	21080	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.1	27934	PASS
197	198	0.00	2	0.3	176	PASS
198	198	100	100	100.0	55762	PASS
199	198	5	9	6.5	3633	PASS
275	198	10	60	21.9	12204	PASS
365	198	1	100	2.6	1429	PASS
441	442	0.01	24	16.1	8819	PASS
442	198	50	100	98.5	54901	PASS
443	442	15	24	18.3	10046	PASS