


Data Path : Z:\HPCHEM1\BNA F\Data\BF071917\
 Data File : BF096892.D
 Acq On : 20 Jul 2017 2:31
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 18 13:33:22 2017



AutoFind: Scans 1595, 1596, 1597; Background Corrected with Scan 1591

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	57.5	22053	PASS
68	69	0.00	2	0.7	104	PASS
69	198	0.00	100	39.1	14995	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.8	20272	PASS
197	198	0.00	2	0.6	227	PASS
198	198	100	100	100.0	38384	PASS
199	198	5	9	6.6	2522	PASS
275	198	10	60	20.0	7679	PASS
365	198	1	100	2.4	930	PASS
441	442	0.01	24	15.8	4993	PASS
442	198	50	100	82.2	31537	PASS
443	442	15	24	18.9	5960	PASS