

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071921\
 Data File : BF124615.D
 Acq On : 19 Jul 2021 15:51
 Operator : JU/CG
 Sample : PB137770BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB137770BL

Quant Time: Jul 19 16:13:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 19 11:33:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.916	152	10275	20.00	ng	0.00
21) Naphthalene-d8	8.192	136	40347	20.00	ng	-0.01
39) Acenaphthene-d10	9.951	164	22307	20.00	ng	0.00
64) Phenanthrene-d10	11.439	188	38642	20.00	ng	0.00
76) Chrysene-d12	14.080	240	23504	20.00	ng	0.00
86) Perylene-d12	15.574	264	17483	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.557	112	79000	133.21	ng	0.02
7) Phenol-d6	6.540	99	90010	124.76	ng	0.00
23) Nitrobenzene-d5	7.475	82	57285	89.63	ng	-0.01
42) 2,4,6-Tribromophenol	10.745	330	23233	133.84	ng	0.00
45) 2-Fluorobiphenyl	9.275	172	131490	85.96	ng	0.00
79) Terphenyl-d14	13.039	244	130904	93.52	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071921\
Data File : BF124615.D
Acq On : 19 Jul 2021 15:51
Operator : JU/CG
Sample : PB137770BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB137770BL

Quant Time: Jul 19 16:13:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 19 11:33:09 2021
Response via : Initial Calibration

