

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072016\
 Data File : BF089155.D
 Acq On : 21 Jul 2016 2:51
 Operator : UM/SJ
 Sample : H4112-09
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 KSB-06-10.0-12.0

Quant Time: Jul 21 03:37:22 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 20 19:03:15 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.60	152	52986	20.00	ng	0.00
21) Naphthalene-d8	7.88	136	233523	20.00	ng	0.00
38) Acenaphthene-d10	9.62	164	105328	20.00	ng	-0.01
63) Phenanthrene-d10	11.10	188	210792	20.00	ng	-0.01
75) Chrysene-d12	13.71	240	166848	20.00	ng	-0.01
86) Perylene-d12	15.06	264	126360	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	368659	105.82	ng	0.00
7) Phenol-d6	6.26	99	515965	114.84	ng	-0.01
23) Nitrobenzene-d5	7.16	82	295690	67.28	ng	-0.01
41) 2,4,6-Tribromophenol	10.42	330	106215	112.01	ng	0.00
44) 2-Fluorobiphenyl	8.96	172	509156	66.60	ng	-0.01
78) Terphenyl-d14	12.69	244	478238	62.17	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.36	163	133324	16.50	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072016\
Data File : BF089155.D
Acq On : 21 Jul 2016 2:51
Operator : UM/SJ
Sample : H4112-09
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
KSB-06-10.0-12.0

Quant Time: Jul 21 03:37:22 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 20 19:03:15 2016
Response via : Initial Calibration

