


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $2.78 \times 10^{-1} \times \text{Amt} - 1.18 \times 10^{-1}$
Coef of Det (r^2) = 0.995 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
Calibration Table Last Updated: Wed Jul 20 19:03:15 2016


2,4-Dinitrophenol

Response Ratio



Resp Ratio = $1.98 \times 10^{-1} \times \text{Amt} - 8.44 \times 10^{-2}$
Coef of Det (r^2) = 0.995 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
Calibration Table Last Updated: Wed Jul 20 19:03:15 2016



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
Calibration Table Last Updated: Wed Jul 20 19:03:15 2016