


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072016\
 Data File : BF089137.D
 Acq On : 20 Jul 2016 17:54
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 20 19:03:15 2016



AutoFind: Scans 856, 857, 858; Background Corrected with Scan 853

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.6	6699	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.5	7418	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.9	9692	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	18310	PASS
199	198	5	9	6.4	1167	PASS
275	198	10	60	24.4	4464	PASS
365	198	1	100	2.1	383	PASS
441	442	0.01	24	16.5	1825	PASS
442	198	50	100	60.5	11086	PASS
443	442	15	24	18.3	2030	PASS