

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072017\
 Data File : BF096926.D
 Acq On : 21 Jul 2017 00:48
 Operator : SJ/JU
 Sample : I4319-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-2(0-6)COMP

Manual Integrations
 APPROVED

Sohil
 7/21/2017 1:06:02 PM

Quant Time: Jul 21 08:11:45 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF071317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 18 13:33:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.57	152	102510	20.00	ng	-0.03
21) Naphthalene-d8	7.85	136	493861	20.00	ng	-0.04
38) Acenaphthene-d10	9.60	164	195200	20.00	ng	-0.04
63) Phenanthrene-d10	11.07	188	311748	20.00	ng	-0.03
75) Chrysene-d12	13.70	240	233862	20.00	ng	-0.03
86) Perylene-d12	15.07	264	210695	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.17	112	775630	113.77	ng	-0.02
7) Phenol-d6	6.23	99	958918	117.21	ng	-0.02
23) Nitrobenzene-d5	7.13	82	536149	67.24	ng	-0.04
41) 2,4,6-Tribromophenol	10.39	330	182071	109.27	ng	-0.04
44) 2-Fluorobiphenyl	8.93	172	896047	72.52	ng	-0.03
78) Terphenyl-d14	12.66	244	603392	44.75	ng	-0.03
Target Compounds						Qvalue
10) Phenol	6.23	94	39027	4.220	ng	85
49) Dimethylphthalate	9.32	163	126733	8.623	ng	98
70) Phenanthrene	11.10	178	42872	2.529	ng	98
74) Fluoranthene	12.28	202	46655	2.876	ng	99
77) Pyrene	12.50	202	57064	2.689	ng	97
85) Indeno(1,2,3-cd)pyrene	16.34	276	10491	4.089	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072017\
Data File : BF096926.D
Acq On : 21 Jul 2017 00:48
Operator : SJ/JU
Sample : I4319-01
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-2(0-6)COMP

Manual Integrations
APPROVED

Sohil
7/21/2017 1:06:02 PM

Quant Time: Jul 21 08:11:45 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF071317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 18 13:33:22 2017
Response via : Initial Calibration

