


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072020\
 Data File : BF120925.D
 Acq On : 20 Jul 2020 8:49
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071620.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 16 14:20:32 2020



AutoFind: Scans 1616, 1617, 1618; Background Corrected with Scan 1611

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.1	63250	PASS
68	69	0.00	2	1.8	1090	PASS
69	198	0.00	100	32.2	59728	PASS
70	69	0.00	2	0.6	364	PASS
127	198	10	80	47.4	87952	PASS
197	198	0.00	2	0.9	1744	PASS
198	198	100	100	100.0	185384	PASS
199	198	5	9	6.8	12664	PASS
275	198	10	60	28.5	52856	PASS
365	198	1	100	3.5	6523	PASS
441	198	0.01	100	13.5	25071	PASS
442	198	50	100	84.9	157458	PASS
443	442	15	24	18.8	29670	PASS