

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072116\  
 Data File : BF089180.D  
 Acq On : 21 Jul 2016 15:12  
 Operator : UM/SJ  
 Sample : H4121-04  
 Misc :  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 TEC-EDER2-COMP

Quant Time: Jul 22 01:35:32 2016  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072016.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jul 20 19:03:15 2016  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4      | 6.60  | 152  | 48309    | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8             | 7.88  | 136  | 210634   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10           | 9.63  | 164  | 82836    | 20.00 | ng    | 0.00     |
| 63) Phenanthrene-d10           | 11.11 | 188  | 134472   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12               | 13.73 | 240  | 98726    | 20.00 | ng    | 0.00     |
| 86) Perylene-d12               | 15.07 | 264  | 82605    | 20.00 | ng    | 0.00     |
| System Monitoring Compounds    |       |      |          |       |       |          |
| 5) 2-Fluorophenol              | 5.19  | 112  | 189799   | 59.75 | ng    | 0.00     |
| 7) Phenol-d6                   | 6.26  | 99   | 279024   | 68.12 | ng    | -0.01    |
| 23) Nitrobenzene-d5            | 7.16  | 82   | 176380   | 44.49 | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol       | 10.41 | 330  | 34398    | 46.12 | ng    | -0.01    |
| 44) 2-Fluorobiphenyl           | 8.96  | 172  | 320005   | 53.23 | ng    | -0.01    |
| 78) Terphenyl-d14              | 12.69 | 244  | 220688   | 48.48 | ng    | 0.00     |
| Target Compounds               |       |      |          |       |       |          |
| 49) Dimethylphthalate          | 9.36  | 163  | 56662    | 8.91  | ng    | 98       |
| 83) Bis(2-ethylhexyl)phthalate | 13.73 | 149  | 9830     | 2.15  | ng    | # 96     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072116\  
Data File : BF089180.D  
Acq On : 21 Jul 2016 15:12  
Operator : UM/SJ  
Sample : H4121-04  
Misc :  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
TEC-EDER2-COMP

Quant Time: Jul 22 01:35:32 2016  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072016.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jul 20 19:03:15 2016  
Response via : Initial Calibration

