

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072117\
 Data File : BF096942.D
 Acq On : 21 Jul 2017 14:17
 Operator : SJ/JU
 Sample : I4253-07 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-07-BASE

Manual Integrations
 APPROVED

Sohil
 7/24/2017 3:05:09 PM

Quant Time: Jul 24 11:24:48 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF071317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 18 13:33:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.55	152	94946	20.00	ng	-0.05
21) Naphthalene-d8	7.83	136	453397	20.00	ng	-0.05
38) Acenaphthene-d10	9.58	164	192553	20.00	ng	-0.05
63) Phenanthrene-d10	11.06	188	312061	20.00	ng	-0.05
75) Chrysene-d12	13.68	240	206168	20.00	ng	-0.05
86) Perylene-d12	15.04	264	189619	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.15	112	76901	12.18	ng	-0.05
7) Phenol-d6	6.20	99	90682	11.97	ng	-0.05
23) Nitrobenzene-d5	7.12	82	56171	7.67	ng	-0.05
41) 2,4,6-Tribromophenol	10.37	330	19929	12.13	ng	-0.05
44) 2-Fluorobiphenyl	8.91	172	92924	7.62	ng	-0.05
78) Terphenyl-d14	12.64	244	58377	4.91	ng	-0.05
Target Compounds						
11) bis(2-Chloroethyl)ether	6.29	93	29017	4.504	ng	86
14) 1,2-Dichlorobenzene	6.72	146	18951	2.841	ng	95
31) Naphthalene	7.86	128	123208	5.736	ng	99
37) 2-Methylnaphthalene	8.55	142	136544	9.972	ng	98
59) Diethylphthalate	10.01	149	33043	2.341	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072117\
Data File : BF096942.D
Acq On : 21 Jul 2017 14:17
Operator : SJ/JU
Sample : I4253-07 10X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-07-BASE

Manual Integrations
APPROVED

Sohil
7/24/2017 3:05:09 PM

Quant Time: Jul 24 11:24:48 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF071317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 18 13:33:22 2017
Response via : Initial Calibration

