


Data Path : Z:\HPCHEM1\BNA F\Data\BF072117\
 Data File : BF096962.D
 Acq On : 22 Jul 2017 00:38
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF071317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jul 24 14:10:38 2017



AutoFind: Scans 1589, 1590, 1591; Background Corrected with Scan 1585

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.3	21487	PASS
68	69	0.00	2	0.6	105	PASS
69	198	0.00	100	40.6	16352	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.7	21239	PASS
197	198	0.00	2	0.8	331	PASS
198	198	100	100	100.0	40288	PASS
199	198	5	9	6.9	2779	PASS
275	198	10	60	21.6	8719	PASS
365	198	1	100	2.3	920	PASS
441	442	0.01	24	15.2	5667	PASS
442	198	50	100	92.3	37166	PASS
443	442	15	24	17.7	6560	PASS