


Data Path : Z:\HPCHEM1\BNA\_F\Data\BF072216\  
 Data File : BF089186.D  
 Acq On : 22 Jul 2016 9:26  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF072016.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Jul 20 19:03:15 2016



AutoFind: Scans 855, 856, 857; Background Corrected with Scan 852

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 38.6      | 7196    | PASS             |
| 68          | 69           | 0.00         | 2            | 1.3       | 106     | PASS             |
| 69          | 198          | 0.00         | 100          | 43.3      | 8070    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 55.1      | 10258   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 18621   | PASS             |
| 199         | 198          | 5            | 9            | 6.6       | 1237    | PASS             |
| 275         | 198          | 10           | 60           | 25.0      | 4647    | PASS             |
| 365         | 198          | 1            | 100          | 2.2       | 404     | PASS             |
| 441         | 442          | 0.01         | 24           | 16.2      | 1892    | PASS             |
| 442         | 198          | 50           | 100          | 62.6      | 11663   | PASS             |
| 443         | 442          | 15           | 24           | 18.5      | 2155    | PASS             |