

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072221\
 Data File : BF124690.D
 Acq On : 22 Jul 2021 15:23
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF072221

Quant Time: Jul 22 15:52:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 15:28:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	41.699	-4.2	96	0.00
3	Pyridine	40.000	42.800	-7.0	97	-0.01
4	n-Nitrosodimethylamine	40.000	40.802	-2.0	96	-0.01
5 S	2-Fluorophenol	80.000	80.531	-0.7	98	0.00
6	Aniline	40.000	40.834	-2.1	100	0.00
7 S	Phenol-d6	80.000	80.585	-0.7	98	0.00
8	2-Chlorophenol	40.000	40.777	-1.9	101	0.00
9	Benzaldehyde	40.000	37.826	5.4	99	0.00
10 C	Phenol	40.000	41.239	-3.1	104	0.00
11	bis(2-Chloroethyl)ether	40.000	41.300	-3.2	103	0.00
12	1,3-Dichlorobenzene	40.000	39.247	1.9	96	0.00
13 C	1,4-Dichlorobenzene	40.000	40.376	-0.9	98	0.00
14	1,2-Dichlorobenzene	40.000	40.266	-0.7	99	0.00
15	Benzyl Alcohol	40.000	44.030	-10.1	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.087	-0.2	98	0.00
17	2-Methylphenol	40.000	41.158	-2.9	100	0.00
18	Hexachloroethane	40.000	41.877	-4.7	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.056	-5.1	100	0.00
20	3+4-Methylphenols	40.000	40.730	-1.8	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	39.465	1.3	99	0.00
23 S	Nitrobenzene-d5	80.000	80.264	-0.3	100	0.00
24	Nitrobenzene	40.000	38.227	4.4	97	0.00
25	Isophorone	40.000	39.425	1.4	99	0.00
26 C	2-Nitrophenol	40.000	39.776	0.6	96	0.00
27	2,4-Dimethylphenol	40.000	40.520	-1.3	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.496	1.3	101	0.00
29 C	2,4-Dichlorophenol	40.000	40.267	-0.7	101	0.00
30	1,2,4-Trichlorobenzene	40.000	39.342	1.6	98	0.00
31	Naphthalene	40.000	39.171	2.1	99	0.00
32	Benzoic acid	40.000	40.812	-2.0	99	0.00
33	4-Chloroaniline	40.000	39.022	2.4	99	0.00
34 C	Hexachlorobutadiene	40.000	38.309	4.2	97	0.00
35	Caprolactam	40.000	42.411	-6.0	109	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.045	-2.6	97	0.00
37	2-Methylnaphthalene	40.000	39.594	1.0	100	0.00
38	1-Methylnaphthalene	40.000	38.585	3.5	98	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.749	-1.9	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.995	-2.5	93	0.00
42 S	2,4,6-Tribromophenol	80.000	84.183	-5.2	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.783	-4.5	101	0.00
44	2,4,5-Trichlorophenol	40.000	42.631	-6.6	101	0.00
45 S	2-Fluorobiphenyl	80.000	79.889	0.1	99	0.00
46	1,1'-Biphenyl	40.000	40.533	-1.3	100	0.00
47	2-Chloronaphthalene	40.000	41.210	-3.0	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072221\
 Data File : BF124690.D
 Acq On : 22 Jul 2021 15:23
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF072221

Quant Time: Jul 22 15:52:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 15:28:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.268	-0.7	96	0.00
49	Acenaphthylene	40.000	41.711	-4.3	102	0.00
50	Dimethylphthalate	40.000	40.640	-1.6	98	0.00
51	2,6-Dinitrotoluene	40.000	41.614	-4.0	101	0.00
52 C	Acenaphthene	40.000	41.145	-2.9	99	0.00
53	3-Nitroaniline	40.000	41.576	-3.9	98	0.00
54 P	2,4-Dinitrophenol	40.000	39.640	0.9	96	0.00
55	Dibenzofuran	40.000	40.840	-2.1	99	0.00
56 P	4-Nitrophenol	40.000	42.846	-7.1	102	0.00
57	2,4-Dinitrotoluene	40.000	41.942	-4.9	100	0.00
58	Fluorene	40.000	40.430	-1.1	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.287	-8.2	103	0.00
60	Diethylphthalate	40.000	39.765	0.6	98	0.00
61	4-Chlorophenyl-phenylether	40.000	39.702	0.7	100	0.00
62	4-Nitroaniline	40.000	40.755	-1.9	98	0.00
63	Azobenzene	40.000	41.007	-2.5	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.003	-2.5	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.784	0.5	101	0.00
67	4-Bromophenyl-phenylether	40.000	39.500	1.3	102	0.00
68	Hexachlorobenzene	40.000	40.633	-1.6	104	0.00
69	Atrazine	40.000	39.499	1.3	101	0.00
70 C	Pentachlorophenol	40.000	41.847	-4.6	101	0.00
71	Phenanthrene	40.000	40.237	-0.6	101	0.00
72	Anthracene	40.000	39.767	0.6	101	0.00
73	Carbazole	40.000	39.559	1.1	101	0.00
74	Di-n-butylphthalate	40.000	40.230	-0.6	100	0.00
75 C	Fluoranthene	40.000	39.906	0.2	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	39.970	0.1	96	0.00
78	Pyrene	40.000	38.394	4.0	101	0.00
79 S	Terphenyl-d14	80.000	77.712	2.9	102	0.00
80	Butylbenzylphthalate	40.000	39.077	2.3	102	0.00
81	Benzo(a)anthracene	40.000	40.297	-0.7	107	0.00
82	3,3'-Dichlorobenzidine	40.000	37.927	5.2	99	0.00
83	Chrysene	40.000	40.274	-0.7	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.800	0.5	104	0.00
85 c	Di-n-octyl phthalate	40.000	40.625	-1.6	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	111	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.975	0.1	114	0.00
88	Benzo(b)fluoranthene	40.000	37.785	5.5	106	0.00
89	Benzo(k)fluoranthene	40.000	41.219	-3.0	114	0.00
90 C	Benzo(a)pyrene	40.000	39.875	0.3	112	0.00
91	Dibenzo(a,h)anthracene	40.000	38.942	2.6	112	0.00
92	Benzo(g,h,i)perylene	40.000	38.255	4.4	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072221\
Data File : BF124690.D
Acq On : 22 Jul 2021 15:23
Operator : JU/CG
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF072221

Quant Time: Jul 22 15:52:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 22 15:28:57 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0