

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072222\
 Data File : BF129326.D
 Acq On : 22 Jul 2022 20:54
 Operator : CG\JU
 Sample : N3814-05 5X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ENV-DE-3(5-10)

Quant Time: Jul 23 01:24:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 20 13:42:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.893	152	137745	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	496467	20.000	ng	0.00
39) Acenaphthene-d10	9.934	164	218069	20.000	ng	0.00
64) Phenanthrene-d10	11.422	188	513898	20.000	ng	0.00
76) Chrysene-d12	14.074	240	391745	20.000	ng	0.00
86) Perylene-d12	15.568	264	345687	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	63251	7.490	ng	-0.01
7) Phenol-d6	6.510	99	79627	7.396	ng	-0.02
23) Nitrobenzene-d5	7.451	82	48034	5.205	ng	-0.01
42) 2,4,6-Tribromophenol	10.722	330	20786	9.880	ng	-0.01
45) 2-Fluorobiphenyl	9.245	172	75236	5.307	ng	-0.01
79) Terphenyl-d14	13.004	244	102120	4.855	ng	-0.01
Target Compounds						Qvalue
25) Isophorone	7.451	82	48033	2.965	ng	# 65
31) Naphthalene	8.198	128	135257	5.312	ng	99
37) 2-Methylnaphthalene	8.887	142	42101	2.552	ng	100
38) 1-Methylnaphthalene	8.986	142	40382	2.530	ng	99
49) Acenaphthylene	9.792	152	41895	2.114	ng	98
51) 2,6-Dinitrotoluene	9.934	165	30666	9.141	ng	# 26
52) Acenaphthene	9.963	154	181014	14.072	ng	98
55) Dibenzofuran	10.139	168	123321	6.899	ng	99
58) Fluorene	10.481	166	184356	12.922	ng	97
71) Phenanthrene	11.451	178	1928627	68.376	ng	99
72) Anthracene	11.498	178	622858	22.286	ng	100
73) Carbazole	11.651	167	265236	10.174	ng	99
75) Fluoranthene	12.651	202	2775872	96.564	ng	99
78) Pyrene	12.880	202	2855230	85.217	ng	99
81) Benzo(a)anthracene	14.063	228	1633928	60.571	ng	98
83) Chrysene	14.104	228	1356545	53.279	ng	97
87) Indeno(1,2,3-cd)pyrene	17.080	276	601324	25.374	ng	# 93
88) Benzo(b)fluoranthene	15.133	252	2032202	88.325	ng	98
89) Benzo(k)fluoranthene	15.133	252	2032202	90.964	ng	98
90) Benzo(a)pyrene	15.510	252	1128292	60.607	ng	96
91) Dibenzo(a,h)anthracene	17.257	278	106359	5.568	ng	95
92) Benzo(g,h,i)perylene	17.545	276	372108	18.748	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072222\
Data File : BF129326.D
Acq On : 22 Jul 2022 20:54
Operator : CG\JU
Sample : N3814-05 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)

Quant Time: Jul 23 01:24:14 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 20 13:42:18 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.893 min Scan# 71
Delta R.T. -0.005 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion:152 Resp: 137745
Ion Ratio Lower Upper
152 100
150 154.2 123.7 185.5
115 63.9 49.9 74.9



#5
2-Fluorophenol
Concen: 7.490 ng
RT: 5.504 min Scan# 547
Delta R.T. -0.012 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54

Tgt Ion:112 Resp: 63251
Ion Ratio Lower Upper
112 100
64 59.5 53.3 79.9
63 35.7 28.0 42.0





#7

Phenol-d6

Concen: 7.396 ng

RT: 6.510 min Scan# 71

Delta R.T. -0.017 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54

Instrument :

BNA_F

ClientSampleId :

ENV-DE-3(5-10)





Tgt Ion: 99 Resp: 79627

Ion Ratio Lower Upper

99 100

42 20.9 17.3 25.9

71 37.9 28.9 43.3



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.175 min Scan# 1001

Delta R.T. -0.006 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54

Tgt Ion: 136 Resp: 496467




Ion Ratio Lower Upper


136 100

137 11.1 8.8 13.2

54 8.2 7.5 11.3

68 4.9 4.7 7.1









#23
Nitrobenzene-d5
Concen: 5.205 ng
RT: 7.451 min Scan# 81
Delta R.T. -0.012 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion: 82 Resp: 48034
Ion Ratio Lower Upper
82 100
128 41.9 34.0 51.0
54 53.0 46.9 70.3



#25
Isophorone
Concen: 2.965 ng
RT: 7.451 min Scan# 878
Delta R.T. -0.265 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54

Tgt Ion: 82 Resp: 48033
Ion Ratio Lower Upper
82 100
95 0.0 5.7 8.5#
138 0.0 14.2 21.2#









#31
Naphthalene
Concen: 5.312 ng
RT: 8.198 min Scan# 1005
Delta R.T. -0.006 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion:128 Resp: 135257
Ion Ratio Lower Upper
128 100
129 10.4 9.2 13.8
127 13.1 10.5 15.7



#37
2-Methylnaphthalene
Concen: 2.552 ng
RT: 8.887 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54

Tgt Ion:142 Resp: 42101
Ion Ratio Lower Upper
142 100
141 88.6 71.0 106.4





#38

1-Methylnaphthalene

Concen: 2.530 ng

RT: 8.986 min Scan# 1139

Delta R.T. -0.006 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54

Instrument :

BNA_F

ClientSampleId :





ENV-DE-3(5-10)

Tgt Ion:142 Resp: 40382

Ion Ratio Lower Upper

142 100

141 93.8 74.2 111.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.934 min Scan# 1300

Delta R.T. -0.006 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54




Tgt Ion:164 Resp: 218069


Ion Ratio Lower Upper

164 100

162 101.9 79.5 119.3

160 45.0 34.2 51.4





#42

2,4,6-Tribromophenol

Concen: 9.880 ng

RT: 10.722 min Scan# 1436

Delta R.T. -0.012 min

Lab File: BF129326.D


Acq: 22 Jul 2022 20:54

Instrument :

BNA_F

ClientSampleId :

ENV-DE-3(5-10)






Tgt Ion:330 Resp: 20786

Ion Ratio Lower Upper

330 100

332 96.3 76.8 115.2

141 44.5 28.0 42.0



#45

2-Fluorobiphenyl

Concen: 5.307 ng

RT: 9.245 min Scan# 1183

Delta R.T. -0.012 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54




Tgt Ion:172 Resp: 75236


Ion Ratio Lower Upper

172 100

171 36.4 28.8 43.2

170 23.1 19.8 29.8









#49
Acenaphthylene
Concen: 2.114 ng
RT: 9.792 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion:152 Resp: 41895
Ion Ratio Lower Upper
152 100
151 19.7 16.2 24.4
153 15.1 10.7 16.1



#51
2,6-Dinitrotoluene
Concen: 9.141 ng
RT: 9.934 min Scan# 1300
Delta R.T. 0.212 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54


Tgt Ion:165 Resp: 30666
Ion Ratio Lower Upper
165 100
63 1.6 44.9 67.3#
89 1.4 43.6 65.4#








#52
Acenaphthene
Concen: 14.072 ng
RT: 9.963 min Scan# 1305
Delta R.T. -0.012 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54


Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)





Tgt Ion:154 Resp: 181014
Ion Ratio Lower Upper
154 100
153 115.3 89.6 134.4
152 53.2 42.4 63.6




#55
Dibenzofuran
Concen: 6.899 ng
RT: 10.139 min Scan# 1335
Delta R.T. -0.006 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54



Tgt Ion:168 Resp: 123321
Ion Ratio Lower Upper
168 100
139 38.6 30.6 46.0
169 13.6 11.0 16.6









#58
Fluorene
Concen: 12.922 ng
RT: 10.481 min Scan# 1393
Delta R.T. -0.012 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion:166 Resp: 184356
Ion Ratio Lower Upper
166 100
165 100.4 78.1 117.1
167 13.9 10.5 15.7



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.422 min Scan# 1553
Delta R.T. -0.006 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54

Tgt Ion:188 Resp: 513898
Ion Ratio Lower Upper
188 100
94 10.0 7.0 10.6
80 11.1 7.6 11.4









#71
Phenanthrene
Concen: 68.376 ng
RT: 11.451 min Scan# 1558
Delta R.T. -0.006 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion:178 Resp: 1928627
Ion Ratio Lower Upper
178 100
176 19.9 16.0 24.0
179 16.3 12.6 19.0



#72
Anthracene
Concen: 22.286 ng
RT: 11.498 min Scan# 1566
Delta R.T. -0.012 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54

Tgt Ion:178 Resp: 622858
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.8 12.6 19.0





#73

Carbazole

Concen: 10.174 ng

RT: 11.651 min Scan# 1

Delta R.T. -0.012 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54

Instrument :

BNA_F

ClientSampleId :

ENV-DE-3(5-10)





Tgt Ion:167 Resp: 265236

Ion Ratio Lower Upper

167 100

166 21.1 17.7 26.5

139 14.0 11.2 16.8



#75

Fluoranthene

Concen: 96.564 ng

RT: 12.651 min Scan# 1762

Delta R.T. 0.006 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54




Tgt Ion:202 Resp: 2775872


Ion Ratio Lower Upper

202 100

101 12.3 0.0 31.4

203 18.3 0.0 38.0









#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 20
Delta R.T. -0.006 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion:240 Resp: 391745
Ion Ratio Lower Upper
240 100
120 13.2 8.9 13.3
236 25.7 20.8 31.2



#78
Pyrene
Concen: 85.217 ng
RT: 12.880 min Scan# 1801
Delta R.T. 0.000 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54

Tgt Ion:202 Resp: 2855230
Ion Ratio Lower Upper
202 100
200 22.1 17.7 26.5
203 18.9 14.6 21.8









#79
 Terphenyl-d14
 Concen: 4.855 ng
 RT: 13.004 min Scan# 11
 Delta R.T. -0.012 min
 Lab File: BF129326.D
 Acq: 22 Jul 2022 20:54




Instrument :
 BNA_F
 ClientSampleId :
 ENV-DE-3(5-10)


Tgt Ion:244 Resp: 102120
 Ion Ratio Lower Upper
 244 100
 212 7.7 6.1 9.1
 122 13.8 8.8 13.2#



#81
 Benzo(a)anthracene
 Concen: 60.571 ng
 RT: 14.063 min Scan# 2002
 Delta R.T. -0.006 min
 Lab File: BF129326.D
 Acq: 22 Jul 2022 20:54

Tgt Ion:228 Resp: 1633928
 Ion Ratio Lower Upper
 228 100
 226 28.7 22.1 33.1
 229 21.2 15.8 23.8





#83

Chrysene

Concen: 53.279 ng

RT: 14.104 min Scan# 2009

Delta R.T. 0.000 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54

Instrument :

BNA_F

ClientSampleId :

ENV-DE-3(5-10)





Tgt Ion:228 Resp: 1356545

Ion Ratio Lower Upper

228 100

226 31.1 24.3 36.5

229 22.6 15.9 23.9



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.568 min Scan# 2258

Delta R.T. 0.000 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54




Tgt Ion:264 Resp: 345687


Ion Ratio Lower Upper

264 100

260 24.3 18.4 27.6

265 23.6 17.2 25.8









#87
Indeno(1,2,3-cd)pyrene
Concen: 25.374 ng
RT: 17.080 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion:276 Resp: 601324
Ion Ratio Lower Upper
276 100
138 20.7 21.1 31.7#
277 24.0 20.3 30.5



#88
Benzo(b)fluoranthene
Concen: 88.325 ng
RT: 15.133 min Scan# 2184
Delta R.T. 0.006 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54

Tgt Ion:252 Resp: 2032202
Ion Ratio Lower Upper
252 100
253 22.8 17.7 26.5
125 11.6 8.0 12.0









#89
Benzo(k)fluoranthene
Concen: 90.964 ng
RT: 15.133 min Scan# 2189
Delta R.T. -0.029 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54




Instrument :
BNA_F
ClientSampleId :
ENV-DE-3(5-10)


Tgt Ion:252 Resp: 2032202
Ion Ratio Lower Upper
252 100
253 22.8 17.8 26.6
125 11.6 8.3 12.5



#90
Benzo(a)pyrene
Concen: 60.607 ng
RT: 15.510 min Scan# 2248
Delta R.T. 0.006 min
Lab File: BF129326.D
Acq: 22 Jul 2022 20:54

Tgt Ion:252 Resp: 1128292
Ion Ratio Lower Upper
252 100
253 23.7 17.0 25.6
125 11.7 8.8 13.2





#91

Dibenzo(a,h)anthracene

Concen: 5.568 ng

RT: 17.257 min Scan# 2545

Delta R.T. 0.153 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54

Instrument :

BNA_F

ClientSampleId :

ENV-DE-3(5-10)





Tgt Ion:278 Resp: 106359

Ion Ratio Lower Upper

278 100

139 17.9 12.8 19.2

279 26.2 18.6 28.0



#92

Benzo(g,h,i)perylene

Concen: 18.748 ng

RT: 17.545 min Scan# 2594

Delta R.T. -0.006 min

Lab File: BF129326.D

Acq: 22 Jul 2022 20:54

Tgt Ion:276 Resp: 372108

Ion Ratio Lower Upper

276 100

277 23.8 19.1 28.7

138 19.9 18.0 27.0

